

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060821\
 Data File : VX022597.D
 Acq On : 09 Jun 2021 05:03
 Operator : JC/MD
 Sample : M2625-36
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 18-B12(198-202)

Quant Time: Jun 09 06:41:17 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060721W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 07 14:27:50 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	57835	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	117788	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	111501	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	43832	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	50036	48.666	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.340%
35) Dibromofluoromethane	5.397	113	35511	48.186	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.380%
50) Toluene-d8	8.653	98	149310	50.777	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.560%
62) 4-Bromofluorobenzene	11.085	95	53103	48.586	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.180%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	10048	21.749	ug/l	100
18) Methyl Acetate	2.721	43	1394	1.094	ug/l	# 82
20) Methylene Chloride	2.800	84	4093	4.243	ug/l	# 78
43) Isopropyl Acetate	6.355	43	10245	4.321	ug/l	# 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060821\
Data File : VX022597.D
Acq On : 09 Jun 2021 05:03
Operator : JC/MD
Sample : M2625-36
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 48 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
18-B12(198-202)

Quant Time: Jun 09 06:41:17 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060721W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 07 14:27:50 2021
Response via : Initial Calibration

