

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060821\
 Data File : VX022553.D
 Acq On : 08 Jun 2021 10:50
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 09 03:58:17 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060721W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 07 14:27:50 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	97	0.00
2 T	Dichlorodifluoromethane	50.000	53.427	-6.9	94	0.00
3 P	Chloromethane	50.000	48.146	3.7	94	0.00
4 C	Vinyl Chloride	50.000	48.599	2.8#	93	0.00
5 T	Bromomethane	50.000	48.221	3.6	94	0.00
6 T	Chloroethane	50.000	47.673	4.7	92	0.00
7 T	Trichlorofluoromethane	50.000	48.970	2.1	92	0.00
8 T	Diethyl Ether	50.000	47.543	4.9	93	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	49.989	0.0	96	0.00
10 T	Methyl Iodide	50.000	44.727	10.5	86	0.00
11 T	Tert butyl alcohol	250.000	225.768	9.7	90	0.00
12 CM	1,1-Dichloroethene	50.000	46.699	6.6#	92	0.00
13 T	Acrolein	250.000	207.860	16.9	87	0.00
14 T	Allyl chloride	50.000	49.271	1.5	93	0.00
15 T	Acrylonitrile	250.000	237.997	4.8	93	0.00
16 T	Acetone	250.000	248.880	0.4	101	0.00
17 T	Carbon Disulfide	50.000	43.811	12.4	93	0.00
18 T	Methyl Acetate	50.000	47.095	5.8	92	0.00
19 T	Methyl tert-butyl Ether	50.000	47.828	4.3	92	0.00
20 T	Methylene Chloride	50.000	43.765	12.5	93	0.00
21 T	trans-1,2-Dichloroethene	50.000	47.311	5.4	94	0.00
22 T	Diisopropyl ether	50.000	47.660	4.7	94	0.00
23 T	Vinyl Acetate	250.000	242.798	2.9	92	0.00
24 P	1,1-Dichloroethane	50.000	47.827	4.3	93	0.00
25 T	2-Butanone	250.000	238.803	4.5	94	0.00
26 T	2,2-Dichloropropane	50.000	50.952	-1.9	95	-0.01
27 T	cis-1,2-Dichloroethene	50.000	46.505	7.0	92	0.00
28 T	Bromochloromethane	50.000	45.650	8.7	95	0.00
29 T	Tetrahydrofuran	250.000	232.964	6.8	90	0.00
30 C	Chloroform	50.000	48.237	3.5#	94	0.00
31 T	Cyclohexane	50.000	49.558	0.9	92	0.00
32 T	1,1,1-Trichloroethane	50.000	50.021	-0.0	93	0.00
33 S	1,2-Dichloroethane-d4	50.000	44.609	10.8	94	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	94	0.00
35 S	Dibromofluoromethane	50.000	47.984	4.0	93	0.00
36 T	1,1-Dichloropropene	50.000	51.973	-3.9	94	0.00
37 T	Ethyl Acetate	50.000	48.077	3.8	90	0.00
38 T	Carbon Tetrachloride	50.000	47.356	5.3	91	0.00
39 T	Methylcyclohexane	50.000	52.209	-4.4	98	0.00
40 TM	Benzene	50.000	50.739	-1.5	93	0.00
41 T	Methacrylonitrile	50.000	49.349	1.3	95	0.00
42 TM	1,2-Dichloroethane	50.000	50.079	-0.2	93	0.00
43 T	Isopropyl Acetate	50.000	50.076	-0.2	92	0.00
44 TM	Trichloroethene	50.000	51.524	-3.0	95	0.00
45 C	1,2-Dichloropropane	50.000	52.210	-4.4#	95	0.00
46 T	Dibromomethane	50.000	48.760	2.5	92	0.00
47 T	Bromodichloromethane	50.000	47.839	4.3	94	0.00
48 T	Methyl methacrylate	50.000	50.761	-1.5	94	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060821\
 Data File : VX022553.D
 Acq On : 08 Jun 2021 10:50
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 09 03:58:17 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060721W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 07 14:27:50 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	1011.683	-1.2	94	0.00
50 S	Toluene-d8	50.000	48.800	2.4	95	0.00
51 T	4-Methyl-2-Pentanone	250.000	258.134	-3.3	93	0.00
52 CM	Toluene	50.000	50.244	-0.5#	95	0.00
53 T	t-1,3-Dichloropropene	50.000	46.373	7.3	95	0.00
54 T	cis-1,3-Dichloropropene	50.000	47.257	5.5	94	0.00
55 T	1,1,2-Trichloroethane	50.000	52.432	-4.9	94	0.00
56 T	Ethyl methacrylate	50.000	47.280	5.4	95	0.00
57 T	1,3-Dichloropropane	50.000	51.122	-2.2	94	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	241.316	3.5	94	0.00
59 T	2-Hexanone	250.000	260.534	-4.2	94	0.00
60 T	Dibromochloromethane	50.000	45.755	8.5	94	0.00
61 T	1,2-Dibromoethane	50.000	51.097	-2.2	94	0.00
62 S	4-Bromofluorobenzene	50.000	49.567	0.9	101	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	96	0.00
64 T	Tetrachloroethene	50.000	51.767	-3.5	96	0.00
65 PM	Chlorobenzene	50.000	51.184	-2.4	96	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	47.149	5.7	93	0.00
67 C	Ethyl Benzene	50.000	52.043	-4.1#	96	0.00
68 T	m/p-Xylenes	100.000	105.781	-5.8	97	0.00
69 T	o-Xylene	50.000	52.688	-5.4	96	0.00
70 T	Styrene	50.000	53.146	-6.3	97	0.00
71 P	Bromoform	50.000	44.274	11.5	94	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	96	0.00
73 T	Isopropylbenzene	50.000	52.303	-4.6	97	0.00
74 T	N-amyl acetate	50.000	49.819	0.4	92	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	50.239	-0.5	95	0.00
76 T	1,2,3-Trichloropropane	50.000	52.004	-4.0	95	0.00
77 T	Bromobenzene	50.000	49.881	0.2	96	0.00
78 T	n-propylbenzene	50.000	51.888	-3.8	98	0.00
79 T	2-Chlorotoluene	50.000	50.934	-1.9	98	0.00
80 T	1,3,5-Trimethylbenzene	50.000	52.088	-4.2	98	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	45.476	9.0	94	0.00
82 T	4-Chlorotoluene	50.000	52.444	-4.9	99	0.00
83 T	tert-Butylbenzene	50.000	52.817	-5.6	96	0.00
84 T	1,2,4-Trimethylbenzene	50.000	52.203	-4.4	97	0.00
85 T	sec-Butylbenzene	50.000	53.418	-6.8	98	0.00
86 T	p-Isopropyltoluene	50.000	54.670	-9.3	100	0.00
87 T	1,3-Dichlorobenzene	50.000	51.301	-2.6	99	0.00
88 T	1,4-Dichlorobenzene	50.000	49.164	1.7	98	0.00
89 T	n-Butylbenzene	50.000	55.317	-10.6	100	0.00
90 T	Hexachloroethane	50.000	45.505	9.0	93	0.00
91 T	1,2-Dichlorobenzene	50.000	51.689	-3.4	98	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	45.720	8.6	94	0.00
93 T	1,2,4-Trichlorobenzene	50.000	52.668	-5.3	102	0.00
94 T	Hexachlorobutadiene	50.000	52.558	-5.1	101	0.00
95 T	Naphthalene	50.000	53.288	-6.6	97	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060821\
Data File : VX022553.D
Acq On : 08 Jun 2021 10:50
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Jun 09 03:58:17 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060721W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 07 14:27:50 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	52.635	-5.3	99	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6