

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060821\
 Data File : VX022580.D
 Acq On : 08 Jun 2021 21:36
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 09 04:02:39 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060721W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 07 14:27:50 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	89	0.00
2 T	Dichlorodifluoromethane	50.000	54.940	-9.9	88	0.00
3 P	Chloromethane	50.000	49.798	0.4	89	0.00
4 C	Vinyl Chloride	50.000	50.359	-0.7#	88	0.00
5 T	Bromomethane	50.000	50.681	-1.4	90	0.00
6 T	Chloroethane	50.000	48.996	2.0	87	0.00
7 T	Trichlorofluoromethane	50.000	50.681	-1.4	87	0.00
8 T	Diethyl Ether	50.000	49.245	1.5	88	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	48.803	2.4	85	0.00
10 T	Methyl Iodide	50.000	47.419	5.2	83	0.00
11 T	Tert butyl alcohol	250.000	262.995	-5.2	96	0.00
12 CM	1,1-Dichloroethene	50.000	47.977	4.0#	86	0.00
13 T	Acrolein	250.000	242.149	3.1	92	0.00
14 T	Allyl chloride	50.000	51.198	-2.4	89	0.00
15 T	Acrylonitrile	250.000	263.639	-5.5	94	0.00
16 T	Acetone	250.000	261.117	-4.4	97	0.00
17 T	Carbon Disulfide	50.000	43.298	13.4	84	0.00
18 T	Methyl Acetate	50.000	53.802	-7.6	96	0.00
19 T	Methyl tert-butyl Ether	50.000	51.639	-3.3	91	0.00
20 T	Methylene Chloride	50.000	47.101	5.8	91	0.00
21 T	trans-1,2-Dichloroethene	50.000	47.926	4.1	87	0.00
22 T	Diisopropyl ether	50.000	51.701	-3.4	93	0.00
23 T	Vinyl Acetate	250.000	264.858	-5.9	91	0.00
24 P	1,1-Dichloroethane	50.000	51.049	-2.1	90	0.00
25 T	2-Butanone	250.000	265.527	-6.2	95	0.00
26 T	2,2-Dichloropropane	50.000	41.374	17.3	70	-0.01
27 T	cis-1,2-Dichloroethene	50.000	49.821	0.4	90	0.00
28 T	Bromochloromethane	50.000	49.358	1.3	94	0.00
29 T	Tetrahydrofuran	250.000	262.353	-4.9	93	0.00
30 C	Chloroform	50.000	51.274	-2.5#	91	0.00
31 T	Cyclohexane	50.000	51.770	-3.5	88	0.00
32 T	1,1,1-Trichloroethane	50.000	52.978	-6.0	89	0.00
33 S	1,2-Dichloroethane-d4	50.000	48.949	2.1	94	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	92	0.00
35 S	Dibromofluoromethane	50.000	48.519	3.0	92	0.00
36 T	1,1-Dichloropropene	50.000	49.156	1.7	87	0.00
37 T	Ethyl Acetate	50.000	53.127	-6.3	97	0.00
38 T	Carbon Tetrachloride	50.000	47.750	4.5	90	0.00
39 T	Methylcyclohexane	50.000	47.167	5.7	87	0.00
40 TM	Benzene	50.000	50.554	-1.1	91	0.00
41 T	Methacrylonitrile	50.000	50.414	-0.8	95	0.00
42 TM	1,2-Dichloroethane	50.000	50.133	-0.3	92	0.00
43 T	Isopropyl Acetate	50.000	51.820	-3.6	94	0.00
44 TM	Trichloroethene	50.000	48.996	2.0	89	0.00
45 C	1,2-Dichloropropane	50.000	52.254	-4.5#	93	0.00
46 T	Dibromomethane	50.000	49.915	0.2	93	0.00
47 T	Bromodichloromethane	50.000	47.233	5.5	91	0.00
48 T	Methyl methacrylate	50.000	52.018	-4.0	94	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060821\
 Data File : VX022580.D
 Acq On : 08 Jun 2021 21:36
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 09 04:02:39 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060721W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 07 14:27:50 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	1016.398	-1.6	93	0.00
50 S	Toluene-d8	50.000	48.910	2.2	94	0.00
51 T	4-Methyl-2-Pentanone	250.000	267.870	-7.1	95	0.00
52 CM	Toluene	50.000	48.918	2.2#	90	0.00
53 T	t-1,3-Dichloropropene	50.000	44.469	11.1	89	0.00
54 T	cis-1,3-Dichloropropene	50.000	45.560	8.9	89	0.00
55 T	1,1,2-Trichloroethane	50.000	53.718	-7.4	95	0.00
56 T	Ethyl methacrylate	50.000	47.919	4.2	95	0.00
57 T	1,3-Dichloropropane	50.000	51.456	-2.9	93	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	241.034	3.6	92	0.00
59 T	2-Hexanone	250.000	271.887	-8.8	96	0.00
60 T	Dibromochloromethane	50.000	45.738	8.5	92	0.00
61 T	1,2-Dibromoethane	50.000	51.345	-2.7	93	0.00
62 S	4-Bromofluorobenzene	50.000	49.256	1.5	98	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	94	0.00
64 T	Tetrachloroethene	50.000	46.026	7.9	83	0.00
65 PM	Chlorobenzene	50.000	50.866	-1.7	94	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	47.720	4.6	92	0.00
67 C	Ethyl Benzene	50.000	51.169	-2.3#	92	0.00
68 T	m/p-Xylenes	100.000	103.598	-3.6	93	0.00
69 T	o-Xylene	50.000	52.185	-4.4	93	0.00
70 T	Styrene	50.000	52.858	-5.7	94	0.00
71 P	Bromoform	50.000	45.507	9.0	95	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	95	0.00
73 T	Isopropylbenzene	50.000	50.534	-1.1	92	0.00
74 T	N-amyl acetate	50.000	52.154	-4.3	95	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	52.371	-4.7	98	0.00
76 T	1,2,3-Trichloropropane	50.000	52.395	-4.8	95	0.00
77 T	Bromobenzene	50.000	49.942	0.1	94	0.00
78 T	n-propylbenzene	50.000	50.303	-0.6	93	0.00
79 T	2-Chlorotoluene	50.000	50.451	-0.9	96	0.00
80 T	1,3,5-Trimethylbenzene	50.000	50.076	-0.2	93	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	44.080	11.8	89	0.00
82 T	4-Chlorotoluene	50.000	49.998	0.0	93	0.00
83 T	tert-Butylbenzene	50.000	52.020	-4.0	93	0.00
84 T	1,2,4-Trimethylbenzene	50.000	50.420	-0.8	93	0.00
85 T	sec-Butylbenzene	50.000	50.779	-1.6	92	0.00
86 T	p-Isopropyltoluene	50.000	50.666	-1.3	92	0.00
87 T	1,3-Dichlorobenzene	50.000	49.228	1.5	94	0.00
88 T	1,4-Dichlorobenzene	50.000	47.016	6.0	92	0.00
89 T	n-Butylbenzene	50.000	49.657	0.7	89	0.00
90 T	Hexachloroethane	50.000	45.731	8.5	93	0.00
91 T	1,2-Dichlorobenzene	50.000	49.966	0.1	93	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	47.405	5.2	97	0.00
93 T	1,2,4-Trichlorobenzene	50.000	48.680	2.6	93	0.00
94 T	Hexachlorobutadiene	50.000	48.640	2.7	92	0.00
95 T	Naphthalene	50.000	52.629	-5.3	95	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060821\
Data File : VX022580.D
Acq On : 08 Jun 2021 21:36
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Jun 09 04:02:39 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060721W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 07 14:27:50 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	50.672	-1.3	94	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6