

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060821\
 Data File : VX022582.D
 Acq On : 08 Jun 2021 22:47
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 09 04:02:55 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060721W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 07 14:27:50 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	88	0.00
2 T	Dichlorodifluoromethane	50.000	56.174	-12.3	89	0.00
3 P	Chloromethane	50.000	50.884	-1.8	90	0.00
4 C	Vinyl Chloride	50.000	51.127	-2.3#	88	0.00
5 T	Bromomethane	50.000	49.849	0.3	88	0.00
6 T	Chloroethane	50.000	51.046	-2.1	90	0.00
7 T	Trichlorofluoromethane	50.000	50.933	-1.9	87	0.00
8 T	Diethyl Ether	50.000	52.072	-4.1	92	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	50.609	-1.2	88	0.00
10 T	Methyl Iodide	50.000	48.213	3.6	84	0.00
11 T	Tert butyl alcohol	250.000	253.704	-1.5	92	0.00
12 CM	1,1-Dichloroethene	50.000	49.378	1.2#	88	0.00
13 T	Acrolein	250.000	219.699	12.1	83	0.00
14 T	Allyl chloride	50.000	51.252	-2.5	88	0.00
15 T	Acrylonitrile	250.000	258.510	-3.4	91	0.00
16 T	Acetone	250.000	249.734	0.1	92	0.00
17 T	Carbon Disulfide	50.000	43.673	12.7	84	0.00
18 T	Methyl Acetate	50.000	52.266	-4.5	93	0.00
19 T	Methyl tert-butyl Ether	50.000	50.855	-1.7	88	0.00
20 T	Methylene Chloride	50.000	46.596	6.8	90	0.00
21 T	trans-1,2-Dichloroethene	50.000	48.385	3.2	87	0.00
22 T	Diisopropyl ether	50.000	50.665	-1.3	90	0.00
23 T	Vinyl Acetate	250.000	259.125	-3.6	89	0.00
24 P	1,1-Dichloroethane	50.000	51.220	-2.4	90	0.00
25 T	2-Butanone	250.000	257.634	-3.1	92	0.00
26 T	2,2-Dichloropropane	50.000	41.119	17.8	69	0.00
27 T	cis-1,2-Dichloroethene	50.000	48.990	2.0	87	0.00
28 T	Bromochloromethane	50.000	48.923	2.2	92	0.00
29 T	Tetrahydrofuran	250.000	254.389	-1.8	89	0.00
30 C	Chloroform	50.000	50.520	-1.0#	89	0.00
31 T	Cyclohexane	50.000	51.958	-3.9	88	0.00
32 T	1,1,1-Trichloroethane	50.000	51.883	-3.8	87	0.00
33 S	1,2-Dichloroethane-d4	50.000	47.994	4.0	92	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	89	0.00
35 S	Dibromofluoromethane	50.000	48.953	2.1	89	0.00
36 T	1,1-Dichloropropene	50.000	50.891	-1.8	87	0.00
37 T	Ethyl Acetate	50.000	51.797	-3.6	91	0.00
38 T	Carbon Tetrachloride	50.000	48.406	3.2	88	0.00
39 T	Methylcyclohexane	50.000	48.809	2.4	86	0.00
40 TM	Benzene	50.000	51.397	-2.8	89	0.00
41 T	Methacrylonitrile	50.000	50.583	-1.2	92	0.00
42 TM	1,2-Dichloroethane	50.000	51.280	-2.6	90	0.00
43 T	Isopropyl Acetate	50.000	51.593	-3.2	90	0.00
44 TM	Trichloroethene	50.000	51.228	-2.5	89	0.00
45 C	1,2-Dichloropropane	50.000	52.760	-5.5#	91	0.00
46 T	Dibromomethane	50.000	50.556	-1.1	91	0.00
47 T	Bromodichloromethane	50.000	48.104	3.8	89	0.00
48 T	Methyl methacrylate	50.000	53.386	-6.8	93	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060821\
 Data File : VX022582.D
 Acq On : 08 Jun 2021 22:47
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 09 04:02:55 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060721W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 07 14:27:50 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	1053.722	-5.4	93	0.00
50 S	Toluene-d8	50.000	49.727	0.5	92	0.00
51 T	4-Methyl-2-Pentanone	250.000	272.948	-9.2	93	0.00
52 CM	Toluene	50.000	50.826	-1.7#	90	0.00
53 T	t-1,3-Dichloropropene	50.000	44.882	10.2	86	0.00
54 T	cis-1,3-Dichloropropene	50.000	46.370	7.3	87	0.00
55 T	1,1,2-Trichloroethane	50.000	54.062	-8.1	92	0.00
56 T	Ethyl methacrylate	50.000	48.371	3.3	92	0.00
57 T	1,3-Dichloropropane	50.000	52.532	-5.1	91	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	247.148	1.1	91	0.00
59 T	2-Hexanone	250.000	272.785	-9.1	93	0.00
60 T	Dibromochloromethane	50.000	46.967	6.1	91	0.00
61 T	1,2-Dibromoethane	50.000	51.886	-3.8	91	0.00
62 S	4-Bromofluorobenzene	50.000	49.704	0.6	96	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	91	0.00
64 T	Tetrachloroethene	50.000	51.927	-3.9	90	0.00
65 PM	Chlorobenzene	50.000	51.759	-3.5	92	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	48.627	2.7	90	0.00
67 C	Ethyl Benzene	50.000	52.742	-5.5#	92	0.00
68 T	m/p-Xylenes	100.000	105.470	-5.5	91	0.00
69 T	o-Xylene	50.000	54.520	-9.0	94	0.00
70 T	Styrene	50.000	53.936	-7.9	93	0.00
71 P	Bromoform	50.000	44.765	10.5	89	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	91	0.00
73 T	Isopropylbenzene	50.000	52.189	-4.4	91	0.00
74 T	N-ethyl acetate	50.000	53.915	-7.8	95	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	51.498	-3.0	92	0.00
76 T	1,2,3-Trichloropropane	50.000	52.746	-5.5	92	0.00
77 T	Bromobenzene	50.000	51.706	-3.4	94	0.00
78 T	n-propylbenzene	50.000	52.135	-4.3	93	0.00
79 T	2-Chlorotoluene	50.000	52.086	-4.2	95	0.00
80 T	1,3,5-Trimethylbenzene	50.000	52.596	-5.2	94	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	42.759	14.5	83	0.00
82 T	4-Chlorotoluene	50.000	52.286	-4.6	94	0.00
83 T	tert-Butylbenzene	50.000	53.407	-6.8	92	0.00
84 T	1,2,4-Trimethylbenzene	50.000	52.536	-5.1	93	0.00
85 T	sec-Butylbenzene	50.000	53.549	-7.1	93	0.00
86 T	p-Isopropyltoluene	50.000	53.307	-6.6	93	0.00
87 T	1,3-Dichlorobenzene	50.000	49.830	0.3	91	0.00
88 T	1,4-Dichlorobenzene	50.000	49.242	1.5	93	0.00
89 T	n-Butylbenzene	50.000	52.809	-5.6	91	0.00
90 T	Hexachloroethane	50.000	47.292	5.4	92	0.00
91 T	1,2-Dichlorobenzene	50.000	52.870	-5.7	95	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	47.494	5.0	93	0.00
93 T	1,2,4-Trichlorobenzene	50.000	50.668	-1.3	93	0.00
94 T	Hexachlorobutadiene	50.000	49.603	0.8	90	0.00
95 T	Naphthalene	50.000	54.143	-8.3	94	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060821\
Data File : VX022582.D
Acq On : 08 Jun 2021 22:47
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Jun 09 04:02:55 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060721W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 07 14:27:50 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	52.277	-4.6	93	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6