



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060821\
 Data File : VX022581.D
 Acq On : 08 Jun 2021 22:00
 Operator : JC/MD
 Sample : BFB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060721W.M
 Title : SW846 8260
 Last Update : Mon Jun 07 14:27:50 2021



AutoFind: Scans 1639, 1640, 1641; Background Corrected with Scan 1632

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.7	5831	PASS
75	95	30	60	52.0	14636	PASS
95	95	100	100	100.0	28120	PASS
96	95	5	9	6.2	1756	PASS
173	174	0.00	2	0.9	155	PASS
174	95	50	100	61.6	17321	PASS
175	174	5	9	6.9	1202	PASS
176	174	95	101	98.3	17026	PASS
177	176	5	9	7.3	1236	PASS