

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060822\
 Data File : VX029314.D
 Acq On : 08 Jun 2022 12:17
 Operator : JC/MD
 Sample : VX0608WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0608WBL01

Quant Time: Jun 08 16:25:52 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 08 03:00:38 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	302081	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	503500	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	440162	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	188989	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	256365	55.961	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	111.920%
35) Dibromofluoromethane	5.385	113	213124	49.790	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	99.580%
50) Toluene-d8	8.647	98	783847	51.777	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	103.560%
62) 4-Bromofluorobenzene	11.079	95	291977	49.284	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	98.560%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060822\
Data File : VX029314.D
Acq On : 08 Jun 2022 12:17
Operator : JC/MD
Sample : VX0608WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0608WBL01

Quant Time: Jun 08 16:25:52 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060722W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 08 03:00:38 2022
Response via : Initial Calibration

