

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060822\
 Data File : VX029319.D
 Acq On : 08 Jun 2022 14:27
 Operator : JC/MD
 Sample : N3174-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MID0622

Quant Time: Jun 09 09:40:50 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 08 03:00:38 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	295586	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	487974	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	419785	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	171392	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	238741	53.171	ug/l	0.00
Spiked Amount	50.000	Range 61 - 141	Recovery	=	106.340%	
35) Dibromofluoromethane	5.391	113	193974	46.758	ug/l	0.00
Spiked Amount	50.000	Range 69 - 133	Recovery	=	93.520%	
50) Toluene-d8	8.653	98	721464	49.173	ug/l	0.00
Spiked Amount	50.000	Range 65 - 126	Recovery	=	98.340%	
62) 4-Bromofluorobenzene	11.085	95	262445	45.709	ug/l	0.00
Spiked Amount	50.000	Range 58 - 135	Recovery	=	91.420%	
Target Compounds						
					Qvalue	
17) Carbon Disulfide	2.520	76	4748	0.434	ug/l	96
19) Methyl tert-butyl Ether	3.117	73	6724	0.438	ug/l #	81
31) Cyclohexane	5.476	56	11239	1.403	ug/l #	95
40) Benzene	6.043	78	82326	4.035	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060822\
Data File : VX029319.D
Acq On : 08 Jun 2022 14:27
Operator : JC/MD
Sample : N3174-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MID0622

Quant Time: Jun 09 09:40:50 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060722W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 08 03:00:38 2022
Response via : Initial Calibration

