

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060822\
 Data File : VX029321.D
 Acq On : 08 Jun 2022 15:14
 Operator : JC/MD
 Sample : PB145329TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB145329TB

Quant Time: Jun 09 07:00:02 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 08 03:00:38 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	289502	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	481165	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	420302	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	180841	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	280020	64.034	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	128.060%
35) Dibromofluoromethane	5.385	113	224400	54.858	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	109.720%
50) Toluene-d8	8.653	98	827713	57.213	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	114.420%
62) 4-Bromofluorobenzene	11.085	95	310332	54.814	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	109.620%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060822\
Data File : VX029321.D
Acq On : 08 Jun 2022 15:14
Operator : JC/MD
Sample : PB145329TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB145329TB

Quant Time: Jun 09 07:00:02 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060722W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 08 03:00:38 2022
Response via : Initial Calibration

