

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060822\
 Data File : VX029333.D
 Acq On : 08 Jun 2022 19:54
 Operator : JC/MD
 Sample : N3028-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP-2-1-5FT

Quant Time: Jun 09 07:02:15 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 08 03:00:38 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	296003	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	501521	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	440468	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	190553	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	256778	57.242	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	114.480%
35) Dibromofluoromethane	5.391	113	210079	49.272	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	98.540%
50) Toluene-d8	8.653	98	793950	52.652	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	105.300%
62) 4-Bromofluorobenzene	11.085	95	299808	50.806	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	101.620%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	86550	36.144	ug/l	94
20) Methylene Chloride	2.794	84	12727	2.380	ug/l	98
25) 2-Butanone	4.568	43	33945	8.961	ug/l	99
43) Isopropyl Acetate	6.348	43	13646	1.249	ug/l	92
52) Toluene	8.720	92	7903	0.587	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060822\
Data File : VX029333.D
Acq On : 08 Jun 2022 19:54
Operator : JC/MD
Sample : N3028-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP-2-1-5FT

Quant Time: Jun 09 07:02:15 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060722W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 08 03:00:38 2022
Response via : Initial Calibration

