

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX060822\  
 Data File : VX029335.D  
 Acq On : 08 Jun 2022 20:40  
 Operator : JC/MD  
 Sample : N3175-21 100X  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 24 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 AUD-1713

Quant Time: Jun 09 07:02:32 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X060722W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Jun 08 03:00:38 2022  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon  | Response | Conc     | Units | Dev(Min)  |
|-----------------------------|--------|-------|----------|----------|-------|-----------|
| -----                       |        |       |          |          |       |           |
| Internal Standards          |        |       |          |          |       |           |
| 1) Pentafluorobenzene       | 5.556  | 168   | 297407   | 50.000   | ug/l  | # 0.00    |
| 34) 1,4-Difluorobenzene     | 6.763  | 114   | 495773   | 50.000   | ug/l  | 0.00      |
| 63) Chlorobenzene-d5        | 10.055 | 117   | 432344   | 50.000   | ug/l  | 0.00      |
| 72) 1,4-Dichlorobenzene-d4  | 12.024 | 152   | 186565   | 50.000   | ug/l  | 0.00      |
| System Monitoring Compounds |        |       |          |          |       |           |
| 33) 1,2-Dichloroethane-d4   | 5.958  | 65    | 264205   | 58.663   | ug/l  | 0.00      |
| Spiked Amount               | 50.000 | Range | 61 - 141 | Recovery | =     | 117.320%  |
| 35) Dibromofluoromethane    | 5.391  | 113   | 213978   | 50.769   | ug/l  | 0.00      |
| Spiked Amount               | 50.000 | Range | 69 - 133 | Recovery | =     | 101.540%  |
| 50) Toluene-d8              | 8.653  | 98    | 806432   | 54.099   | ug/l  | 0.00      |
| Spiked Amount               | 50.000 | Range | 65 - 126 | Recovery | =     | 108.200%  |
| 62) 4-Bromofluorobenzene    | 11.085 | 95    | 300652   | 51.539   | ug/l  | 0.00      |
| Spiked Amount               | 50.000 | Range | 58 - 135 | Recovery | =     | 103.080%  |
| Target Compounds            |        |       |          |          |       |           |
| 16) Acetone                 | 2.392  | 43    | 7388     | 3.071    | ug/l  | Qvalue 91 |
| -----                       |        |       |          |          |       |           |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX060822\  
Data File : VX029335.D  
Acq On : 08 Jun 2022 20:40  
Operator : JC/MD  
Sample : N3175-21 100X  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 24 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
AUD-1713

Quant Time: Jun 09 07:02:32 2022  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X060722W.M  
Quant Title : SW846 8260  
QLast Update : Wed Jun 08 03:00:38 2022  
Response via : Initial Calibration

