

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060823\
 Data File : VX036088.D
 Acq On : 08 Jun 2023 22:08
 Operator : JC/MD
 Sample : 03097-21
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EZEQ5

Quant Time: Jun 09 02:25:18 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML060623WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Jun 07 05:37:48 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.763	114	294059	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	270063	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.018	152	124337	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.368	65	90230	45.163	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	90.320%
7) Chloroethane-d5	1.672	69	73368	47.883	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	95.760%
11) 1,1-Dichloroethene-d2	2.307	65	48040	47.425	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	94.860%
21) 2-Butanone-d5	4.471	46	152436	93.085	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	93.090%
24) Chloroform-d	5.056	84	191855	46.052	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.100%
26) 1,2-Dichloroethane-d4	5.952	65	148478	48.842	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.680%
32) Benzene-d6	5.970	84	387620	47.712	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.420%
36) 1,2-Dichloropropane-d6	7.306	67	121809	48.238	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	96.480%
41) Toluene-d8	8.647	98	347010	46.408	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	92.820%
43) trans-1,3-Dichloroprop...	8.952	79	49132	40.533	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	81.060%
47) 2-Hexanone-d5	9.384	63	105331	93.676	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	93.680%
56) 1,1,2,2-Tetrachloroeth...	11.189	84	160367	46.495	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	92.980%
66) 1,2-Dichlorobenzene-d4	12.317	152	118108	48.926	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	97.860%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060823\
Data File : VX036088.D
Acq On : 08 Jun 2023 22:08
Operator : JC/MD
Sample : 03097-21
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EZEQ5

Quant Time: Jun 09 02:25:18 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML060623WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Jun 07 05:37:48 2023
Response via : Initial Calibration

