

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060823\
 Data File : VX036080.D
 Acq On : 08 Jun 2023 19:02
 Operator : JC/MD
 Sample : 03097-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EZEN2

Quant Time: Jun 09 02:24:21 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML060623WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Jun 07 05:37:48 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.763	114	296702	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	281078	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	128812	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.368	65	99058	49.140	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	98.280%
7) Chloroethane-d5	1.672	69	77117	49.881	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	99.760%
11) 1,1-Dichloroethene-d2	2.307	65	51598	50.484	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	100.960%
21) 2-Butanone-d5	4.471	46	163035	98.671	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	98.670%
24) Chloroform-d	5.056	84	202801	48.246	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.500%
26) 1,2-Dichloroethane-d4	5.958	65	156427	50.999	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	102.000%
32) Benzene-d6	5.970	84	415518	49.142	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.280%
36) 1,2-Dichloropropane-d6	7.306	67	131120	49.890	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	99.780%
41) Toluene-d8	8.647	98	370427	47.598	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	95.200%
43) trans-1,3-Dichloroprop...	8.952	79	54495	43.196	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	86.400%
47) 2-Hexanone-d5	9.384	63	112654	96.262	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	96.260%
56) 1,1,2,2-Tetrachloroeth...	11.189	84	171389	47.743	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	95.480%
66) 1,2-Dichlorobenzene-d4	12.317	152	129505	51.783	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	103.560%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060823\
Data File : VX036080.D
Acq On : 08 Jun 2023 19:02
Operator : JC/MD
Sample : 03097-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EZEN2

Quant Time: Jun 09 02:24:21 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML060623WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Jun 07 05:37:48 2023
Response via : Initial Calibration

