

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX060920\  
 Data File : VX016654.D  
 Acq On : 09 Jun 2020 10:59  
 Operator : JC/SP  
 Sample : VX0609WBS01  
 Misc : 5.0mL/MSVOA X/WATER  
 ALS Vial : 4 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 VX0609WBS01

Quant Time: Jun 10 04:09:28 2020  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_X\METHOD\82X052720W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed May 27 16:31:05 2020  
 Response via : Initial Calibration

| Internal Standards         | R.T.  | QIon | Response | Conc  | Units | Dev(Min) |
|----------------------------|-------|------|----------|-------|-------|----------|
| 1) Pentafluorobenzene      | 5.63  | 168  | 239242   | 50.00 | ug/l  | 0.00     |
| 34) 1,4-Difluorobenzene    | 6.83  | 114  | 352435   | 50.00 | ug/l  | 0.00     |
| 63) Chlorobenzene-d5       | 10.10 | 117  | 326960   | 50.00 | ug/l  | 0.00     |
| 72) 1,4-Dichlorobenzene-d4 | 12.06 | 152  | 164387   | 50.00 | ug/l  | 0.00     |

| System Monitoring Compounds | R.T.  | QIon | Response | Conc  | Units   | Dev(Min) |
|-----------------------------|-------|------|----------|-------|---------|----------|
| 33) 1,2-Dichloroethane-d4   | 6.03  | 65   | 143334   | 48.24 | ug/l    | 0.00     |
| Spiked Amount               |       |      | Recovery | =     | 96.48%  |          |
| 35) Dibromofluoromethane    | 5.46  | 113  | 116452   | 53.66 | ug/l    | 0.00     |
| Spiked Amount               |       |      | Recovery | =     | 107.32% |          |
| 50) Toluene-d8              | 8.70  | 98   | 443168   | 52.47 | ug/l    | 0.00     |
| Spiked Amount               |       |      | Recovery | =     | 104.94% |          |
| 62) 4-Bromofluorobenzene    | 11.12 | 95   | 167017   | 51.17 | ug/l    | 0.00     |
| Spiked Amount               |       |      | Recovery | =     | 102.34% |          |

| Target Compounds              | R.T. | QIon | Response | Conc   | Units | Qvalue |
|-------------------------------|------|------|----------|--------|-------|--------|
| 2) Dichlorodifluoromethane    | 1.18 | 85   | 40252    | 17.096 | ug/l  | 99     |
| 3) Chloromethane              | 1.31 | 50   | 44810    | 15.749 | ug/l  | 97     |
| 4) Vinyl Chloride             | 1.39 | 62   | 49726    | 16.526 | ug/l  | 100    |
| 5) Bromomethane               | 1.62 | 94   | 34173    | 23.358 | ug/l  | 96     |
| 6) Chloroethane               | 1.71 | 64   | 31222    | 17.789 | ug/l  | 98     |
| 7) Trichlorofluoromethane     | 1.92 | 101  | 79086    | 20.014 | ug/l  | 98     |
| 8) Diethyl Ether              | 2.17 | 74   | 29030    | 17.618 | ug/l  | 94     |
| 9) 1,1,2-Trichlorotrifluoroet | 2.36 | 101  | 43157    | 19.844 | ug/l  | 98     |
| 10) Methyl Iodide             | 2.49 | 142  | 41062    | 17.893 | ug/l  | 98     |
| 11) Tert butyl alcohol        | 3.01 | 59   | 63451    | 83.283 | ug/l  | 100    |
| 12) 1,1-Dichloroethene        | 2.36 | 96   | 41745    | 18.134 | ug/l  | 93     |
| 13) Acrolein                  | 2.27 | 56   | 24953    | 76.993 | ug/l  | 98     |
| 14) Allyl chloride            | 2.70 | 41   | 67719    | 15.996 | ug/l  | 96     |
| 15) Acrylonitrile             | 3.11 | 53   | 129485   | 82.968 | ug/l  | 100    |
| 16) Acetone                   | 2.42 | 43   | 105144   | 80.775 | ug/l  | 97     |
| 17) Carbon Disulfide          | 2.55 | 76   | 110225   | 16.124 | ug/l  | 98     |
| 18) Methyl Acetate            | 2.75 | 43   | 57651    | 17.268 | ug/l  | 95     |
| 19) Methyl tert-butyl Ether   | 3.16 | 73   | 150627   | 18.222 | ug/l  | 96     |
| 20) Methylene Chloride        | 2.83 | 84   | 47647    | 17.692 | ug/l  | 94     |
| 21) trans-1,2-Dichloroethene  | 3.14 | 96   | 45132    | 17.725 | ug/l  | 96     |
| 22) Diisopropyl ether         | 3.82 | 45   | 137232   | 16.874 | ug/l  | 93     |
| 23) Vinyl Acetate             | 3.78 | 43   | 585773   | 82.287 | ug/l  | 96     |
| 24) 1,1-Dichloroethane        | 3.67 | 63   | 81487    | 17.752 | ug/l  | 99     |
| 25) 2-Butanone                | 4.63 | 43   | 174677   | 79.194 | ug/l  | 97     |
| 26) 2,2-Dichloropropane       | 4.54 | 77   | 78451    | 20.886 | ug/l  | 99     |
| 27) cis-1,2-Dichloroethene    | 4.56 | 96   | 53169    | 18.049 | ug/l  | 99     |
| 28) Bromochloromethane        | 4.98 | 49   | 34540    | 16.192 | ug/l  | 83     |
| 29) Tetrahydrofuran           | 5.09 | 42   | 112766   | 79.136 | ug/l  | 94     |
| 30) Chloroform                | 5.17 | 83   | 86415    | 18.684 | ug/l  | 96     |
| 31) Cyclohexane               | 5.55 | 56   | 65513    | 15.864 | ug/l  | 94     |
| 32) 1,1,1-Trichloroethane     | 5.46 | 97   | 80669    | 19.265 | ug/l  | 98     |
| 36) 1,1-Dichloropropene       | 5.76 | 75   | 62279    | 18.664 | ug/l  | 97     |
| 37) Ethyl Acetate             | 4.79 | 43   | 65699    | 16.399 | ug/l  | 97     |
| 38) Carbon Tetrachloride      | 5.75 | 117  | 71775    | 20.468 | ug/l  | 95     |

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX060920\  
 Data File : VX016654.D  
 Acq On : 09 Jun 2020 10:59  
 Operator : JC/SP  
 Sample : VX0609WBS01  
 Misc : 5.0mL/MSVOA X/WATER  
 ALS Vial : 4 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 VX0609WBS01

Quant Time: Jun 10 04:09:28 2020  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_X\METHOD\82X052720W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed May 27 16:31:05 2020  
 Response via : Initial Calibration

| Internal Standards             | R.T.  | QIon | Response | Conc    | Units  | Dev(Min) |
|--------------------------------|-------|------|----------|---------|--------|----------|
| 39) Methylcyclohexane          | 7.44  | 83   | 67367    | 18.587  | ug/l   | 97       |
| 40) Benzene                    | 6.11  | 78   | 186912   | 18.810  | ug/l   | 100      |
| 41) Methacrylonitrile          | 4.99  | 41   | 36806    | 16.726  | ug/l   | 97       |
| 42) 1,2-Dichloroethane         | 6.16  | 62   | 68237    | 19.309  | ug/l   | 99       |
| 43) Isopropyl Acetate          | 6.41  | 43   | 105612   | 16.901  | ug/l   | 97       |
| 44) Trichloroethene            | 7.18  | 130  | 63227    | 20.054  | ug/l   | 98       |
| 45) 1,2-Dichloropropane        | 7.49  | 63   | 46883    | 18.533  | ug/l   | 99       |
| 46) Dibromomethane             | 7.64  | 93   | 33102    | 19.481  | ug/l   | 97       |
| 47) Bromodichloromethane       | 7.87  | 83   | 69352    | 19.712  | ug/l   | 96       |
| 48) Methyl methacrylate        | 7.74  | 41   | 49364    | 16.499  | ug/l   | 95       |
| 49) 1,4-Dioxane                | 7.71  | 88   | 25887    | 369.705 | ug/l   | 96       |
| 51) 4-Methyl-2-Pentanone       | 8.62  | 43   | 330676   | 84.985  | ug/l   | 97       |
| 52) Toluene                    | 8.76  | 92   | 120805   | 19.279  | ug/l   | 100      |
| 53) t-1,3-Dichloropropene      | 9.02  | 75   | 77669    | 19.414  | ug/l   | 97       |
| 54) cis-1,3-Dichloropropene    | 8.42  | 75   | 82002    | 19.365  | ug/l   | 97       |
| 55) 1,1,2-Trichloroethane      | 9.20  | 97   | 50000    | 20.068  | ug/l   | 95       |
| 56) Ethyl methacrylate         | 9.16  | 69   | 73065    | 18.007  | ug/l   | 96       |
| 57) 1,3-Dichloropropane        | 9.35  | 76   | 81328    | 19.017  | ug/l   | 100      |
| 58) 2-Chloroethyl Vinyl ether  | 8.29  | 63   | 169127   | 79.854  | ug/l   | 95       |
| 59) 2-Hexanone                 | 9.47  | 43   | 255251   | 84.356  | ug/l   | 98       |
| 60) Dibromochloromethane       | 9.56  | 129  | 57158    | 20.386  | ug/l   | 98       |
| 61) 1,2-Dibromoethane          | 9.65  | 107  | 52621    | 19.723  | ug/l   | 100      |
| 64) Tetrachloroethene          | 9.32  | 164  | 73168    | 20.002  | ug/l   | 96       |
| 65) Chlorobenzene              | 10.12 | 112  | 131549   | 19.735  | ug/l   | 99       |
| 66) 1,1,1,2-Tetrachloroethane  | 10.20 | 131  | 51947    | 20.415  | ug/l   | 99       |
| 67) Ethyl Benzene              | 10.23 | 91   | 224558   | 18.921  | ug/l   | 99       |
| 68) m/p-Xylenes                | 10.34 | 106  | 174379   | 39.629  | ug/l   | 97       |
| 69) o-Xylene                   | 10.68 | 106  | 81681    | 19.167  | ug/l   | 98       |
| 70) Styrene                    | 10.70 | 104  | 142353   | 19.434  | ug/l   | 100      |
| 71) Bromoform                  | 10.84 | 173  | 44520    | 20.703  | ug/l # | 98       |
| 73) Isopropylbenzene           | 11.00 | 105  | 222855   | 19.566  | ug/l   | 100      |
| 74) N-amyl acetate             | 10.88 | 43   | 84620    | 16.152  | ug/l   | 95       |
| 75) 1,1,2,2-Tetrachloroethane  | 11.25 | 83   | 54075    | 18.943  | ug/l   | 99       |
| 76) 1,2,3-Trichloropropane     | 11.28 | 75   | 89193    | 24.685  | ug/l   | 84       |
| 77) Bromobenzene               | 11.24 | 156  | 58968    | 19.739  | ug/l   | 95       |
| 78) n-propylbenzene            | 11.34 | 91   | 242567   | 19.062  | ug/l   | 98       |
| 79) 2-Chlorotoluene            | 11.40 | 91   | 150715   | 19.404  | ug/l   | 100      |
| 80) 1,3,5-Trimethylbenzene     | 11.49 | 105  | 190983   | 20.081  | ug/l   | 100      |
| 81) trans-1,4-Dichloro-2-buten | 11.06 | 75   | 27053    | 18.417  | ug/l   | 96       |
| 82) 4-Chlorotoluene            | 11.49 | 91   | 177354   | 19.335  | ug/l   | 99       |
| 83) tert-Butylbenzene          | 11.76 | 119  | 160089   | 19.700  | ug/l   | 99       |
| 84) 1,2,4-Trimethylbenzene     | 11.79 | 105  | 185270   | 19.263  | ug/l   | 99       |
| 85) sec-Butylbenzene           | 11.93 | 105  | 199868   | 19.606  | ug/l   | 99       |
| 86) p-Isopropyltoluene         | 12.05 | 119  | 190553   | 19.955  | ug/l   | 99       |
| 87) 1,3-Dichlorobenzene        | 12.01 | 146  | 102230   | 19.752  | ug/l   | 99       |
| 88) 1,4-Dichlorobenzene        | 12.08 | 146  | 105115   | 19.858  | ug/l   | 98       |
| 89) n-Butylbenzene             | 12.37 | 91   | 153564   | 19.307  | ug/l   | 99       |
| 90) Hexachloroethane           | 12.58 | 117  | 34127    | 20.690  | ug/l   | 96       |
| 91) 1,2-Dichlorobenzene        | 12.38 | 146  | 98633    | 19.670  | ug/l   | 99       |
| 92) 1,2-Dibromo-3-Chloropropan | 12.98 | 75   | 16883    | 18.259  | ug/l   | 98       |

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX060920\  
Data File : VX016654.D  
Acq On : 09 Jun 2020 10:59  
Operator : JC/SP  
Sample : VX0609WBS01  
Misc : 5.0mL/MSVOA X/WATER  
ALS Vial : 4 Sample Multiplier: 1


Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0609WBS01

Quant Time: Jun 10 04:09:28 2020  
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA\_X\METHOD\82X052720W.M  
Quant Title : SW846 8260  
QLast Update : Wed May 27 16:31:05 2020  
Response via : Initial Calibration

| Internal Standards         | R.T.  | QIon | Response | Conc   | Units | Dev(Min) |
|----------------------------|-------|------|----------|--------|-------|----------|
| 93) 1,2,4-Trichlorobenzene | 13.63 | 180  | 64827    | 19.466 | ug/l  | 97       |
| 94) Hexachlorobutadiene    | 13.77 | 225  | 32960    | 25.208 | ug/l  | 97       |
| 95) Naphthalene            | 13.82 | 128  | 215758   | 18.899 | ug/l  | 100      |
| 96) 1,2,3-Trichlorobenzene | 14.00 | 180  | 68440    | 20.977 | ug/l  | 98       |

(#) = qualifier out of range (m) = manual integration (+) = signals summed





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.63 min Scan# 745

Delta R.T. -0.00 min

Lab File: VX016654.D

Acq: 09 Jun 2020 10:59

Instrument :

MSVOA\_X

ClientSampled :


VX0609WBS01

Tot Ion:168 Resp: 239242

Ion Ratio Lower Upper

168 100

99 52.2 45.2 67.8



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0

5.63

80000



60000

40000

20000

0

Time-->



#2

Dichlorodifluoromethane

Concen: 17.096 ug/l

RT: 1.18 min Scan# 16

Delta R.T. 0.01 min

Lab File: VX016654.D


Acq: 09 Jun 2020 10:59

Tgt Ion: 85 Resp: 40252

Ion Ratio Lower Upper

85 100

87 31.5 15.4 46.2



Abundance Ion 84.90 (84.60 to 85.60): VX0

Ion 86.90 (86.60 to 87.60): VX0

1.18

50000

40000


30000


20000

10000

0

Time-->





#3

Chloromethane

Concen: 15.749 ug/l

RT: 1.31 min Scan# 37

Delta R.T. -0.00 min

Lab File: VX016654.D

Acq: 09 Jun 2020 10:59

Instrument :

MSVOA\_X

ClientSampled :





VX0609WBS01

Tot Ion: 50 Resp: 44810

Ion Ratio Lower Upper

50 100

52 34.2 26.2 39.2



#4

Vinyl Chloride

Concen: 16.526 ug/l

RT: 1.39 min Scan# 50

Delta R.T. -0.00 min

Lab File: VX016654.D




Acq: 09 Jun 2020 10:59


Tgt Ion: 62 Resp: 49726

Ion Ratio Lower Upper

62 100

64 32.7 25.9 38.9





#5

Bromomethane

Concen: 23.358 ug/l

RT: 1.62 min Scan# 88

Delta R.T. -0.00 min

Lab File: VX016654.D

Acq: 09 Jun 2020 10:59

Instrument :

MSVOA\_X

ClientSampleId :


VX0609WBS01

Tot Ion: 94 Resp: 34173

Ion Ratio Lower Upper

94 100

96 90.1 75.1 112.7



Abundance Ion 93.90 (93.60 to 94.60): VX0

Ion 95.90 (95.60 to 96.60): VX0

30000

25000

20000

15000



10000

5000

0

1.55 1.60 1.65 1.70 1.75

Time-->



#6

Chloroethane

Concen: 17.789 ug/l

RT: 1.71 min Scan# 102

Delta R.T. 0.01 min

Lab File: VX016654.D


Acq: 09 Jun 2020 10:59

Tgt Ion: 64 Resp: 31222

Ion Ratio Lower Upper

64 100

66 32.2 26.7 40.1



Abundance Ion 63.90 (63.60 to 64.60): VX0

Ion 65.90 (65.60 to 66.60): VX0

30000

25000

20000

15000


10000


5000

0

1.65 1.70 1.75 1.80

Time-->









#7

Trichlorofluoromethane  
Concen: 20.014 ug/l  
RT: 1.92 min Scan# 136  
Delta R.T. 0.01 min  
Lab File: VX016654.D  
Acq: 09 Jun 2020 10:59

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0609WBS01




Tot Ion:101 Resp: 79086  
Ion Ratio Lower Upper  
101 100  
103 65.1 50.9 76.3




#8

Diethyl Ether  
Concen: 17.618 ug/l  
RT: 2.17 min Scan# 177  
Delta R.T. -0.00 min  
Lab File: VX016654.D  
Acq: 09 Jun 2020 10:59

Tgt Ion: 74 Resp: 29030  
Ion Ratio Lower Upper  
74 100  
45 88.7 47.5 142.5







#9

1,1,2-Trichlorotrifluoroethane

Concen: 19.844 ug/l

RT: 2.36 min Scan# 209

Delta R.T. -0.00 min

Lab File: VX016654.D

Acq: 09 Jun 2020 10:59

Instrument :

MSVOA\_X

ClientSampleId :

VX0609WBS01


Tot Ion:101 Resp: 43157

Ion Ratio Lower Upper

101 100

85 41.7 34.4 51.6




151 81.8 63.8 95.6



Abundance Ion 100.90 (100.60 to 101.60): V

Ion 84.90 (84.60 to 85.60): VX0

Ion 150.80 (150.50 to 151.50): V



#10

Methyl Iodide

Concen: 17.893 ug/l

RT: 2.49 min Scan# 230

Delta R.T. 0.00 min

Lab File: VX016654.D

Acq: 09 Jun 2020 10:59


Tgt Ion:142 Resp: 41062

Ion Ratio Lower Upper

142 100

127 45.2 35.3 52.9



141 15.2 11.4 17.2




Abundance Ion 141.80 (141.50 to 142.50): V

Ion 126.80 (126.50 to 127.50): V

Ion 140.80 (140.50 to 141.50): V









#11

Tert butyl alcohol  
 Concen: 83.283 ug/l  
 RT: 3.01 min Scan# 315  
 Delta R.T. -0.01 min  
 Lab File: VX016654.D  
 Acq: 09 Jun 2020 10:59

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 VX0609WBS01




Tot Ion: 59 Resp: 63451  
 Ion Ratio Lower Upper  
 59 100  
 57 10.3 8.2 12.4




#12

1,1-Dichloroethene  
 Concen: 18.134 ug/l  
 RT: 2.36 min Scan# 208  
 Delta R.T. -0.00 min  
 Lab File: VX016654.D  
 Acq: 09 Jun 2020 10:59

Tgt Ion: 96 Resp: 41745  
 Ion Ratio Lower Upper  
 96 100  
 61 167.0 126.6 189.8  
 98 69.8 51.3 76.9





#13

Acrolein

Concen: 76.993 ug/l

RT: 2.27 min Scan# 194

Delta R.T. -0.00 min

Lab File: VX016654.D

Acq: 09 Jun 2020 10:59

Instrument :

MSVOA\_X

ClientSampleId :





VX0609WBS01

Tot Ion: 56 Resp: 24953

Ion Ratio Lower Upper

56 100

55 71.0 58.3 87.5



#14

Allyl chloride

Concen: 15.996 ug/l

RT: 2.70 min Scan# 265

Delta R.T. -0.00 min

Lab File: VX016654.D

Acq: 09 Jun 2020 10:59




Tgt Ion: 41 Resp: 67719


Ion Ratio Lower Upper

41 100

39 62.3 47.7 71.5

76 36.6 26.6 40.0





#15

Acrylonitrile

Concen: 82.968 ug/l

RT: 3.11 min Scan# 332

Delta R.T. -0.01 min

Lab File: VX016654.D

Acq: 09 Jun 2020 10:59

Instrument :

MSVOA\_X

ClientSampleId :

VX0609WBS01





Tot Ion: 53 Resp: 129485

Ion Ratio Lower Upper

53 100

52 83.6 66.7 100.1

51 35.8 28.6 43.0



#16

Acetone

Concen: 80.775 ug/l

RT: 2.42 min Scan# 218

Delta R.T. -0.00 min

Lab File: VX016654.D




Acq: 09 Jun 2020 10:59


Tgt Ion: 43 Resp: 105144

Ion Ratio Lower Upper

43 100

58 34.4 26.1 39.1





#17

Carbon Disulfide

Concen: 16.124 ug/l

RT: 2.55 min Scan# 240

Delta R.T. -0.00 min

Lab File: VX016654.D

Acq: 09 Jun 2020 10:59

Instrument :

MSVOA\_X

ClientSampleId :





VX0609WBS01

Tot Ion: 76 Resp: 110225

Ion Ratio Lower Upper

76 100

78 8.5 7.4 11.0



#18

Methyl Acetate

Concen: 17.268 ug/l

RT: 2.75 min Scan# 272

Delta R.T. -0.00 min

Lab File: VX016654.D




Acq: 09 Jun 2020 10:59


Tgt Ion: 43 Resp: 57651

Ion Ratio Lower Upper

43 100

74 26.9 19.7 29.5









#19

Methyl tert-butyl Ether  
 Concen: 18.222 ug/l  
 RT: 3.16 min Scan# 340  
 Delta R.T. -0.01 min  
 Lab File: VX016654.D  
 Acq: 09 Jun 2020 10:59

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 VX0609WBS01




Tot Ion: 73 Resp: 150627  
 Ion Ratio Lower Upper  
 73 100  
 57 20.3 17.8 26.6




#20

Methylene Chloride  
 Concen: 17.692 ug/l  
 RT: 2.83 min Scan# 286  
 Delta R.T. -0.00 min  
 Lab File: VX016654.D  
 Acq: 09 Jun 2020 10:59

Tgt Ion: 84 Resp: 47647  
 Ion Ratio Lower Upper  
 84 100  
 49 110.7 94.5 141.7  
 51 33.2 30.2 45.4  
 86 65.8 51.1 76.7





#21

trans-1,2-Dichloroethene

Concen: 17.725 ug/l

RT: 3.14 min Scan# 337

Delta R.T. -0.00 min

Lab File: VX016654.D

Acq: 09 Jun 2020 10:59

Instrument :

MSVOA\_X

ClientSampleId :

VX0609WBS01


Tot Ion: 96 Resp: 45132

Ion Ratio Lower Upper

96 100

61 137.6 115.3 172.9




98 65.6 51.9 77.9



Abundance Ion 95.90 (95.60 to 96.60): VX0

Ion 60.90 (60.60 to 61.60): VX0

Ion 97.90 (97.60 to 98.60): VX0



#22

Diisopropyl ether

Concen: 16.874 ug/l

RT: 3.82 min Scan# 448

Delta R.T. -0.01 min

Lab File: VX016654.D

Acq: 09 Jun 2020 10:59

Tgt Ion: 45 Resp: 137232


Ion Ratio Lower Upper

45 100

43 78.1 57.0 85.6

87 30.3 22.5 33.7

59 12.5 9.1 13.7






Abundance Ion 45.00 (44.70 to 45.70): VX0

Ion 43.00 (42.70 to 43.70): VX0

Ion 87.00 (86.70 to 87.70): VX0

Ion 59.00 (58.70 to 59.70): VX0





#23

Vinyl Acetate

Concen: 82.287 ug/l

RT: 3.78 min Scan# 442

Delta R.T. -0.01 min

Lab File: VX016654.D

Acq: 09 Jun 2020 10:59

Instrument :

MSVOA\_X

ClientSampled :


VX0609WBS01

Tot Ion: 43 Resp: 585773

Ion Ratio Lower Upper




43 100

86 12.1 8.6 13.0



Abundance Ion 43.00 (42.70 to 43.70): VX0

Ion 86.00 (85.70 to 86.70): VX0



#24

1,1-Dichloroethane

Concen: 17.752 ug/l

RT: 3.67 min Scan# 423

Delta R.T. -0.01 min

Lab File: VX016654.D

Acq: 09 Jun 2020 10:59


Tgt Ion: 63 Resp: 81487

Ion Ratio Lower Upper

63 100

98 7.7 3.6 10.9



100 4.8 2.4 7.1




Abundance Ion 62.90 (62.60 to 63.60): VX0

Ion 97.90 (97.60 to 98.60): VX0

Ion 99.90 (99.60 to 100.60): VX0











#25  
 2-Butanone  
 Concen: 79.194 ug/l  
 RT: 4.63 min Scan# 581  
 Delta R.T. -0.01 min  
 Lab File: VX016654.D  
 Acq: 09 Jun 2020 10:59




Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 VX0609WBS01


Tot Ion: 43 Resp: 174677  
 Ion Ratio Lower Upper  
 43 100  
 72 27.9 21.0 31.6



#26  
 2,2-Dichloropropane  
 Concen: 20.886 ug/l  
 RT: 4.54 min Scan# 567  
 Delta R.T. -0.01 min  
 Lab File: VX016654.D  
 Acq: 09 Jun 2020 10:59

Tgt Ion: 77 Resp: 78451  
 Ion Ratio Lower Upper  
 77 100  
 97 22.7 11.5 34.4









#27

cis-1,2-Dichloroethene  
 Concen: 18.049 ug/l  
 RT: 4.56 min Scan# 570  
 Delta R.T. -0.00 min  
 Lab File: VX016654.D  
 Acq: 09 Jun 2020 10:59

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 VX0609WBS01




| Tot Ion   | 96    | Resp  | 53169 |
|-----------|-------|-------|-------|
| Ion Ratio | Lower | Upper |       |
| 96        | 100   |       |       |
| 61        | 139.8 | 0.0   | 284.6 |
| 98        | 63.6  | 0.0   | 127.2 |




#28

Bromochloromethane  
 Concen: 16.192 ug/l  
 RT: 4.98 min Scan# 638  
 Delta R.T. -0.01 min  
 Lab File: VX016654.D  
 Acq: 09 Jun 2020 10:59

| Tgt Ion   | 49    | Resp  | 34540 |
|-----------|-------|-------|-------|
| Ion Ratio | Lower | Upper |       |
| 49        | 100   |       |       |
| 129       | 1.8   | 0.0   | 4.0   |
| 130       | 94.1  | 63.0  | 94.4  |









#29  
Tetrahydrofuran  
Concen: 79.136 ug/l  
RT: 5.09 min Scan# 656  
Delta R.T. -0.01 min  
Lab File: VX016654.D  
Acq: 09 Jun 2020 10:59




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0609WBS01


Tot Ion: 42 Resp: 112766  
Ion Ratio Lower Upper  
42 100  
72 50.6 37.0 55.4  
71 46.0 34.1 51.1



#30  
Chloroform  
Concen: 18.684 ug/l  
RT: 5.17 min Scan# 670  
Delta R.T. -0.01 min  
Lab File: VX016654.D  
Acq: 09 Jun 2020 10:59

Tgt Ion: 83 Resp: 86415  
Ion Ratio Lower Upper  
83 100  
85 62.0 52.0 78.0









#31  
Cyclohexane  
Concen: 15.864 ug/l  
RT: 5.55 min Scan# 732  
Delta R.T. -0.00 min  
Lab File: VX016654.D  
Acq: 09 Jun 2020 10:59




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0609WBS01


Tot Ion: 56 Resp: 65513  
Ion Ratio Lower Upper  
56 100  
69 33.8 25.7 38.5  
84 94.3 70.0 105.0



#32  
1,1,1-Trichloroethane  
Concen: 19.265 ug/l  
RT: 5.46 min Scan# 717  
Delta R.T. -0.01 min  
Lab File: VX016654.D  
Acq: 09 Jun 2020 10:59

Tgt Ion: 97 Resp: 80669  
Ion Ratio Lower Upper  
97 100  
99 65.2 51.4 77.2  
61 39.5 33.5 50.3









#33  
 1,2-Dichloroethane-d4  
 Concen: 48.243 ug/l  
 RT: 6.03 min Scan# 810  
 Delta R.T. -0.01 min  
 Lab File: VX016654.D  
 Acq: 09 Jun 2020 10:59




Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 VX0609WBS01


Tot Ion: 65 Resp: 143334  
 Ion Ratio Lower Upper  
 65 100  
 67 51.7 0.0 104.8



#34  
 1,4-Difluorobenzene  
 Concen: 50.000 ug/l  
 RT: 6.83 min Scan# 942  
 Delta R.T. -0.00 min  
 Lab File: VX016654.D  
 Acq: 09 Jun 2020 10:59

Tgt Ion: 114 Resp: 352435  
 Ion Ratio Lower Upper  
 114 100  
 63 19.7 0.0 42.6  
 88 16.6 0.0 32.8





#35

Dibromofluoromethane

Concen: 53.662 ug/l

RT: 5.46 min Scan# 717

Delta R.T. -0.01 min

Lab File: VX016654.D

Acq: 09 Jun 2020 10:59

Instrument :

MSVOA\_X

ClientSampled :

VX0609WBS01


Tot Ion:113 Resp: 116452

Ion Ratio Lower Upper

113 100

111 102.7 83.0 124.6




192 22.6 17.7 26.5



Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V



#36

1,1-Dichloropropene

Concen: 18.664 ug/l

RT: 5.76 min Scan# 767

Delta R.T. -0.01 min

Lab File: VX016654.D

Acq: 09 Jun 2020 10:59


Tgt Ion: 75 Resp: 62279

Ion Ratio Lower Upper

75 100

110 37.0 17.4 52.3



77 31.2 24.2 36.4




Abundance Ion 74.90 (74.60 to 75.60): VX0

Ion 109.90 (109.60 to 110.60): V

Ion 76.90 (76.60 to 77.60): VX0









#37  
Ethyl Acetate  
Concen: 16.399 ug/l  
RT: 4.79 min Scan# 607  
Delta R.T. -0.01 min  
Lab File: VX016654.D  
Acq: 09 Jun 2020 10:59




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0609WBS01


Tot Ion: 43 Resp: 65699  
Ion Ratio Lower Upper  
43 100  
61 15.5 11.3 16.9  
70 12.9 9.3 13.9



#38  
Carbon Tetrachloride  
Concen: 20.468 ug/l  
RT: 5.75 min Scan# 765  
Delta R.T. -0.01 min  
Lab File: VX016654.D  
Acq: 09 Jun 2020 10:59

Tgt Ion: 117 Resp: 71775  
Ion Ratio Lower Upper  
117 100  
119 99.4 76.3 114.5  
121 35.0 24.6 37.0





#39

Methylvlcyclohexane

Concen: 18.587 ug/l

RT: 7.44 min Scan# 1042

Delta R.T. -0.00 min

Lab File: VX016654.D

Acq: 09 Jun 2020 10:59

Instrument :

MSVOA\_X

ClientSampleId :

VX0609WBS01





Tot Ion: 83 Resp: 67367

Ion Ratio Lower Upper

83 100

55 73.7 61.7 92.5

98 47.1 38.7 58.1



#40

Benzene

Concen: 18.810 ug/l

RT: 6.11 min Scan# 824

Delta R.T. -0.00 min

Lab File: VX016654.D




Acq: 09 Jun 2020 10:59

Tgt Ion: 78 Resp: 186912


Ion Ratio Lower Upper

78 100

77 23.5 18.9 28.3











#41  
Methacrylonitrile  
Concen: 16.726 ug/l  
RT: 4.99 min Scan# 641  
Delta R.T. -0.01 min  
Lab File: VX016654.D  
Acq: 09 Jun 2020 10:59




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0609WBS01


|          |       |       |       |
|----------|-------|-------|-------|
| Tot Ion: | 41    | Resp: | 36806 |
| Ion      | Ratio | Lower | Upper |
| 41       | 100   |       |       |
| 39       | 53.2  | 43.0  | 64.4  |
| 67       | 77.2  | 58.9  | 88.3  |
| 52       | 30.9  | 23.1  | 34.7  |



#42  
1,2-Dichloroethane  
Concen: 19.309 ug/l  
RT: 6.16 min Scan# 832  
Delta R.T. -0.01 min  
Lab File: VX016654.D  
Acq: 09 Jun 2020 10:59

|          |       |       |       |
|----------|-------|-------|-------|
| Tgt Ion: | 62    | Resp: | 68237 |
| Ion      | Ratio | Lower | Upper |
| 62       | 100   |       |       |
| 98       | 10.3  | 0.0   | 19.8  |





#43

Isopropyl Acetate

Concen: 16.901 ug/l

RT: 6.41 min Scan# 873

Delta R.T. -0.01 min

Lab File: VX016654.D

Acq: 09 Jun 2020 10:59

Instrument :

MSVOA\_X

ClientSampled :

VX0609WBS01





Tot Ion: 43 Resp: 105612

Ion Ratio Lower Upper

43 100

61 22.0 16.8 25.2

87 15.5 11.2 16.8



#44

Trichloroethene

Concen: 20.054 ug/l

RT: 7.18 min Scan# 1000

Delta R.T. -0.01 min

Lab File: VX016654.D




Acq: 09 Jun 2020 10:59


Tgt Ion:130 Resp: 63227

Ion Ratio Lower Upper

130 100

95 97.5 0.0 198.2









#45  
 1,2-Dichloropropane  
 Concen: 18.533 ug/l  
 RT: 7.49 min Scan# 1050  
 Delta R.T. -0.00 min  
 Lab File: VX016654.D  
 Acq: 09 Jun 2020 10:59




Instrument :  
 MSVOA\_X  
 ClientSampled :  
 VX0609WBS01


Tot Ion: 63 Resp: 46883  
 Ion Ratio Lower Upper  
 63 100  
 65 30.9 24.2 36.2



#46  
 Dibromomethane  
 Concen: 19.481 ug/l  
 RT: 7.64 min Scan# 1075  
 Delta R.T. -0.00 min  
 Lab File: VX016654.D  
 Acq: 09 Jun 2020 10:59

Tgt Ion: 93 Resp: 33102  
 Ion Ratio Lower Upper  
 93 100  
 95 83.4 66.6 100.0  
 174 113.7 86.2 129.4





#47

Bromodichloromethane

Concen: 19.712 ug/l

RT: 7.87 min Scan# 1113

Delta R.T. -0.01 min

Lab File: VX016654.D

Acq: 09 Jun 2020 10:59

Instrument :

MSVOA\_X

ClientSampleId :

VX0609WBS01


Tot Ion: 83 Resp: 69352

Ion Ratio Lower Upper

83 100

85 61.8 52.0 78.0




127 8.6 7.4 11.0



Abundance Ion 82.90 (82.60 to 83.60): VX0

Ion 84.90 (84.60 to 85.60): VX0

Ion 126.80 (126.50 to 127.50): V



#48

Methyl methacrylate

Concen: 16.499 ug/l

RT: 7.74 min Scan# 1092

Delta R.T. -0.00 min

Lab File: VX016654.D

Acq: 09 Jun 2020 10:59


Tgt Ion: 41 Resp: 49364

Ion Ratio Lower Upper

41 100

69 92.6 68.6 102.8



39 54.5 42.4 63.6




Abundance Ion 41.00 (40.70 to 41.70): VX0

Ion 69.00 (68.70 to 69.70): VX0

Ion 39.00 (38.70 to 39.70): VX0





#49

1,4-Dioxane

Concen: 369.705 ug/l

RT: 7.71 min Scan# 1087

Delta R.T. -0.00 min

Lab File: VX016654.D

Acq: 09 Jun 2020 10:59

Instrument :

MSVOA\_X

ClientSampleId :

VX0609WBS01


Tot Ion: 88 Resp: 25887

Ion Ratio Lower Upper

88 100

43 31.1 24.2 36.4




58 66.3 56.5 84.7



Abundance Ion 88.00 (87.70 to 88.70): VX0

Ion 43.00 (42.70 to 43.70): VX0

Ion 58.00 (57.70 to 58.70): VX0



#50

Toluene-d8

Concen: 52.468 ug/l

RT: 8.70 min Scan# 1248

Delta R.T. -0.00 min

Lab File: VX016654.D


Acq: 09 Jun 2020 10:59

Tgt Ion: 98 Resp: 443168

Ion Ratio Lower Upper



98 100


100 66.4 53.8 80.6



Abundance Ion 98.00 (97.70 to 98.70): VX0

Ion 100.00 (99.70 to 100.70): VX0





#51

4-Methyl-2-Pentanone

Concen: 84.985 ug/l

RT: 8.62 min Scan# 1235

Delta R.T. -0.01 min

Lab File: VX016654.D

Acq: 09 Jun 2020 10:59

Instrument :

MSVOA\_X

ClientSampleId :


VX0609WBS01

Tot Ion: 43 Resp: 330676

Ion Ratio Lower Upper

43 100

58 42.9 33.0 49.6



Abundance Ion 43.00 (42.70 to 43.70): VX0

Ion 58.00 (57.70 to 58.70): VX0

8.62

200000



150000

100000

50000

0

Time-->



#52

Toluene

Concen: 19.279 ug/l

RT: 8.76 min Scan# 1259

Delta R.T. -0.01 min

Lab File: VX016654.D


Acq: 09 Jun 2020 10:59

Tgt Ion: 92 Resp: 120805

Ion Ratio Lower Upper

92 100

91 169.3 135.0 202.4



Abundance Ion 92.00 (91.70 to 92.70): VX0

Ion 91.00 (90.70 to 91.70): VX0

8.76


150000


100000

50000

0

Time-->





#53

t-1,3-Dichloropropene

Concen: 19.414 ug/l

RT: 9.02 min Scan# 1301

Delta R.T. -0.01 min

Lab File: VX016654.D

Acq: 09 Jun 2020 10:59

Instrument :

MSVOA\_X

ClientSampleId :





VX0609WBS01

Tot Ion: 75 Resp: 77669

Ion Ratio Lower Upper

75 100

77 33.7 25.5 38.3



#54

cis-1,3-Dichloropropene

Concen: 19.365 ug/l

RT: 8.42 min Scan# 1202

Delta R.T. -0.00 min

Lab File: VX016654.D

Acq: 09 Jun 2020 10:59




Tgt Ion: 75 Resp: 82002


Ion Ratio Lower Upper

75 100

77 31.3 24.6 36.8

39 41.1 35.3 52.9





#55

1,1,2-Trichloroethane

Concen: 20.068 ug/l

RT: 9.20 min Scan# 1330

Delta R.T. -0.00 min

Lab File: VX016654.D

Acq: 09 Jun 2020 10:59

Instrument :

MSVOA\_X

ClientSampled :

VX0609WBS01

Tot Ion: 97 Resp: 50000


Ion Ratio Lower Upper

97 100

83 79.9 70.9 106.3

85 53.7 43.4 65.2

99 61.3 51.0 76.6






Abundance Ion 96.90 (96.60 to 97.60): VX0

Ion 82.90 (82.60 to 83.60): VX0

Ion 84.90 (84.60 to 85.60): VX0

Ion 98.90 (98.60 to 99.60): VX0



#56

Ethyl methacrylate

Concen: 18.007 ug/l

RT: 9.16 min Scan# 1324

Delta R.T. -0.00 min

Lab File: VX016654.D

Acq: 09 Jun 2020 10:59


Tgt Ion: 69 Resp: 73065

Ion Ratio Lower Upper

69 100

41 59.2 50.7 76.1



39 31.7 25.8 38.6




Abundance Ion 69.00 (68.70 to 69.70): VX0

Ion 41.00 (40.70 to 41.70): VX0

Ion 39.00 (38.70 to 39.70): VX0











#57  
 1,3-Dichloropropane  
 Concen: 19.017 ug/l  
 RT: 9.35 min Scan# 1356  
 Delta R.T. -0.00 min  
 Lab File: VX016654.D  
 Acq: 09 Jun 2020 10:59




Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 VX0609WBS01


Tot Ion: 76 Resp: 81328  
 Ion Ratio Lower Upper  
 76 100  
 78 32.6 26.2 39.4



#58  
 2-Chloroethyl Vinyl ether  
 Concen: 79.854 ug/l  
 RT: 8.29 min Scan# 1181  
 Delta R.T. -0.01 min  
 Lab File: VX016654.D  
 Acq: 09 Jun 2020 10:59

Tgt Ion: 63 Resp: 169127  
 Ion Ratio Lower Upper  
 63 100  
 106 31.7 23.2 34.8









#59  
2-Hexanone  
Concen: 84.356 ug/l  
RT: 9.47 min Scan# 1375  
Delta R.T. -0.01 min  
Lab File: VX016654.D  
Acq: 09 Jun 2020 10:59




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0609WBS01


Tot Ion: 43 Resp: 255251  
Ion Ratio Lower Upper  
43 100  
58 59.3 28.8 86.4



#60  
Dibromochloromethane  
Concen: 20.386 ug/l  
RT: 9.56 min Scan# 1390  
Delta R.T. -0.01 min  
Lab File: VX016654.D  
Acq: 09 Jun 2020 10:59

Tgt Ion: 129 Resp: 57158  
Ion Ratio Lower Upper  
129 100  
127 79.4 38.6 116.0









#61  
 1,2-Dibromoethane  
 Concen: 19.723 ug/l  
 RT: 9.65 min Scan# 1405  
 Delta R.T. -0.00 min  
 Lab File: VX016654.D  
 Acq: 09 Jun 2020 10:59




Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 VX0609WBS01


Tot Ion:107 Resp: 52621  
 Ion Ratio Lower Upper  
 107 100  
 109 94.8 75.5 113.3



#62  
 4-Bromofluorobenzene  
 Concen: 51.174 ug/l  
 RT: 11.12 min Scan# 1646  
 Delta R.T. -0.00 min  
 Lab File: VX016654.D  
 Acq: 09 Jun 2020 10:59

Tgt Ion: 95 Resp: 167017  
 Ion Ratio Lower Upper  
 95 100  
 174 82.9 0.0 163.0  
 176 83.5 0.0 156.8









#63  
Chlorobenzene-d5  
Concen: 50.000 ug/l  
RT: 10.10 min Scan# 1478  
Delta R.T. -0.00 min  
Lab File: VX016654.D  
Acq: 09 Jun 2020 10:59




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0609WBS01


Tot Ion:117 Resp: 326960  
Ion Ratio Lower Upper  
117 100  
82 54.0 45.0 67.6  
119 31.9 25.8 38.8



#64  
Tetrachloroethene  
Concen: 20.002 ug/l  
RT: 9.32 min Scan# 1350  
Delta R.T. -0.00 min  
Lab File: VX016654.D  
Acq: 09 Jun 2020 10:59

Tgt Ion:164 Resp: 73168  
Ion Ratio Lower Upper  
164 100  
166 135.8 100.9 151.3  
129 92.9 73.2 109.8  
131 88.6 70.8 106.2





#65

Chlorobenzene

Concen: 19.735 ug/l

RT: 10.12 min Scan# 1482

Delta R.T. -0.00 min

Lab File: VX016654.D

Acq: 09 Jun 2020 10:59

Instrument :

MSVOA\_X

ClientSampled :


VX0609WBS01

Tot Ion:112 Resp: 131549

Ion Ratio Lower Upper




112 100

114 33.1 26.0 39.0



Abundance Ion 111.90 (111.60 to 112.60): V

Ion 113.90 (113.60 to 114.60): V



#66

1,1,1,2-Tetrachloroethane

Concen: 20.415 ug/l

RT: 10.20 min Scan# 1495

Delta R.T. -0.01 min

Lab File: VX016654.D

Acq: 09 Jun 2020 10:59


Tgt Ion:131 Resp: 51947

Ion Ratio Lower Upper

131 100

133 97.1 48.4 145.3



119 63.0 32.0 96.2




Abundance Ion 130.80 (130.50 to 131.50): V

Ion 132.80 (132.50 to 133.50): V

Ion 118.80 (118.50 to 119.50): V









#67  
Ethyl Benzene  
Concen: 18.921 ug/l  
RT: 10.23 min Scan# 1500  
Delta R.T. -0.01 min  
Lab File: VX016654.D  
Acq: 09 Jun 2020 10:59




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0609WBS01


Tot Ion: 91 Resp: 224558  
Ion Ratio Lower Upper  
91 100  
106 30.4 25.0 37.4



#68  
m/p-Xylenes  
Concen: 39.629 ug/l  
RT: 10.34 min Scan# 1518  
Delta R.T. -0.00 min  
Lab File: VX016654.D  
Acq: 09 Jun 2020 10:59

Tgt Ion: 106 Resp: 174379  
Ion Ratio Lower Upper  
106 100  
91 200.6 163.8 245.8









#69  
o-Xylene  
Concen: 19.167 ug/l  
RT: 10.68 min Scan# 1574  
Delta R.T. -0.00 min  
Lab File: VX016654.D  
Acq: 09 Jun 2020 10:59




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0609WBS01


Tot Ion:106 Resp: 81681  
Ion Ratio Lower Upper  
106 100  
91 216.6 109.6 328.8



#70  
Styrene  
Concen: 19.434 ug/l  
RT: 10.70 min Scan# 1576  
Delta R.T. -0.00 min  
Lab File: VX016654.D  
Acq: 09 Jun 2020 10:59

Tgt Ion:104 Resp: 142353  
Ion Ratio Lower Upper  
104 100  
78 50.6 40.2 60.2  
103 54.0 43.1 64.7





#71

Bromoform

Concen: 20.703 ug/l

RT: 10.84 min Scan# 1600

Delta R.T. -0.00 min

Lab File: VX016654.D

Acq: 09 Jun 2020 10:59

Instrument :

MSVOA\_X

ClientSampleId :

VX0609WBS01


Tot Ion:173 Resp: 44520

Ion Ratio Lower Upper




173 100

175 48.8 23.9 71.7

254 0.2 0.1 0.1#



Abundance Ion 172.70 (172.40 to 173.40): V  
Ion 174.70 (174.40 to 175.40): V  
Ion 254.40 (254.10 to 255.10): V



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.06 min Scan# 1800

Delta R.T. -0.00 min

Lab File: VX016654.D

Acq: 09 Jun 2020 10:59


Tgt Ion:152 Resp: 164387

Ion Ratio Lower Upper



152 100

115 66.2 39.9 119.6


150 164.3 0.0 341.0



Abundance Ion 151.90 (151.60 to 152.60): V  
Ion 114.90 (114.60 to 115.60): V  
Ion 149.90 (149.60 to 150.60): V







#73

Isopropylbenzene

Concen: 19.566 ug/l

RT: 11.00 min Scan# 1626

Delta R.T. -0.00 min

Lab File: VX016654.D

Acq: 09 Jun 2020 10:59

Instrument :

MSVOA\_X

ClientSampleId :





VX0609WBS01

Tot Ion:105 Resp: 222855

Ion Ratio Lower Upper

105 100

120 26.4 13.1 39.3



#74

N-ethyl acetate

Concen: 16.152 ug/l

RT: 10.88 min Scan# 1606

Delta R.T. -0.00 min

Lab File: VX016654.D

Acq: 09 Jun 2020 10:59

Tgt Ion: 43 Resp: 84620




Ion Ratio Lower Upper


43 100

70 49.6 36.1 54.1

55 27.2 20.2 30.4

61 27.4 20.3 30.5





#75

1,1,2,2-Tetrachloroethane

Concen: 18.943 ug/l

RT: 11.25 min Scan# 1667

Delta R.T. -0.01 min

Lab File: VX016654.D

Acq: 09 Jun 2020 10:59

Instrument :

MSVOA\_X

ClientSampled :

VX0609WBS01


Tot Ion: 83 Resp: 54075

Ion Ratio Lower Upper

83 100

131 11.1 5.1 15.3




85 65.0 32.0 96.2



Abundance Ion 82.90 (82.60 to 83.60): VX0

Ion 130.80 (130.50 to 131.50): VX0

Ion 84.90 (84.60 to 85.60): VX0



#76

1,2,3-Trichloropropane

Concen: 24.685 ug/l

RT: 11.28 min Scan# 1672

Delta R.T. -0.00 min

Lab File: VX016654.D


Acq: 09 Jun 2020 10:59

Tgt Ion: 75 Resp: 89193

Ion Ratio Lower Upper



75 100


77 32.1 21.3 63.7



Abundance Ion 74.90 (74.60 to 75.60): VX0

Ion 77.00 (76.70 to 77.70): VX0





#77

Bromobenzene

Concen: 19.739 ug/l

RT: 11.24 min Scan# 1665

Delta R.T. -0.00 min

Lab File: VX016654.D

Acq: 09 Jun 2020 10:59

Instrument :

MSVOA\_X

ClientSampleId :

VX0609WBS01


Tot Ion:156 Resp: 58968

Ion Ratio Lower Upper

156 100

77 149.5 79.6 238.8




158 97.4 48.1 144.4



Abundance Ion 155.80 (155.50 to 156.50): V

Ion 77.00 (76.70 to 77.70): VX0

Ion 157.80 (157.50 to 158.50): V



#78

n-propylbenzene

Concen: 19.062 ug/l

RT: 11.34 min Scan# 1682

Delta R.T. -0.00 min

Lab File: VX016654.D


Acq: 09 Jun 2020 10:59

Tgt Ion: 91 Resp: 242567

Ion Ratio Lower Upper



91 100


120 23.8 11.6 34.7



Abundance Ion 91.00 (90.70 to 91.70): VX0

Ion 120.00 (119.70 to 120.70): V





#79

2-Chlorotoluene

Concen: 19.404 ug/l

RT: 11.40 min Scan# 1692

Delta R.T. -0.00 min

Lab File: VX016654.D

Acq: 09 Jun 2020 10:59

Instrument :

MSVOA\_X

ClientSampleId :





VX0609WBS01

Tot Ion: 91 Resp: 150715

Ion Ratio Lower Upper

91 100

126 33.4 16.8 50.3



#80

1,3,5-Trimethylbenzene

Concen: 20.081 ug/l

RT: 11.49 min Scan# 1706

Delta R.T. -0.00 min

Lab File: VX016654.D




Acq: 09 Jun 2020 10:59


Tgt Ion: 105 Resp: 190983

Ion Ratio Lower Upper

105 100

120 48.4 24.3 72.8





#81

trans-1,4-Dichloro-2-butene

Concen: 18.417 ug/l

RT: 11.06 min Scan# 1636

Delta R.T. -0.00 min

Lab File: VX016654.D

Acq: 09 Jun 2020 10:59

Instrument :

MSVOA\_X

ClientSampleId :

VX0609WBS01


Tot Ion: 75 Resp: 27053

Ion Ratio Lower Upper

75 100

53 86.1 71.3 106.9




89 49.4 36.6 55.0



Abundance Ion 74.90 (74.60 to 75.60): VX0

Ion 53.00 (52.70 to 53.70): VX0

Ion 88.90 (88.60 to 89.60): VX0



#82

4-Chlorotoluene

Concen: 19.335 ug/l

RT: 11.49 min Scan# 1707

Delta R.T. -0.00 min

Lab File: VX016654.D


Acq: 09 Jun 2020 10:59

Tgt Ion: 91 Resp: 177354

Ion Ratio Lower Upper



91 100


126 29.8 14.7 44.1



Abundance Ion 91.00 (90.70 to 91.70): VX0

Ion 125.90 (125.60 to 126.60): V









#83  
 tert-Butylbenzene  
 Concen: 19.700 ug/l  
 RT: 11.76 min Scan# 1750  
 Delta R.T. -0.00 min  
 Lab File: VX016654.D  
 Acq: 09 Jun 2020 10:59




Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 VX0609WBS01


| Tot Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 119     | 100   |       |       |
| 91      | 62.0  | 31.3  | 93.8  |
| 134     | 24.0  | 12.0  | 36.0  |



#84  
 1,2,4-Trimethylbenzene  
 Concen: 19.263 ug/l  
 RT: 11.79 min Scan# 1756  
 Delta R.T. -0.00 min  
 Lab File: VX016654.D  
 Acq: 09 Jun 2020 10:59

| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 105     | 100   |       |       |
| 120     | 45.9  | 22.6  | 67.7  |





#85

sec-Butylbenzene

Concen: 19.606 ug/l

RT: 11.93 min Scan# 1779

Delta R.T. -0.00 min

Lab File: VX016654.D

Acq: 09 Jun 2020 10:59

Instrument :

MSVOA\_X

ClientSampleId :


VX0609WBS01

Tot Ion:105 Resp: 199868

Ion Ratio Lower Upper




105 100

134 20.7 10.1 30.3



Abundance Ion 105.00 (104.70 to 105.70): V

Ion 134.00 (133.70 to 134.70): V



#86

p-Isopropyltoluene

Concen: 19.955 ug/l

RT: 12.05 min Scan# 1798

Delta R.T. -0.00 min

Lab File: VX016654.D

Acq: 09 Jun 2020 10:59


Tgt Ion:119 Resp: 190553

Ion Ratio Lower Upper

119 100

134 25.8 12.7 38.0



91 24.2 11.7 35.0




Abundance Ion 119.00 (118.70 to 119.70): V

Ion 134.00 (133.70 to 134.70): V

Ion 91.00 (90.70 to 91.70): VX0





#87

1,3-Dichlorobenzene

Concen: 19.752 ug/l

RT: 12.01 min Scan# 1792





Delta R.T. -0.00 min

Lab File: VX016654.D

Acq: 09 Jun 2020 10:59

Instrument :  
MSVOA\_X  
ClientSampled :  
VX0609WBS01

| Tot Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 146     | 100   |       |       |
| 111     | 39.3  | 20.0  | 60.0  |
| 148     | 64.3  | 31.9  | 95.5  |



#88

1,4-Dichlorobenzene

Concen: 19.858 ug/l




RT: 12.08 min Scan# 1803

Delta R.T. -0.00 min


Lab File: VX016654.D

Acq: 09 Jun 2020 10:59

| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 146     | 100   |       |       |
| 111     | 40.5  | 19.4  | 58.2  |
| 148     | 64.5  | 31.4  | 94.3  |











#89  
n-Butylbenzene  
Concen: 19.307 ug/l  
RT: 12.37 min Scan# 1851  
Delta R.T. -0.00 min  
Lab File: VX016654.D  
Acq: 09 Jun 2020 10:59




Instrument :  
MSVOA\_X  
ClientSampled :  
VX0609WBS01


Tot Ion: 91 Resp: 153564  
Ion Ratio Lower Upper  
91 100  
92 54.4 27.2 81.5  
134 27.3 13.1 39.3



#90  
Hexachloroethane  
Concen: 20.690 ug/l  
RT: 12.58 min Scan# 1885  
Delta R.T. -0.00 min  
Lab File: VX016654.D  
Acq: 09 Jun 2020 10:59

Tgt Ion: 117 Resp: 34127  
Ion Ratio Lower Upper  
117 100  
201 89.5 43.1 129.3





#91

1,2-Dichlorobenzene

Concen: 19.670 ug/l

RT: 12.38 min Scan# 1852

Delta R.T. -0.00 min

Lab File: VX016654.D

Acq: 09 Jun 2020 10:59

Instrument :

MSVOA\_X

ClientSampleId :

VX0609WBS01


Tot Ion:146 Resp: 98633

Ion Ratio Lower Upper

146 100

111 40.9 21.4 64.3




148 64.0 31.9 95.7



Abundance Ion 145.90 (145.60 to 146.60): V

Ion 110.90 (110.60 to 111.60): V

Ion 147.90 (147.60 to 148.60): V



#92

1,2-Dibromo-3-Chloropropane

Concen: 18.259 ug/l

RT: 12.98 min Scan# 1951

Delta R.T. -0.00 min

Lab File: VX016654.D

Acq: 09 Jun 2020 10:59


Tgt Ion: 75 Resp: 16883

Ion Ratio Lower Upper

75 100

155 92.3 44.1 132.4



157 116.4 57.8 173.3




Abundance Ion 74.90 (74.60 to 75.60): VX0

Ion 154.80 (154.50 to 155.50): V

Ion 156.80 (156.50 to 157.50): V









#93  
 1,2,4-Trichlorobenzene  
 Concen: 19.466 ug/l  
 RT: 13.63 min Scan# 2057  
 Delta R.T. -0.00 min  
 Lab File: VX016654.D  
 Acq: 09 Jun 2020 10:59




Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 VX0609WBS01


Tot Ion:180 Resp: 64827  
 Ion Ratio Lower Upper  
 180 100  
 182 100.0 48.3 144.8  
 145 31.6 15.4 46.2



#94  
 Hexachlorobutadiene  
 Concen: 25.208 ug/l  
 RT: 13.77 min Scan# 2080  
 Delta R.T. -0.00 min  
 Lab File: VX016654.D  
 Acq: 09 Jun 2020 10:59

Tgt Ion:225 Resp: 32960  
 Ion Ratio Lower Upper  
 225 100  
 223 61.3 32.3 96.8  
 227 64.9 31.6 94.7





#95

Naphthalene

Concen: 18.899 ug/l

RT: 13.82 min Scan# 2088

Delta R.T. -0.00 min

Lab File: VX016654.D

Acq: 09 Jun 2020 10:59

Instrument :

MSVOA\_X

ClientSampleId :

VX0609WBS01


Tot Ion:128 Resp: 215758

Ion Ratio Lower Upper

128 100

127 13.2 10.6 15.8




129 10.9 8.6 13.0



Abundance Ion 128.00 (127.70 to 128.70): V

Ion 127.00 (126.70 to 127.70): V

Ion 129.00 (128.70 to 129.70): V



#96

1,2,3-Trichlorobenzene

Concen: 20.977 ug/l

RT: 14.00 min Scan# 2118

Delta R.T. -0.01 min

Lab File: VX016654.D

Acq: 09 Jun 2020 10:59


Tgt Ion:180 Resp: 68440

Ion Ratio Lower Upper

180 100

182 95.2 48.6 145.8

145 31.8 16.6 49.8



Abundance Ion 179.80 (179.50 to 180.50): V

Ion 181.80 (181.50 to 182.50): V

Ion 144.90 (144.60 to 145.60): V

