

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060921\
 Data File : VX022609.D
 Acq On : 09 Jun 2021 13:26
 Operator : JC/MD
 Sample : M2644-08 100X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 1824-25

Quant Time: Jun 10 05:41:31 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060721W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 07 14:27:50 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	56248	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	113301	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	99549	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	37375	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	47896	47.899	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.800%
35) Dibromofluoromethane	5.397	113	33660	47.483	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.960%
50) Toluene-d8	8.653	98	140224	49.576	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.160%
62) 4-Bromofluorobenzene	11.085	95	46233	43.975	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	87.960%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	5345	11.896	ug/l	97
52) Toluene	8.720	92	13878	6.533	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060921\
Data File : VX022609.D
Acq On : 09 Jun 2021 13:26
Operator : JC/MD
Sample : M2644-08 100X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
1824-25

Quant Time: Jun 10 05:41:31 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060721W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 07 14:27:50 2021
Response via : Initial Calibration

