

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060921\
 Data File : VX022610.D
 Acq On : 09 Jun 2021 13:50
 Operator : JC/MD
 Sample : M2644-09 100X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 1847

Quant Time: Jun 10 05:41:48 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060721W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 07 14:27:50 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	56188	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	111387	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	99612	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	39649	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	47044	47.097	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	94.200%
35) Dibromofluoromethane	5.397	113	32973	47.313	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.620%
50) Toluene-d8	8.653	98	137463	49.435	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.860%
62) 4-Bromofluorobenzene	11.085	95	48411	46.838	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.680%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060921\
Data File : VX022610.D
Acq On : 09 Jun 2021 13:50
Operator : JC/MD
Sample : M2644-09 100X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
1847

Quant Time: Jun 10 05:41:48 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060721W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 07 14:27:50 2021
Response via : Initial Calibration

