

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060921\
 Data File : VX022612.D
 Acq On : 09 Jun 2021 14:37
 Operator : JC/MD
 Sample : M2652-09
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB05-20210608

Quant Time: Jun 10 05:42:23 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060721W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 07 14:27:50 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	57553	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	115693	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	100651	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	38852	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	49329	48.213	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.420%
35) Dibromofluoromethane	5.391	113	34098	47.106	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.220%
50) Toluene-d8	8.653	98	143184	49.575	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.160%
62) 4-Bromofluorobenzene	11.085	95	48181	44.881	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.760%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060921\
Data File : VX022612.D
Acq On : 09 Jun 2021 14:37
Operator : JC/MD
Sample : M2652-09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB05-20210608

Quant Time: Jun 10 05:42:23 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060721W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 07 14:27:50 2021
Response via : Initial Calibration

