

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060921\
 Data File : VX022613.D
 Acq On : 09 Jun 2021 15:00
 Operator : JC/MD
 Sample : M2652-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RB09-20210608

Quant Time: Jun 10 05:42:37 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060721W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 07 14:27:50 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	56780	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	115377	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	104005	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	40085	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	48796	48.342	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.680%
35) Dibromofluoromethane	5.397	113	34505	47.799	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.600%
50) Toluene-d8	8.653	98	142469	49.463	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.920%
62) 4-Bromofluorobenzene	11.085	95	50422	47.097	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.200%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	2394	5.278	ug/l	96
20) Methylene Chloride	2.794	84	1112	1.174	ug/l	# 83
52) Toluene	8.726	92	1145	0.529	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060921\
Data File : VX022613.D
Acq On : 09 Jun 2021 15:00
Operator : JC/MD
Sample : M2652-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RB09-20210608

Quant Time: Jun 10 05:42:37 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060721W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 07 14:27:50 2021
Response via : Initial Calibration

