

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060921\
 Data File : VX022616.D
 Acq On : 09 Jun 2021 16:06
 Operator : JC/MD
 Sample : M2652-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 40BR-20210608

Quant Time: Jun 10 05:43:24 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060721W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 07 14:27:50 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	56914	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	115514	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	101848	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	39043	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	48294	47.732	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.460%
35) Dibromofluoromethane	5.391	113	34862	48.236	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.480%
50) Toluene-d8	8.653	98	141458	49.054	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.100%
62) 4-Bromofluorobenzene	11.085	95	48785	45.514	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.020%
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.172	85	1503	2.264	ug/l	93
4) Vinyl Chloride	1.374	62	19086	24.394	ug/l	94
16) Acetone	2.386	43	2182	4.799	ug/l	97
21) trans-1,2-Dichloroethene	3.093	96	1442	1.774	ug/l	88
27) cis-1,2-Dichloroethene	4.495	96	220208	230.770	ug/l	92
37) Ethyl Acetate	4.727	43	4273	2.981	ug/l	# 68
44) Trichloroethene	7.129	130	2939	3.612	ug/l	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060921\
Data File : VX022616.D
Acq On : 09 Jun 2021 16:06
Operator : JC/MD
Sample : M2652-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
40BR-20210608

Quant Time: Jun 10 05:43:24 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060721W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 07 14:27:50 2021
Response via : Initial Calibration

