

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060921\
 Data File : VX022618.D
 Acq On : 09 Jun 2021 16:53
 Operator : JC/MD
 Sample : M2652-02 100X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 7BR-20210608

Quant Time: Jun 10 05:43:57 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060721W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 07 14:27:50 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	56709	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	115351	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	105153	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	41508	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	48632	48.239	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.480%
35) Dibromofluoromethane	5.397	113	34330	47.568	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.140%
50) Toluene-d8	8.653	98	142661	49.541	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.080%
62) 4-Bromofluorobenzene	11.085	95	50914	47.567	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.140%
Target Compounds						
					Qvalue	
4) Vinyl Chloride	1.374	62	5256	6.742	ug/l	91
21) trans-1,2-Dichloroethene	3.093	96	1324	1.635	ug/l	88
27) cis-1,2-Dichloroethene	4.495	96	117569	123.653	ug/l	93
44) Trichloroethene	7.135	130	7686	9.459	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060921\
Data File : VX022618.D
Acq On : 09 Jun 2021 16:53
Operator : JC/MD
Sample : M2652-02 100X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
7BR-20210608

Quant Time: Jun 10 05:43:57 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060721W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 07 14:27:50 2021
Response via : Initial Calibration

