

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060921\
 Data File : VX022619.D
 Acq On : 09 Jun 2021 17:17
 Operator : JC/MD
 Sample : M2652-04 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BRP2-20210608

Quant Time: Jun 10 05:44:17 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060721W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 07 14:27:50 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	56933	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	115883	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	103838	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	40396	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	49356	48.765	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.520%
35) Dibromofluoromethane	5.391	113	35615	49.121	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.240%
50) Toluene-d8	8.653	98	144929	50.097	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.200%
62) 4-Bromofluorobenzene	11.085	95	48938	45.511	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.020%
Target Compounds						
						Qvalue
4) Vinyl Chloride	1.374	62	6509	8.316	ug/l	100
9) 1,1,2-Trichlorotrifluo...	2.331	101	1832	2.621	ug/l	92
12) 1,1-Dichloroethene	2.325	96	1680	2.327	ug/l	94
21) trans-1,2-Dichloroethene	3.099	96	6380	7.847	ug/l	92
27) cis-1,2-Dichloroethene	4.495	96	346987	363.508	ug/l	92
44) Trichloroethene	7.135	130	142455	174.514	ug/l	94
52) Toluene	8.720	92	1621	0.746	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060921\
Data File : VX022619.D
Acq On : 09 Jun 2021 17:17
Operator : JC/MD
Sample : M2652-04 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BRP2-20210608

Quant Time: Jun 10 05:44:17 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060721W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 07 14:27:50 2021
Response via : Initial Calibration

