

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060921\
 Data File : VX022620.D
 Acq On : 09 Jun 2021 17:40
 Operator : JC/MD
 Sample : M2652-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 60BR-20210608

Quant Time: Jun 10 05:44:36 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060721W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 07 14:27:50 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.568	168	54658	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	111340	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	97968	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	35279	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	46918	48.286	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.580%
35) Dibromofluoromethane	5.391	113	33271	47.761	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.520%
50) Toluene-d8	8.653	98	139005	50.010	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.020%
62) 4-Bromofluorobenzene	11.085	95	45208	43.758	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	87.520%
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.172	85	2368	3.715	ug/l	95
16) Acetone	2.392	43	2214	5.071	ug/l	91
27) cis-1,2-Dichloroethene	4.501	96	77308	84.360	ug/l	94
37) Ethyl Acetate	4.727	43	4412	3.193	ug/l #	75
44) Trichloroethene	7.135	130	458122	584.122	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060921\
Data File : VX022620.D
Acq On : 09 Jun 2021 17:40
Operator : JC/MD
Sample : M2652-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
60BR-20210608

Quant Time: Jun 10 05:44:36 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060721W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 07 14:27:50 2021
Response via : Initial Calibration

