

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060921\
 Data File : VX022623.D
 Acq On : 09 Jun 2021 18:51
 Operator : JC/MD
 Sample : PB136943TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB136943TB

Quant Time: Jun 10 05:45:30 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060721W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 07 14:27:50 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	56009	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	116096	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	108610	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	42096	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	49379	49.593	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	99.180%
35) Dibromofluoromethane	5.397	113	34573	47.597	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.200%
50) Toluene-d8	8.653	98	146926	50.694	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.380%
62) 4-Bromofluorobenzene	11.085	95	52168	48.426	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.860%
Target Compounds						
						Qvalue
16) Acetone	2.386	43	6772	15.136	ug/l	# 81
20) Methylene Chloride	2.794	84	7195	7.702	ug/l	91
25) 2-Butanone	4.568	43	4836	6.540	ug/l	# 80
43) Isopropyl Acetate	6.355	43	9589	4.103	ug/l	# 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060921\
Data File : VX022623.D
Acq On : 09 Jun 2021 18:51
Operator : JC/MD
Sample : PB136943TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB136943TB

Quant Time: Jun 10 05:45:30 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060721W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 07 14:27:50 2021
Response via : Initial Calibration

