

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060921\
 Data File : VX022625.D
 Acq On : 09 Jun 2021 19:38
 Operator : JC/MD
 Sample : PB136928TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB136928TB

Quant Time: Jun 10 05:46:00 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060721W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 07 14:27:50 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	55216	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	112319	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	101724	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	35773	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	47906	48.804	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.600%
35) Dibromofluoromethane	5.397	113	33807	48.107	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.220%
50) Toluene-d8	8.653	98	143036	51.012	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	102.020%
62) 4-Bromofluorobenzene	11.085	95	46809	44.912	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.820%
Target Compounds						
16) Acetone	2.386	43	4800	10.882	ug/l	Qvalue 97
20) Methylene Chloride	2.794	84	3691	4.008	ug/l	# 83
43) Isopropyl Acetate	6.354	43	11251	4.977	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060921\
Data File : VX022625.D
Acq On : 09 Jun 2021 19:38
Operator : JC/MD
Sample : PB136928TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB136928TB

Quant Time: Jun 10 05:46:00 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060721W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 07 14:27:50 2021
Response via : Initial Calibration

