

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060921\
 Data File : VX022604.D
 Acq On : 09 Jun 2021 11:13
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 10 05:38:40 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060721W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 07 14:27:50 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	105	0.00
2 T	Dichlorodifluoromethane	0.583	0.626	-7.4	102	0.00
3 P	Chloromethane	0.704	0.740	-5.1	111	0.00
4 C	Vinyl Chloride	0.687	0.668	2.8#	101	0.00
5 T	Bromomethane	0.361	0.333	7.8	97	0.00
6 T	Chloroethane	0.433	0.412	4.8	100	0.00
7 T	Trichlorofluoromethane	1.001	0.986	1.5	101	0.00
8 T	Diethyl Ether	0.408	0.392	3.9	102	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.614	0.615	-0.2	104	0.00
10 T	Methyl Iodide	0.722	0.607	15.9	87	0.00
11 T	Tert butyl alcohol	0.199	0.178	10.6	97	0.00
12 CM	1,1-Dichloroethene	0.634	0.600	5.4#	101	0.00
13 T	Acrolein	0.101	0.080	20.8	95	0.00
14 T	Allyl chloride	1.163	1.141	1.9	101	0.00
15 T	Acrylonitrile	0.416	0.386	7.2	98	0.00
16 T	Acetone	0.399	0.396	0.8	109	0.00
17 T	Carbon Disulfide	1.531	1.348	12.0	99	0.00
18 T	Methyl Acetate	1.101	1.026	6.8	99	0.00
19 T	Methyl tert-butyl Ether	2.114	2.067	2.2	102	0.00
20 T	Methylene Chloride	0.834	0.741	11.2	102	0.00
21 T	trans-1,2-Dichloroethene	0.714	0.676	5.3	102	0.00
22 T	Diisopropyl ether	2.332	2.283	2.1	104	0.00
23 T	Vinyl Acetate	2.195	2.159	1.6	101	0.00
24 P	1,1-Dichloroethane	1.327	1.288	2.9	102	0.00
25 T	2-Butanone	0.660	0.612	7.3	99	0.00
26 T	2,2-Dichloropropane	1.017	1.058	-4.0	105	0.00
27 T	cis-1,2-Dichloroethene	0.838	0.795	5.1	101	0.00
28 T	Bromochloromethane	0.671	0.619	7.7	104	0.00
29 T	Tetrahydrofuran	0.387	0.352	9.0	95	0.00
30 C	Chloroform	1.333	1.281	3.9#	101	0.00
31 T	Cyclohexane	1.051	1.061	-1.0	102	0.00
32 T	1,1,1-Trichloroethane	1.057	1.052	0.5	100	0.00
33 S	1,2-Dichloroethane-d4	0.889	0.802	9.8	103	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	101	0.00
35 S	Dibromofluoromethane	0.313	0.306	2.2	102	0.00
36 T	1,1-Dichloropropene	0.484	0.504	-4.1	102	0.00
37 T	Ethyl Acetate	0.620	0.611	1.5	99	0.00
38 T	Carbon Tetrachloride	0.406	0.438	-7.9	100	0.00
39 T	Methylcyclohexane	0.523	0.557	-6.5	104	0.00
40 TM	Benzene	1.493	1.523	-2.0	101	0.00
41 T	Methacrylonitrile	0.341	0.334	2.1	102	0.00
42 TM	1,2-Dichloroethane	0.550	0.543	1.3	99	0.00
43 T	Isopropyl Acetate	1.006	1.013	-0.7	100	0.00
44 TM	Trichloroethene	0.352	0.353	-0.3	100	0.00
45 C	1,2-Dichloropropane	0.384	0.400	-4.2#	103	0.00
46 T	Dibromomethane	0.262	0.258	1.5	101	0.00
47 T	Bromodichloromethane	0.454	0.475	-4.6	98	0.00
48 T	Methyl methacrylate	0.468	0.485	-3.6	103	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060921\
 Data File : VX022604.D
 Acq On : 09 Jun 2021 11:13
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 10 05:38:40 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060721W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 07 14:27:50 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.009	0.009	0.0	94	0.00
50 S	Toluene-d8	1.248	1.229	1.5	104	0.00
51 T	4-Methyl-2-Pentanone	0.627	0.639	-1.9	99	0.00
52 CM	Toluene	0.938	0.957	-2.0#	104	0.00
53 T	t-1,3-Dichloropropene	0.536	0.581	-8.4	104	0.00
54 T	cis-1,3-Dichloropropene	0.576	0.628	-9.0	103	0.00
55 T	1,1,2-Trichloroethane	0.365	0.387	-6.0	103	0.00
56 T	Ethyl methacrylate	0.602	0.634	-5.3	102	0.00
57 T	1,3-Dichloropropane	0.660	0.662	-0.3	99	0.00
58 T	2-Chloroethyl Vinyl ether	0.353	0.340	3.7	101	0.00
59 T	2-Hexanone	0.479	0.492	-2.7	100	0.00
60 T	Dibromochloromethane	0.304	0.335	-10.2	102	0.00
61 T	1,2-Dibromoethane	0.385	0.384	0.3	99	0.00
62 S	4-Bromofluorobenzene	0.464	0.463	0.2	110	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	104	0.00
64 T	Tetrachloroethene	0.335	0.326	2.7	98	0.00
65 PM	Chlorobenzene	1.039	1.058	-1.8	104	0.00
66 T	1,1,1,2-Tetrachloroethane	0.328	0.347	-5.8	100	0.00
67 C	Ethyl Benzene	1.912	1.991	-4.1#	104	0.00
68 T	m/p-Xylenes	0.695	0.728	-4.7	104	0.00
69 T	o-Xylene	0.679	0.709	-4.4	103	0.00
70 T	Styrene	1.149	1.210	-5.3	104	0.00
71 P	Bromoform	0.194	0.209	-7.7	100	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	104	0.00
73 T	Isopropylbenzene	4.075	4.238	-4.0	104	0.00
74 T	N-amyl acetate	2.186	2.238	-2.4	103	0.00
75 P	1,1,2,2-Tetrachloroethane	1.467	1.487	-1.4	104	0.00
76 T	1,2,3-Trichloropropane	1.295	1.369	-5.7	105	0.00
77 T	Bromobenzene	0.878	0.894	-1.8	106	0.00
78 T	n-propylbenzene	4.946	5.164	-4.4	106	0.00
79 T	2-Chlorotoluene	2.943	3.062	-4.0	108	0.00
80 T	1,3,5-Trimethylbenzene	3.482	3.612	-3.7	106	0.00
81 T	trans-1,4-Dichloro-2-butene	0.449	0.456	-1.6	101	0.00
82 T	4-Chlorotoluene	3.407	3.538	-3.8	106	0.00
83 T	tert-Butylbenzene	3.102	3.258	-5.0	104	0.00
84 T	1,2,4-Trimethylbenzene	3.502	3.649	-4.2	105	0.00
85 T	sec-Butylbenzene	4.128	4.391	-6.4	106	0.00
86 T	p-Isopropyltoluene	3.412	3.677	-7.8	107	0.00
87 T	1,3-Dichlorobenzene	1.705	1.752	-2.8	107	0.00
88 T	1,4-Dichlorobenzene	1.776	1.749	1.5	106	0.00
89 T	n-Butylbenzene	3.276	3.589	-9.6	108	0.00
90 T	Hexachloroethane	0.477	0.511	-7.1	102	0.00
91 T	1,2-Dichlorobenzene	1.618	1.677	-3.6	106	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.269	0.294	-9.3	100	0.00
93 T	1,2,4-Trichlorobenzene	0.973	0.996	-2.4	108	0.00
94 T	Hexachlorobutadiene	0.349	0.366	-4.9	109	0.00
95 T	Naphthalene	3.900	4.039	-3.6	102	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060921\
Data File : VX022604.D
Acq On : 09 Jun 2021 11:13
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Jun 10 05:38:40 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060721W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 07 14:27:50 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.972	1.015	-4.4	106	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6