

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX060921\  
 Data File : VX022607.D  
 Acq On : 09 Jun 2021 12:36  
 Operator : JC/MD  
 Sample : VX0609WBS01  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 5 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 VX0609WBS01

Manual Integrations  
 APPROVED

MMDadoda  
 6/10/2021 12:11:56 PM

Quant Time: Jun 10 05:39:56 2021  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X060721W.M  
 Quant Title : SW846 8260  
 QLast Update : Mon Jun 07 14:27:50 2021  
 Response via : Initial Calibration

| Compound                     | R.T.   | QIon  | Response | Conc     | Units  | Dev(Min) |
|------------------------------|--------|-------|----------|----------|--------|----------|
| -----                        |        |       |          |          |        |          |
| Internal Standards           |        |       |          |          |        |          |
| 1) Pentafluorobenzene        | 5.562  | 168   | 75459    | 50.000   | ug/l   | 0.00     |
| 34) 1,4-Difluorobenzene      | 6.769  | 114   | 147983   | 50.000   | ug/l   | 0.00     |
| 63) Chlorobenzene-d5         | 10.055 | 117   | 134167   | 50.000   | ug/l   | 0.00     |
| 72) 1,4-Dichlorobenzene-d4   | 12.024 | 152   | 57935    | 50.000   | ug/l   | 0.00     |
| System Monitoring Compounds  |        |       |          |          |        |          |
| 33) 1,2-Dichloroethane-d4    | 5.964  | 65    | 59849    | 44.615   | ug/l   | 0.00     |
| Spiked Amount                | 50.000 | Range | 78 - 117 | Recovery | =      | 89.220%  |
| 35) Dibromofluoromethane     | 5.397  | 113   | 44329    | 47.878   | ug/l   | 0.00     |
| Spiked Amount                | 50.000 | Range | 75 - 124 | Recovery | =      | 95.760%  |
| 50) Toluene-d8               | 8.653  | 98    | 174952   | 47.357   | ug/l   | 0.00     |
| Spiked Amount                | 50.000 | Range | 92 - 112 | Recovery | =      | 94.720%  |
| 62) 4-Bromofluorobenzene     | 11.085 | 95    | 64267    | 46.802   | ug/l   | 0.00     |
| Spiked Amount                | 50.000 | Range | 83 - 123 | Recovery | =      | 93.600%  |
| Target Compounds             |        |       |          |          |        |          |
|                              |        |       |          |          | Qvalue |          |
| 2) Dichlorodifluoromethane   | 1.173  | 85    | 16828    | 19.123   | ug/l   | 100      |
| 3) Chloromethane             | 1.301  | 50    | 19372    | 18.227   | ug/l   | 95       |
| 4) Vinyl Chloride            | 1.380  | 62    | 18453    | 17.789   | ug/l   | 100      |
| 5) Bromomethane              | 1.599  | 94    | 11246    | 20.634   | ug/l   | 97       |
| 6) Chloroethane              | 1.685  | 64    | 11628    | 17.802   | ug/l   | 100      |
| 7) Trichlorofluoromethane    | 1.886  | 101   | 26754    | 17.709   | ug/l   | 100      |
| 8) Diethyl Ether             | 2.136  | 74    | 11009    | 17.874   | ug/l   | 90       |
| 9) 1,1,2-Trichlorotrifluo... | 2.331  | 101   | 16996    | 18.349   | ug/l   | 90       |
| 10) Methyl Iodide            | 2.453  | 142   | 15988    | 14.663   | ug/l   | 94       |
| 11) Tert butyl alcohol       | 2.977  | 59    | 25524    | 84.827   | ug/l   | 99       |
| 12) 1,1-Dichloroethene       | 2.319  | 96    | 15977    | 16.698   | ug/l   | 94       |
| 13) Acrolein                 | 2.239  | 56    | 10795    | 77.376   | ug/l   | 97       |
| 14) Allyl chloride           | 2.666  | 41    | 30653    | 17.471   | ug/l   | 93       |
| 15) Acrylonitrile            | 3.075  | 53    | 54612    | 87.058   | ug/l   | 100      |
| 16) Acetone                  | 2.386  | 43    | 50564    | 83.885   | ug/l   | 94       |
| 17) Carbon Disulfide         | 2.514  | 76    | 34677    | 16.888   | ug/l   | 99       |
| 18) Methyl Acetate           | 2.715  | 43    | 29635    | 17.827   | ug/l   | 93       |
| 19) Methyl tert-butyl Ether  | 3.123  | 73    | 56648    | 17.756   | ug/l   | 97       |
| 20) Methylene Chloride       | 2.794  | 84    | 20891    | 16.600   | ug/l # | 87       |
| 21) trans-1,2-Dichloroethene | 3.093  | 96    | 18349    | 17.027   | ug/l   | 97       |
| 22) Diisopropyl ether        | 3.770  | 45    | 62769    | 17.833   | ug/l   | 91       |
| 23) Vinyl Acetate            | 3.733  | 43    | 289102   | 87.284   | ug/l   | 94       |
| 24) 1,1-Dichloroethane       | 3.617  | 63    | 35887    | 17.913   | ug/l   | 98       |
| 25) 2-Butanone               | 4.568  | 43    | 84428    | 84.750   | ug/l   | 92       |
| 26) 2,2-Dichloropropane      | 4.483  | 77    | 27806    | 18.111   | ug/l   | 98       |
| 27) cis-1,2-Dichloroethene   | 4.495  | 96    | 21824    | 17.250   | ug/l   | 95       |
| 28) Bromochloromethane       | 4.904  | 49    | 17148    | 16.933   | ug/l # | 79       |
| 29) Tetrahydrofuran          | 5.019  | 42    | 49870    | 85.303   | ug/l   | 88       |
| 30) Chloroform               | 5.099  | 83    | 35340    | 17.564   | ug/l   | 99       |
| 31) Cyclohexane              | 5.483  | 56    | 29797    | 18.777   | ug/l   | 94       |
| 32) 1,1,1-Trichloroethane    | 5.391  | 97    | 28930    | 18.127   | ug/l   | 97       |
| 36) 1,1-Dichloropropene      | 5.702  | 75    | 26325    | 18.389   | ug/l   | 96       |
| 37) Ethyl Acetate            | 4.721  | 43    | 32936    | 17.937   | ug/l   | 94       |
| 38) Carbon Tetrachloride     | 5.678  | 117   | 22278    | 17.960   | ug/l   | 92       |
| 39) Methylcyclohexane        | 7.385  | 83    | 29426    | 19.527   | ug/l   | 90       |
| 40) Benzene                  | 6.050  | 78    | 81269    | 18.395   | ug/l   | 100      |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX060921\  
 Data File : VX022607.D  
 Acq On : 09 Jun 2021 12:36  
 Operator : JC/MD  
 Sample : VX0609WBS01  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 5 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 VX0609WBS01

Manual Integrations  
 APPROVED

MMDadoda  
 6/10/2021 12:11:56 PM

Quant Time: Jun 10 05:39:56 2021  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X060721W.M  
 Quant Title : SW846 8260  
 QLast Update : Mon Jun 07 14:27:50 2021  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon | Response | Conc    | Units  | Dev(Min) |
|-------------------------------|--------|------|----------|---------|--------|----------|
| 41) Methacrylonitrile         | 4.934  | 41   | 18160    | 17.974  | ug/l   | 94       |
| 42) 1,2-Dichloroethane        | 6.098  | 62   | 29303    | 17.993  | ug/l   | 98       |
| 43) Isopropyl Acetate         | 6.354  | 43   | 53788    | 18.058  | ug/l # | 91       |
| 44) Trichloroethene           | 7.129  | 130  | 18929    | 18.159  | ug/l   | 95       |
| 45) 1,2-Dichloropropane       | 7.434  | 63   | 21616    | 19.021  | ug/l   | 98       |
| 46) Dibromomethane            | 7.586  | 93   | 14231    | 18.371  | ug/l # | 81       |
| 47) Bromodichloromethane      | 7.824  | 83   | 24242    | 17.591  | ug/l   | 97       |
| 48) Methyl methacrylate       | 7.696  | 41   | 25319    | 18.267  | ug/l   | 87       |
| 49) 1,4-Dioxane               | 7.671  | 88   | 9574     | 341.544 | ug/l # | 90       |
| 51) 4-Methyl-2-Pentanone      | 8.580  | 43   | 169429   | 91.306  | ug/l   | 91       |
| 52) Toluene                   | 8.720  | 92   | 50014    | 18.025  | ug/l   | 100      |
| 53) t-1,3-Dichloropropene     | 8.982  | 75   | 27143    | 17.097  | ug/l   | 99       |
| 54) cis-1,3-Dichloropropene   | 8.372  | 75   | 31316    | 17.793  | ug/l   | 92       |
| 55) 1,1,2-Trichloroethane     | 9.153  | 97   | 20299    | 18.770  | ug/l   | 96       |
| 56) Ethyl methacrylate        | 9.122  | 69   | 32232    | 17.881  | ug/l # | 84       |
| 57) 1,3-Dichloropropane       | 9.311  | 76   | 36589    | 18.721  | ug/l   | 98       |
| 58) 2-Chloroethyl Vinyl ether | 8.244  | 63   | 90500    | 86.569  | ug/l   | 96       |
| 59) 2-Hexanone                | 9.433  | 43   | 127160   | 89.702  | ug/l   | 90       |
| 60) Dibromochloromethane      | 9.525  | 129  | 16017    | 17.343  | ug/l   | 98       |
| 61) 1,2-Dibromoethane         | 9.610  | 107  | 20204    | 17.722  | ug/l   | 100      |
| 64) Tetrachloroethene         | 9.275  | 164  | 16333    | 18.184  | ug/l   | 89       |
| 65) Chlorobenzene             | 10.086 | 112  | 52525    | 18.840  | ug/l   | 99       |
| 66) 1,1,1,2-Tetrachloroethane | 10.165 | 131  | 16331    | 17.547  | ug/l   | 96       |
| 67) Ethyl Benzene             | 10.195 | 91   | 97072    | 18.922  | ug/l   | 98       |
| 68) m/p-Xylenes               | 10.305 | 106  | 71806    | 38.511  | ug/l   | 100      |
| 69) o-Xylene                  | 10.646 | 106  | 34930    | 19.160  | ug/l   | 99       |
| 70) Styrene                   | 10.659 | 104  | 57341    | 18.603  | ug/l   | 98       |
| 71) Bromoform                 | 10.805 | 173  | 8925     | 16.721  | ug/l # | 97       |
| 73) Isopropylbenzene          | 10.963 | 105  | 91616    | 19.402  | ug/l   | 98       |
| 74) N-amyl acetate            | 10.848 | 43   | 45387    | 17.915  | ug/l   | 91       |
| 75) 1,1,2,2-Tetrachloroethane | 11.213 | 83   | 31154    | 18.330  | ug/l   | 98       |
| 76) 1,2,3-Trichloropropane    | 11.244 | 75   | 28208m   | 18.794  | ug/l   |          |
| 77) Bromobenzene              | 11.201 | 156  | 18647    | 18.329  | ug/l   | 80       |
| 78) n-propylbenzene           | 11.305 | 91   | 109971   | 19.189  | ug/l   | 97       |
| 79) 2-Chlorotoluene           | 11.366 | 91   | 65242    | 19.130  | ug/l   | 97       |
| 80) 1,3,5-Trimethylbenzene    | 11.457 | 105  | 77380    | 19.180  | ug/l   | 99       |
| 81) trans-1,4-Dichloro-2-b... | 11.024 | 75   | 7981     | 17.737  | ug/l # | 85       |
| 82) 4-Chlorotoluene           | 11.457 | 91   | 75587    | 19.148  | ug/l   | 94       |
| 83) tert-Butylbenzene         | 11.719 | 119  | 69793    | 19.420  | ug/l   | 91       |
| 84) 1,2,4-Trimethylbenzene    | 11.756 | 105  | 76946    | 18.964  | ug/l   | 98       |
| 85) sec-Butylbenzene          | 11.896 | 105  | 92716    | 19.386  | ug/l   | 98       |
| 86) p-Isopropyltoluene        | 12.012 | 119  | 75982    | 19.222  | ug/l   | 97       |
| 87) 1,3-Dichlorobenzene       | 11.969 | 146  | 36327    | 18.388  | ug/l   | 95       |
| 88) 1,4-Dichlorobenzene       | 12.043 | 146  | 37541    | 18.240  | ug/l   | 95       |
| 89) n-Butylbenzene            | 12.335 | 91   | 73259    | 19.302  | ug/l   | 98       |
| 90) Hexachloroethane          | 12.542 | 117  | 9277     | 16.635  | ug/l   | 73       |
| 91) 1,2-Dichlorobenzene       | 12.335 | 146  | 36042    | 19.222  | ug/l   | 96       |
| 92) 1,2-Dibromo-3-Chloropr... | 12.945 | 75   | 5912     | 17.894  | ug/l   | 74       |
| 93) 1,2,4-Trichlorobenzene    | 13.591 | 180  | 20752    | 18.409  | ug/l   | 97       |
| 94) Hexachlorobutadiene       | 13.725 | 225  | 7855     | 19.443  | ug/l   | 93       |
| 95) Naphthalene               | 13.780 | 128  | 82255    | 18.201  | ug/l   | 99       |
| 96) 1,2,3-Trichlorobenzene    | 13.963 | 180  | 20230    | 17.956  | ug/l   | 97       |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX060921\  
 Data File : VX022607.D  
 Acq On : 09 Jun 2021 12:36  
 Operator : JC/MD  
 Sample : VX0609WBS01  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 5 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 VX0609WBS01

Manual Integrations  
 APPROVED

MMDadoda  
 6/10/2021 12:11:56 PM

Quant Time: Jun 10 05:39:56 2021  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X060721W.M  
 Quant Title : SW846 8260  
 QLast Update : Mon Jun 07 14:27:50 2021  
 Response via : Initial Calibration

| Compound | R.T. | QIon | Response | Conc | Units | Dev(Min) |
|----------|------|------|----------|------|-------|----------|
|----------|------|------|----------|------|-------|----------|

-----

(#) = qualifier out of range (m) = manual integration (+) = signals summed


MMDadoda  
6/10/2021 12:11:56 PM

Abundance


TIC: VX022607.D\data.ms

Time-->




Chlorodifluoromethane, T  
Vinyl Chloride, P  
Benzonitrile, T  
Chloromethane, T  
Trichlorofluoromethane, T  
Diethyl Ether, T  
Acrolein, T  
Methyl Chloride, T  
Methyl Acetylene Chloride, T  
Tert butyl alcohol, T  
Methyl isopropyl ether, T  
1,1-Dichloroethane, P  
Vinyl Acetate, T  
Diisopropyl ether, T  
2-Butanol, T  
Ethyl Acetate, T  
Benzonitrile, T  
Chloroform, T  
Cyclohexane, T  
Carbon tetrachloride, T  
Dibromochloromethane, S  
Perfluorobenzene, I  
1,2-Dichloroethane-d4, S  
1,2-Dichloroethane, T  
Isopropyl Acetate, T  
1,4-Difluorobenzene, I  
Trichloroethene, TM  
1,2-Dichloroethane, T  
1,4-Dichlorobenzene, T  
Bromochloromethane, T  
cis-1,3-Dichloropropene, T  
2-Chloroethyl Vinyl ether, T  
4-Methyl-2-Pentanol, S  
Toluene, CM  
t-1,3-Dichloropropene, T  
Ethyl propyl ether, T  
1,3-Dichlorobenzene, T  
1,2-Dichloropropane, T  
1,2-Dichloroethane, I  
1,1,1,2-Tetrachloroethane, PM  
Chlorobenzene, C  
Ethyl Benzene, C  
Styrene, T  
o-Xylene, T  
N-amyl acetate, T  
Isopropylbenzene, T  
4-Bromofluorobenzene, S  
trans-1,4-Dichloro-2-butene, T  
1,2,3-Trichlorobenzene, T  
2-Chlorobenzene, T  
1,2,4-Trichlorobenzene, T  
1,2,4-Trimethylbenzene, T  
p-Isopropylbenzene, T  
1,3-Dichlorobenzene, T  
1,4-Dichlorobenzene, T  
p-Isopropylbenzene, T  
1,2-Dichlorobenzene, T  
Hexachloroethane, T  
1,2-Dibromo-3-Chloropropane, T  
Hexachlorobutadiene, T  
1,2,4-Trichlorobenzene, T  
Naphthalene, T  
1,2,3-Trichlorobenzene, T




#1  
Pentafluorobenzene  
Concen: 50.000 ug/l  
RT: 5.562 min Scan# 734  
Delta R.T. -0.000 min  
Lab File: VX022607.D  
Acq: 09 Jun 2021 12:36





Tgt Ion: 168 Resp: 75459  
Ion Ratio Lower Upper  
168 100  
99 66.5 45.8 68.6



#2  
Dichlorodifluoromethane  
Concen: 19.123 ug/l  
RT: 1.173 min Scan# 14  
Delta R.T. 0.007 min  
Lab File: VX022607.D  
Acq: 09 Jun 2021 12:36



Tgt Ion: 85 Resp: 16828  
Ion Ratio Lower Upper  
85 100  
87 32.7 16.4 49.2




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0609WBS01

Manual Integrations  
APPROVED

MMDadoda  
6/10/2021 12:11:56 PM

Abundance Scan 34 (1.294 min): VX022528.D\data.ms (-30)







#3  
Chloromethane  
Concen: 18.227 ug/l  
RT: 1.301 min Scan# 35  
Delta R.T. 0.007 min  
Lab File: VX022607.D  
Acq: 09 Jun 2021 12:36

Tgt Ion: 50 Resp: 19372  
Ion Ratio Lower Upper  
50 100  
52 35.7 26.4 39.6

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0609WBS01




Manual Integrations  
APPROVED


MMDadoda  
6/10/2021 12:11:56 PM



#4  
Vinyl Chloride  
Concen: 17.789 ug/l  
RT: 1.380 min Scan# 48  
Delta R.T. 0.006 min  
Lab File: VX022607.D  
Acq: 09 Jun 2021 12:36

Tgt Ion: 62 Resp: 18453  
Ion Ratio Lower Upper  
62 100  
64 32.0 25.7 38.5









#5  
Bromomethane  
Concen: 20.634 ug/l  
RT: 1.599 min Scan# 84  
Delta R.T. 0.000 min  
Lab File: VX022607.D  
Acq: 09 Jun 2021 12:36

Tgt Ion: 94 Resp: 11246  
Ion Ratio Lower Upper  
94 100  
96 98.6 76.4 114.6

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0609WBS01




Manual Integrations  
APPROVED


MMDadoda  
6/10/2021 12:11:56 PM



#6  
Chloroethane  
Concen: 17.802 ug/l  
RT: 1.685 min Scan# 98  
Delta R.T. 0.006 min  
Lab File: VX022607.D  
Acq: 09 Jun 2021 12:36

Tgt Ion: 64 Resp: 11628  
Ion Ratio Lower Upper  
64 100  
66 31.8 25.4 38.0









#7  
Trichlorofluoromethane  
Concen: 17.709 ug/l  
RT: 1.886 min Scan# 131  
Delta R.T. -0.000 min  
Lab File: VX022607.D  
Acq: 09 Jun 2021 12:36

Tgt Ion:101 Resp: 26754  
Ion Ratio Lower Upper  
101 100  
103 66.8 53.2 79.8

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0609WBS01




Manual Integrations  
APPROVED


MMDadoda  
6/10/2021 12:11:56 PM



#8  
Diethyl Ether  
Concen: 17.874 ug/l  
RT: 2.136 min Scan# 172  
Delta R.T. -0.006 min  
Lab File: VX022607.D  
Acq: 09 Jun 2021 12:36

Tgt Ion: 74 Resp: 11009  
Ion Ratio Lower Upper  
74 100  
45 98.3 44.5 133.5









#9  
1,1,2-Trichlorotrifluoroethane  
Concen: 18.349 ug/l  
RT: 2.331 min Scan# 204  
Delta R.T. 0.006 min  
Lab File: VX022607.D  
Acq: 09 Jun 2021 12:36

|          |       |       |       |
|----------|-------|-------|-------|
| Tgt Ion: | 101   | Resp: | 16996 |
| Ion      | Ratio | Lower | Upper |
| 101      | 100   |       |       |
| 85       | 49.2  | 34.9  | 52.3  |
| 151      | 67.8  | 61.4  | 92.2  |

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0609WBS01




Manual Integrations  
APPROVED


MMDadoda  
6/10/2021 12:11:56 PM



#10  
Methyl Iodide  
Concen: 14.663 ug/l  
RT: 2.453 min Scan# 224  
Delta R.T. -0.000 min  
Lab File: VX022607.D  
Acq: 09 Jun 2021 12:36

|          |       |       |       |
|----------|-------|-------|-------|
| Tgt Ion: | 142   | Resp: | 15988 |
| Ion      | Ratio | Lower | Upper |
| 142      | 100   |       |       |
| 127      | 46.0  | 33.3  | 49.9  |
| 141      | 13.6  | 11.4  | 17.0  |









#11  
Tert butyl alcohol  
Concen: 84.827 ug/l  
RT: 2.977 min Scan# 310  
Delta R.T. -0.006 min  
Lab File: VX022607.D  
Acq: 09 Jun 2021 12:36

Tgt Ion: 59 Resp: 25524  
Ion Ratio Lower Upper  
59 100  
57 9.4 7.8 11.8

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0609WBS01




Manual Integrations  
APPROVED


MMDadoda  
6/10/2021 12:11:56 PM



#12  
1,1-Dichloroethene  
Concen: 16.698 ug/l  
RT: 2.319 min Scan# 202  
Delta R.T. -0.000 min  
Lab File: VX022607.D  
Acq: 09 Jun 2021 12:36


Tgt Ion: 96 Resp: 15977  
Ion Ratio Lower Upper  
96 100  
61 163.2 124.6 187.0  
98 59.0 50.6 76.0








#13  
Acrolein  
Concen: 77.376 ug/l  
RT: 2.239 min Scan# 189  
Delta R.T. 0.000 min  
Lab File: VX022607.D  
Acq: 09 Jun 2021 12:36


Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0609WBS01





Tgt Ion: 56 Resp: 10795  
Ion Ratio Lower Upper  
56 100  
55 68.1 56.4 84.6




#14  
Allyl chloride  
Concen: 17.471 ug/l  
RT: 2.666 min Scan# 259  
Delta R.T. 0.000 min  
Lab File: VX022607.D  
Acq: 09 Jun 2021 12:36




Tgt Ion: 41 Resp: 30653  
Ion Ratio Lower Upper  
41 100  
39 69.2 51.6 77.4  
76 32.9 30.5 45.7









#15  
Acrylonitrile  
Concen: 87.058 ug/l  
RT: 3.075 min Scan# 326  
Delta R.T. -0.000 min  
Lab File: VX022607.D  
Acq: 09 Jun 2021 12:36





Tgt Ion: 53 Resp: 54612  
Ion Ratio Lower Upper  
53 100  
52 81.9 65.3 97.9  
51 35.6 28.2 42.2



#16  
Acetone  
Concen: 83.885 ug/l  
RT: 2.386 min Scan# 213  
Delta R.T. -0.006 min  
Lab File: VX022607.D  
Acq: 09 Jun 2021 12:36




Tgt Ion: 43 Resp: 50564  
Ion Ratio Lower Upper  
43 100  
58 31.5 27.8 41.8




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0609WBS01

Manual Integrations  
APPROVED




MMDadoda  
6/10/2021 12:11:56 PM




#17  
Carbon Disulfide  
Concen: 16.888 ug/l  
RT: 2.514 min Scan# 234  
Delta R.T. -0.000 min  
Lab File: VX022607.D  
Acq: 09 Jun 2021 12:36





Tgt Ion: 76 Resp: 34677  
Ion Ratio Lower Upper  
76 100  
78 8.6 7.2 10.8



#18  
Methyl Acetate  
Concen: 17.827 ug/l  
RT: 2.715 min Scan# 267  
Delta R.T. 0.006 min  
Lab File: VX022607.D  
Acq: 09 Jun 2021 12:36




Tgt Ion: 43 Resp: 29635  
Ion Ratio Lower Upper  
43 100  
74 23.8 21.9 32.9



Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0609WBS01

Manual Integrations  
APPROVED

MMDadoda  
6/10/2021 12:11:56 PM







#19  
Methyl tert-butyl Ether  
Concen: 17.756 ug/l  
RT: 3.123 min Scan# 334  
Delta R.T. 0.000 min  
Lab File: VX022607.D  
Acq: 09 Jun 2021 12:36

Tgt Ion: 73 Resp: 56648  
Ion Ratio Lower Upper  
73 100  
57 23.0 17.1 25.7

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0609WBS01




Manual Integrations  
APPROVED


MMDadoda  
6/10/2021 12:11:56 PM




#20  
Methylene Chloride  
Concen: 16.600 ug/l  
RT: 2.794 min Scan# 280  
Delta R.T. 0.000 min  
Lab File: VX022607.D  
Acq: 09 Jun 2021 12:36

Tgt Ion: 84 Resp: 20891  
Ion Ratio Lower Upper  
84 100  
49 133.9 88.6 132.8#  
51 39.0 27.0 40.6  
86 63.4 50.3 75.5









#21  
trans-1,2-Dichloroethene  
Concen: 17.027 ug/l  
RT: 3.093 min Scan# 329  
Delta R.T. -0.000 min  
Lab File: VX022607.D  
Acq: 09 Jun 2021 12:36





Tgt Ion: 96 Resp: 18349  
Ion Ratio Lower Upper  
96 100  
61 138.9 107.7 161.5  
98 60.9 50.4 75.6



#22  
Diisopropyl ether  
Concen: 17.833 ug/l  
RT: 3.770 min Scan# 440  
Delta R.T. -0.000 min  
Lab File: VX022607.D  
Acq: 09 Jun 2021 12:36




Tgt Ion: 45 Resp: 62769  
Ion Ratio Lower Upper  
45 100  
43 81.5 58.0 87.0  
87 28.4 25.4 38.0  
59 11.8 10.2 15.2



Instrument :  
MSVOA\_X  
ClientSampled :  
VX0609WBS01

Manual Integrations  
APPROVED

MMDadoda  
6/10/2021 12:11:56 PM







#23  
 Vinyl Acetate  
 Concen: 87.284 ug/l  
 RT: 3.733 min Scan# 434  
 Delta R.T. 0.006 min  
 Lab File: VX022607.D  
 Acq: 09 Jun 2021 12:36

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 VX0609WBS01

Tgt Ion: 43 Resp: 289102  
 Ion Ratio Lower Upper  
 43 100  
 86 10.2 9.9 14.9




Manual Integrations  
 APPROVED


MMDadoda  
 6/10/2021 12:11:56 PM



#24  
 1,1-Dichloroethane  
 Concen: 17.913 ug/l  
 RT: 3.617 min Scan# 415  
 Delta R.T. 0.006 min  
 Lab File: VX022607.D  
 Acq: 09 Jun 2021 12:36


Tgt Ion: 63 Resp: 35887  
 Ion Ratio Lower Upper  
 63 100  
 98 7.9 3.8 11.2  
 100 3.7 2.5 7.5





#25  
2-Butanone  
Concen: 84.750 ug/l  
RT: 4.568 min Scan# 571  
Delta R.T. 0.000 min  
Lab File: VX022607.D  
Acq: 09 Jun 2021 12:36




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0609WBS01



Tgt Ion: 43 Resp: 84428  
Ion Ratio Lower Upper  
43 100  
72 25.5 23.7 35.5




Manual Integrations  
APPROVED


MMDadoda  
6/10/2021 12:11:56 PM



#26  
2,2-Dichloropropane  
Concen: 18.111 ug/l  
RT: 4.483 min Scan# 557  
Delta R.T. -0.006 min  
Lab File: VX022607.D  
Acq: 09 Jun 2021 12:36

Tgt Ion: 77 Resp: 27806  
Ion Ratio Lower Upper  
77 100  
97 22.3 11.7 35.1





#27  
cis-1,2-Dichloroethene  
Concen: 17.250 ug/l  
RT: 4.495 min Scan# 559  
Delta R.T. 0.000 min  
Lab File: VX022607.D  
Acq: 09 Jun 2021 12:36





Tgt Ion: 96 Resp: 21824

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 96  | 100   |       |       |
| 61  | 150.8 | 0.0   | 285.6 |
| 98  | 63.5  | 0.0   | 129.6 |

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0609WBS01

Manual Integrations  
APPROVED




MMDadoda  
6/10/2021 12:11:56 PM




#28  
Bromochloromethane  
Concen: 16.933 ug/l  
RT: 4.904 min Scan# 626  
Delta R.T. -0.000 min  
Lab File: VX022607.D  
Acq: 09 Jun 2021 12:36

Tgt Ion: 49 Resp: 17148

| Ion | Ratio | Lower | Upper  |
|-----|-------|-------|--------|
| 49  | 100   |       |        |
| 129 | 1.6   | 0.0   | 4.0    |
| 130 | 65.9  | 68.2  | 102.2# |









#29  
Tetrahydrofuran  
Concen: 85.303 ug/l  
RT: 5.019 min Scan# 645  
Delta R.T. 0.006 min  
Lab File: VX022607.D  
Acq: 09 Jun 2021 12:36

Tgt Ion: 42 Resp: 49870  
Ion Ratio Lower Upper  
42 100  
72 42.6 41.7 62.5  
71 41.0 38.4 57.6

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0609WBS01




Manual Integrations  
APPROVED


MMDadoda  
6/10/2021 12:11:56 PM



#30  
Chloroform  
Concen: 17.564 ug/l  
RT: 5.099 min Scan# 658  
Delta R.T. -0.006 min  
Lab File: VX022607.D  
Acq: 09 Jun 2021 12:36


Tgt Ion: 83 Resp: 35340  
Ion Ratio Lower Upper  
83 100  
85 64.3 52.0 78.0





#31  
Cyclohexane  
Concen: 18.777 ug/l  
RT: 5.483 min Scan# 721  
Delta R.T. 0.006 min  
Lab File: VX022607.D  
Acq: 09 Jun 2021 12:36




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0609WBS01




Tgt Ion: 56 Resp: 29797  
Ion Ratio Lower Upper  
56 100  
69 28.4 25.3 37.9  
84 83.4 71.4 107.0

Manual Integrations  
APPROVED



MMDadoda  
6/10/2021 12:11:56 PM




#32  
1,1,1-Trichloroethane  
Concen: 18.127 ug/l  
RT: 5.391 min Scan# 706  
Delta R.T. 0.000 min  
Lab File: VX022607.D  
Acq: 09 Jun 2021 12:36



Tgt Ion: 97 Resp: 28930  
Ion Ratio Lower Upper  
97 100  
99 65.4 51.5 77.3  
61 47.1 34.2 51.4









#33  
1,2-Dichloroethane-d4  
Concen: 44.615 ug/l  
RT: 5.964 min Scan# 800  
Delta R.T. 0.000 min  
Lab File: VX022607.D  
Acq: 09 Jun 2021 12:36

Tgt Ion: 65 Resp: 59849  
Ion Ratio Lower Upper  
65 100  
67 54.9 0.0 105.4

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0609WBS01




Manual Integrations  
APPROVED


MMDadoda  
6/10/2021 12:11:56 PM




#34  
1,4-Difluorobenzene  
Concen: 50.000 ug/l  
RT: 6.769 min Scan# 932  
Delta R.T. 0.000 min  
Lab File: VX022607.D  
Acq: 09 Jun 2021 12:36

Tgt Ion: 114 Resp: 147983  
Ion Ratio Lower Upper  
114 100  
63 22.4 0.0 40.8  
88 15.8 0.0 32.8









#35  
Dibromofluoromethane  
Concen: 47.878 ug/l  
RT: 5.397 min Scan# 707  
Delta R.T. 0.006 min  
Lab File: VX022607.D  
Acq: 09 Jun 2021 12:36





Tgt Ion: 113 Resp: 44329  
Ion Ratio Lower Upper  
113 100  
111 103.1 81.9 122.9  
192 15.2 16.7 25.1#



#36  
1,1-Dichloropropene  
Concen: 18.389 ug/l  
RT: 5.702 min Scan# 757  
Delta R.T. 0.000 min  
Lab File: VX022607.D  
Acq: 09 Jun 2021 12:36




Tgt Ion: 75 Resp: 26325  
Ion Ratio Lower Upper  
75 100  
110 32.6 18.1 54.4  
77 31.9 25.3 37.9



Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0609WBS01

Manual Integrations  
APPROVED

MMDadoda  
6/10/2021 12:11:56 PM



#37

Ethyl Acetate

Concen: 17.937 ug/l

RT: 4.721 min Scan# 596

Delta R.T. -0.006 min

Lab File: VX022607.D





Acq: 09 Jun 2021 12:36

|          |       |       |       |
|----------|-------|-------|-------|
| Tgt Ion: | 43    | Resp: | 32936 |
| Ion      | Ratio | Lower | Upper |
| 43       | 100   |       |       |
| 61       | 13.1  | 12.8  | 19.2  |
| 70       | 10.9  | 10.5  | 15.7  |

Instrument :  
MSVOA\_X  
ClientSampled :  
VX0609WBS01

Manual Integrations  
APPROVED

MMDadoda  
6/10/2021 12:11:56 PM



#38

Carbon Tetrachloride

Concen: 17.960 ug/l




RT: 5.678 min Scan# 753


Delta R.T. -0.000 min

Lab File: VX022607.D

Acq: 09 Jun 2021 12:36

|          |       |       |       |
|----------|-------|-------|-------|
| Tgt Ion: | 117   | Resp: | 22278 |
| Ion      | Ratio | Lower | Upper |
| 117      | 100   |       |       |
| 119      | 84.6  | 76.2  | 114.2 |
| 121      | 31.4  | 24.8  | 37.2  |









#39  
Methylcyclohexane  
Concen: 19.527 ug/l  
RT: 7.385 min Scan# 1033  
Delta R.T. -0.000 min  
Lab File: VX022607.D  
Acq: 09 Jun 2021 12:36

Tgt Ion: 83 Resp: 29426  
Ion Ratio Lower Upper  
83 100  
55 82.5 58.9 88.3  
98 53.7 38.1 57.1

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0609WBS01




Manual Integrations  
APPROVED


MMDadoda  
6/10/2021 12:11:56 PM




#40  
Benzene  
Concen: 18.395 ug/l  
RT: 6.050 min Scan# 814  
Delta R.T. -0.000 min  
Lab File: VX022607.D  
Acq: 09 Jun 2021 12:36

Tgt Ion: 78 Resp: 81269  
Ion Ratio Lower Upper  
78 100  
77 23.5 18.9 28.3









#41  
Methacrylonitrile  
Concen: 17.974 ug/l  
RT: 4.934 min Scan# 631  
Delta R.T. 0.006 min  
Lab File: VX022607.D  
Acq: 09 Jun 2021 12:36





|          |       |       |       |
|----------|-------|-------|-------|
| Tgt Ion: | 41    | Resp: | 18160 |
| Ion      | Ratio | Lower | Upper |
| 41       | 100   |       |       |
| 39       | 54.1  | 42.5  | 63.7  |
| 67       | 72.5  | 65.0  | 97.6  |
| 52       | 28.2  | 25.3  | 37.9  |



#42  
1,2-Dichloroethane  
Concen: 17.993 ug/l  
RT: 6.098 min Scan# 822  
Delta R.T. 0.000 min  
Lab File: VX022607.D  
Acq: 09 Jun 2021 12:36




|          |       |       |       |
|----------|-------|-------|-------|
| Tgt Ion: | 62    | Resp: | 29303 |
| Ion      | Ratio | Lower | Upper |
| 62       | 100   |       |       |
| 98       | 9.2   | 0.0   | 20.2  |



Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0609WBS01

Manual Integrations  
APPROVED

MMDadoda  
6/10/2021 12:11:56 PM







#43  
Isopropyl Acetate  
Concen: 18.058 ug/l  
RT: 6.354 min Scan# 864  
Delta R.T. 0.006 min  
Lab File: VX022607.D  
Acq: 09 Jun 2021 12:36

Tgt Ion: 43 Resp: 53788  
Ion Ratio Lower Upper  
43 100  
61 18.3 18.2 27.4  
87 12.1 12.5 18.7#

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0609WBS01




Manual Integrations  
APPROVED


MMDadoda  
6/10/2021 12:11:56 PM



#44  
Trichloroethene  
Concen: 18.159 ug/l  
RT: 7.129 min Scan# 991  
Delta R.T. -0.000 min  
Lab File: VX022607.D  
Acq: 09 Jun 2021 12:36

Tgt Ion:130 Resp: 18929  
Ion Ratio Lower Upper  
130 100  
95 104.4 0.0 198.0





#45





1,2-Dichloropropane  
Concen: 19.021 ug/l  
RT: 7.434 min Scan# 1041  
Delta R.T. -0.006 min  
Lab File: VX022607.D  
Acq: 09 Jun 2021 12:36

Tgt Ion: 63 Resp: 21616  
Ion Ratio Lower Upper  
63 100  
65 31.0 25.7 38.5

Instrument :  
MSVOA\_X  
ClientSampled :  
VX0609WBS01

Manual Integrations  
APPROVED




MMDadoda  
6/10/2021 12:11:56 PM




#46

Dibromomethane  
Concen: 18.371 ug/l  
RT: 7.586 min Scan# 1066  
Delta R.T. -0.000 min  
Lab File: VX022607.D  
Acq: 09 Jun 2021 12:36

Tgt Ion: 93 Resp: 14231  
Ion Ratio Lower Upper  
93 100  
95 78.3 66.8 100.2  
174 73.5 83.1 124.7#









#47  
Bromodichloromethane  
Concen: 17.591 ug/l  
RT: 7.824 min Scan# 1105  
Delta R.T. -0.006 min  
Lab File: VX022607.D  
Acq: 09 Jun 2021 12:36

Tgt Ion: 83 Resp: 24242  
Ion Ratio Lower Upper  
83 100  
85 66.5 51.4 77.0  
127 7.1 6.8 10.2

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0609WBS01




Manual Integrations  
APPROVED


MMDadoda  
6/10/2021 12:11:56 PM



#48  
Methyl methacrylate  
Concen: 18.267 ug/l  
RT: 7.696 min Scan# 1084  
Delta R.T. -0.000 min  
Lab File: VX022607.D  
Acq: 09 Jun 2021 12:36

Tgt Ion: 41 Resp: 25319  
Ion Ratio Lower Upper  
41 100  
69 80.1 76.0 114.0  
39 62.0 44.6 67.0









#49  
1,4-Dioxane  
Concen: 341.544 ug/l  
RT: 7.671 min Scan# 1080  
Delta R.T. 0.006 min  
Lab File: VX022607.D  
Acq: 09 Jun 2021 12:36

Tgt Ion: 88 Resp: 9574  
Ion Ratio Lower Upper  
88 100  
43 40.4 23.2 34.8#  
58 72.6 55.0 82.4

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0609WBS01




Manual Integrations  
APPROVED


MMDadoda  
6/10/2021 12:11:56 PM



#50  
Toluene-d8  
Concen: 47.357 ug/l  
RT: 8.653 min Scan# 1241  
Delta R.T. -0.000 min  
Lab File: VX022607.D  
Acq: 09 Jun 2021 12:36

Tgt Ion: 98 Resp: 174952  
Ion Ratio Lower Upper  
98 100  
100 65.3 51.9 77.9









#51  
4-Methyl-2-Pentanone  
Concen: 91.306 ug/l  
RT: 8.580 min Scan# 1229  
Delta R.T. -0.000 min  
Lab File: VX022607.D  
Acq: 09 Jun 2021 12:36

Tgt Ion: 43 Resp: 169429  
Ion Ratio Lower Upper  
43 100  
58 38.1 35.4 53.0

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0609WBS01




Manual Integrations  
APPROVED


MMDadoda  
6/10/2021 12:11:56 PM



#52  
Toluene  
Concen: 18.025 ug/l  
RT: 8.720 min Scan# 1252  
Delta R.T. -0.000 min  
Lab File: VX022607.D  
Acq: 09 Jun 2021 12:36

Tgt Ion: 92 Resp: 50014  
Ion Ratio Lower Upper  
92 100  
91 171.8 137.1 205.7









#53  
t-1,3-Dichloropropene  
Concen: 17.097 ug/l  
RT: 8.982 min Scan# 1295  
Delta R.T. 0.000 min  
Lab File: VX022607.D  
Acq: 09 Jun 2021 12:36

Tgt Ion: 75 Resp: 27143  
Ion Ratio Lower Upper  
75 100  
77 30.6 24.9 37.3

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0609WBS01




Manual Integrations  
APPROVED


MMDadoda  
6/10/2021 12:11:56 PM




#54  
cis-1,3-Dichloropropene  
Concen: 17.793 ug/l  
RT: 8.372 min Scan# 1195  
Delta R.T. 0.000 min  
Lab File: VX022607.D  
Acq: 09 Jun 2021 12:36

Tgt Ion: 75 Resp: 31316  
Ion Ratio Lower Upper  
75 100  
77 33.9 25.7 38.5  
39 50.9 35.1 52.7








#55  
1,1,2-Trichloroethane  
Concen: 18.770 ug/l  
RT: 9.153 min Scan# 1323  
Delta R.T. -0.006 min  
Lab File: VX022607.D  
Acq: 09 Jun 2021 12:36




Tgt Ion: 97 Resp: 20299

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 97  | 100   |       |       |
| 83  | 93.9  | 69.3  | 103.9 |
| 85  | 58.3  | 44.3  | 66.5  |
| 99  | 62.7  | 50.3  | 75.5  |





#56  
Ethyl methacrylate  
Concen: 17.881 ug/l  
RT: 9.122 min Scan# 1318  
Delta R.T. 0.000 min  
Lab File: VX022607.D  
Acq: 09 Jun 2021 12:36



Tgt Ion: 69 Resp: 32232


| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 69  | 100   |       |       |
| 41  | 69.6  | 46.8  | 70.2  |
| 39  | 40.8  | 24.7  | 37.1# |



Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0609WBS01

Manual Integrations  
APPROVED

MMDadoda  
6/10/2021 12:11:56 PM







#57  
1,3-Dichloropropane  
Concen: 18.721 ug/l  
RT: 9.311 min Scan# 1349  
Delta R.T. 0.000 min  
Lab File: VX022607.D  
Acq: 09 Jun 2021 12:36

Tgt Ion: 76 Resp: 36589  
Ion Ratio Lower Upper  
76 100  
78 33.6 26.2 39.2

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0609WBS01




Manual Integrations  
APPROVED


MMDadoda  
6/10/2021 12:11:56 PM



#58  
2-Chloroethyl Vinyl ether  
Concen: 86.569 ug/l  
RT: 8.244 min Scan# 1174  
Delta R.T. -0.007 min  
Lab File: VX022607.D  
Acq: 09 Jun 2021 12:36

Tgt Ion: 63 Resp: 90500  
Ion Ratio Lower Upper  
63 100  
106 27.5 23.5 35.3









#59  
2-Hexanone  
Concen: 89.702 ug/l  
RT: 9.433 min Scan# 1369  
Delta R.T. 0.000 min  
Lab File: VX022607.D  
Acq: 09 Jun 2021 12:36

Tgt Ion: 43 Resp: 127160  
Ion Ratio Lower Upper  
43 100  
58 54.3 30.9 92.8

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0609WBS01




Manual Integrations  
APPROVED


MMDadoda  
6/10/2021 12:11:56 PM



#60  
Dibromochloromethane  
Concen: 17.343 ug/l  
RT: 9.525 min Scan# 1384  
Delta R.T. -0.000 min  
Lab File: VX022607.D  
Acq: 09 Jun 2021 12:36

Tgt Ion: 129 Resp: 16017  
Ion Ratio Lower Upper  
129 100  
127 75.4 38.4 115.2









#61  
1,2-Dibromoethane  
Concen: 17.722 ug/l  
RT: 9.610 min Scan# 1398  
Delta R.T. -0.006 min  
Lab File: VX022607.D  
Acq: 09 Jun 2021 12:36

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0609WBS01

Tgt Ion:107 Resp: 20204  
Ion Ratio Lower Upper  
107 100  
109 94.8 75.8 113.6




Manual Integrations  
APPROVED


MMDadoda  
6/10/2021 12:11:56 PM




#62  
4-Bromofluorobenzene  
Concen: 46.802 ug/l  
RT: 11.085 min Scan# 1640  
Delta R.T. 0.000 min  
Lab File: VX022607.D  
Acq: 09 Jun 2021 12:36

Tgt Ion: 95 Resp: 64267  
Ion Ratio Lower Upper  
95 100  
174 62.2 0.0 154.6  
176 59.9 0.0 155.8









#63  
Chlorobenzene-d5  
Concen: 50.000 ug/l  
RT: 10.055 min Scan# 1471  
Delta R.T. 0.000 min  
Lab File: VX022607.D  
Acq: 09 Jun 2021 12:36





Tgt Ion:117 Resp: 134167  
Ion Ratio Lower Upper  
117 100  
82 61.9 47.4 71.2  
119 31.8 26.6 39.8



#64  
Tetrachloroethene  
Concen: 18.184 ug/l  
RT: 9.275 min Scan# 1343  
Delta R.T. -0.006 min  
Lab File: VX022607.D  
Acq: 09 Jun 2021 12:36




Tgt Ion:164 Resp: 16333  
Ion Ratio Lower Upper  
164 100  
166 118.9 103.4 155.2  
129 101.5 70.8 106.2  
131 100.9 72.5 108.7




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0609WBS01

Manual Integrations  
APPROVED



MMDadoda  
6/10/2021 12:11:56 PM



#65  
Chlorobenzene  
Concen: 18.840 ug/l  
RT: 10.086 min Scan# 1476  
Delta R.T. -0.000 min  
Lab File: VX022607.D  
Acq: 09 Jun 2021 12:36




Tgt Ion:112 Resp: 52525  
Ion Ratio Lower Upper  
112 100  
114 31.6 25.8 38.6




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0609WBS01

Manual Integrations  
APPROVED



MMDadoda  
6/10/2021 12:11:56 PM




#66  
1,1,1,2-Tetrachloroethane  
Concen: 17.547 ug/l  
RT: 10.165 min Scan# 1489  
Delta R.T. -0.000 min  
Lab File: VX022607.D  
Acq: 09 Jun 2021 12:36




Tgt Ion:131 Resp: 16331  
Ion Ratio Lower Upper  
131 100  
133 93.3 47.7 143.1  
119 69.8 32.1 96.5







#67  
Ethyl Benzene  
Concen: 18.922 ug/l  
RT: 10.195 min Scan# 1494  
Delta R.T. 0.000 min  
Lab File: VX022607.D  
Acq: 09 Jun 2021 12:36




Tgt Ion: 91 Resp: 97072  
Ion Ratio Lower Upper  
91 100  
106 29.9 24.6 37.0




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0609WBS01

Manual Integrations  
APPROVED



MMDadoda  
6/10/2021 12:11:56 PM




#68  
m/p-Xylenes  
Concen: 38.511 ug/l  
RT: 10.305 min Scan# 1512  
Delta R.T. 0.000 min  
Lab File: VX022607.D  
Acq: 09 Jun 2021 12:36



Tgt Ion: 106 Resp: 71806  
Ion Ratio Lower Upper  
106 100  
91 204.1 162.8 244.2









#69  
o-Xylene  
Concen: 19.160 ug/l  
RT: 10.646 min Scan# 1568  
Delta R.T. 0.000 min  
Lab File: VX022607.D  
Acq: 09 Jun 2021 12:36

Tgt Ion:106 Resp: 34930  
Ion Ratio Lower Upper  
106 100  
91 216.9 108.1 324.2

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0609WBS01




Manual Integrations  
APPROVED


MMDadoda  
6/10/2021 12:11:56 PM



#70  
Styrene  
Concen: 18.603 ug/l  
RT: 10.659 min Scan# 1570  
Delta R.T. -0.000 min  
Lab File: VX022607.D  
Acq: 09 Jun 2021 12:36

Tgt Ion:104 Resp: 57341  
Ion Ratio Lower Upper  
104 100  
78 52.4 40.0 60.0  
103 53.5 43.4 65.0









#71  
Bromoform  
Concen: 16.721 ug/l  
RT: 10.805 min Scan# 1594  
Delta R.T. -0.000 min  
Lab File: VX022607.D  
Acq: 09 Jun 2021 12:36

Tgt Ion:173 Resp: 8925  
Ion Ratio Lower Upper  
173 100  
175 46.1 23.9 71.8  
254 0.0 0.1 0.1#

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0609WBS01




Manual Integrations  
APPROVED


MMDadoda  
6/10/2021 12:11:56 PM



#72  
1,4-Dichlorobenzene-d4  
Concen: 50.000 ug/l  
RT: 12.024 min Scan# 1794  
Delta R.T. -0.000 min  
Lab File: VX022607.D  
Acq: 09 Jun 2021 12:36

Tgt Ion:152 Resp: 57935  
Ion Ratio Lower Upper  
152 100  
115 78.0 41.1 123.3  
150 166.9 0.0 349.0









#73  
Isopropylbenzene  
Concen: 19.402 ug/l  
RT: 10.963 min Scan# 1620  
Delta R.T. 0.000 min  
Lab File: VX022607.D  
Acq: 09 Jun 2021 12:36

Tgt Ion: 105 Resp: 91616  
Ion Ratio Lower Upper  
105 100  
120 25.5 13.2 39.6

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0609WBS01




Manual Integrations  
APPROVED


MMDadoda  
6/10/2021 12:11:56 PM



#74  
N-ethyl acetate  
Concen: 17.915 ug/l  
RT: 10.848 min Scan# 1601  
Delta R.T. -0.000 min  
Lab File: VX022607.D  
Acq: 09 Jun 2021 12:36

Tgt Ion: 43 Resp: 45387  
Ion Ratio Lower Upper  
43 100  
70 41.1 39.0 58.6  
55 23.6 21.5 32.3  
61 22.9 22.1 33.1









#75  
1,1,2,2-Tetrachloroethane  
Concen: 18.330 ug/l  
RT: 11.213 min Scan# 1661  
Delta R.T. 0.000 min  
Lab File: VX022607.D  
Acq: 09 Jun 2021 12:36

Tgt Ion: 83 Resp: 31154  
Ion Ratio Lower Upper  
83 100  
131 9.1 5.1 15.2  
85 65.4 31.8 95.4

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0609WBS01




Manual Integrations  
APPROVED


MMDadoda  
6/10/2021 12:11:56 PM



#76  
1,2,3-Trichloropropane  
Concen: 18.794 ug/l m  
RT: 11.244 min Scan# 1666  
Delta R.T. -0.000 min  
Lab File: VX022607.D  
Acq: 09 Jun 2021 12:36

Tgt Ion: 75 Resp: 28208  
Ion Ratio Lower Upper  
75 100  
77 38.6 19.1 57.5





#77  
Bromobenzene  
Concen: 18.329 ug/l  
RT: 11.201 min Scan# 1659  
Delta R.T. 0.000 min  
Lab File: VX022607.D  
Acq: 09 Jun 2021 12:36





Tgt Ion: 156 Resp: 18647  

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 156 | 100   |       |       |
| 77  | 204.7 | 82.8  | 248.6 |
| 158 | 93.0  | 49.1  | 147.4 |

Instrument :  
MSVOA\_X  
ClientSampled :  
VX0609WBS01

Manual Integrations  
APPROVED




MMDadoda  
6/10/2021 12:11:56 PM




#78  
n-propylbenzene  
Concen: 19.189 ug/l  
RT: 11.305 min Scan# 1676  
Delta R.T. -0.000 min  
Lab File: VX022607.D  
Acq: 09 Jun 2021 12:36

Tgt Ion: 91 Resp: 109971  

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 91  | 100   |       |       |
| 120 | 21.6  | 11.5  | 34.5  |









#79  
2-Chlorotoluene  
Concen: 19.130 ug/l  
RT: 11.366 min Scan# 1686  
Delta R.T. -0.000 min  
Lab File: VX022607.D  
Acq: 09 Jun 2021 12:36

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0609WBS01

Tgt Ion: 91 Resp: 65242  
Ion Ratio Lower Upper  
91 100  
126 31.5 16.6 49.8




Manual Integrations  
APPROVED


MMDadoda  
6/10/2021 12:11:56 PM



#80  
1,3,5-Trimethylbenzene  
Concen: 19.180 ug/l  
RT: 11.457 min Scan# 1701  
Delta R.T. 0.000 min  
Lab File: VX022607.D  
Acq: 09 Jun 2021 12:36


Tgt Ion:105 Resp: 77380  
Ion Ratio Lower Upper  
105 100  
120 47.0 23.8 71.5





#81  
trans-1,4-Dichloro-2-butene  
Concen: 17.737 ug/l  
RT: 11.024 min Scan# 1630  
Delta R.T. 0.000 min  
Lab File: VX022607.D  
Acq: 09 Jun 2021 12:36




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0609WBS01




Tgt Ion: 75 Resp: 7981  
Ion Ratio Lower Upper  
75 100  
53 113.2 74.5 111.7#  
89 44.6 37.9 56.9

Manual Integrations  
APPROVED



MMDadoda  
6/10/2021 12:11:56 PM




#82  
4-Chlorotoluene  
Concen: 19.148 ug/l  
RT: 11.457 min Scan# 1701  
Delta R.T. 0.000 min  
Lab File: VX022607.D  
Acq: 09 Jun 2021 12:36



Tgt Ion: 91 Resp: 75587  
Ion Ratio Lower Upper  
91 100  
126 26.6 14.9 44.5






#83  
tert-Butylbenzene  
Concen: 19.420 ug/l  
RT: 11.719 min Scan# 1744  
Delta R.T. 0.000 min  
Lab File: VX022607.D  
Acq: 09 Jun 2021 12:36




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0609WBS01

Manual Integrations  
APPROVED

MMDadoda  
6/10/2021 12:11:56 PM







Tgt Ion:119 Resp: 69793  
Ion Ratio Lower Upper  
119 100  
91 64.9 27.9 83.7  
134 23.0 11.3 33.9



#84  
1,2,4-Trimethylbenzene  
Concen: 18.964 ug/l  
RT: 11.756 min Scan# 1750  
Delta R.T. -0.000 min  
Lab File: VX022607.D  
Acq: 09 Jun 2021 12:36

Tgt Ion:105 Resp: 76946  
Ion Ratio Lower Upper  
105 100  
120 43.7 22.4 67.3









#85  
sec-Butylbenzene  
Concen: 19.386 ug/l  
RT: 11.896 min Scan# 1773  
Delta R.T. 0.000 min  
Lab File: VX022607.D  
Acq: 09 Jun 2021 12:36

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0609WBS01

Tgt Ion:105 Resp: 92716  
Ion Ratio Lower Upper  
105 100  
134 18.9 10.0 29.8




Manual Integrations  
APPROVED


MMDadoda  
6/10/2021 12:11:56 PM



#86  
p-Isopropyltoluene  
Concen: 19.222 ug/l  
RT: 12.012 min Scan# 1792  
Delta R.T. 0.000 min  
Lab File: VX022607.D  
Acq: 09 Jun 2021 12:36

Tgt Ion:119 Resp: 75982  
Ion Ratio Lower Upper  
119 100  
134 26.4 12.9 38.7  
91 26.2 11.7 35.0









#87  
1,3-Dichlorobenzene  
Concen: 18.388 ug/l  
RT: 11.969 min Scan# 1785  
Delta R.T. -0.006 min  
Lab File: VX022607.D  
Acq: 09 Jun 2021 12:36

Tgt Ion:146 Resp: 36327  
Ion Ratio Lower Upper  
146 100  
111 46.5 20.2 60.5  
148 64.9 31.9 95.7

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0609WBS01




Manual Integrations  
APPROVED


MMDadoda  
6/10/2021 12:11:56 PM



#88  
1,4-Dichlorobenzene  
Concen: 18.240 ug/l  
RT: 12.043 min Scan# 1797  
Delta R.T. -0.000 min  
Lab File: VX022607.D  
Acq: 09 Jun 2021 12:36

Tgt Ion:146 Resp: 37541  
Ion Ratio Lower Upper  
146 100  
111 46.0 19.6 58.8  
148 62.6 31.6 95.0





#89  
n-Butylbenzene  
Concen: 19.302 ug/l  
RT: 12.335 min Scan# 1845  
Delta R.T. 0.000 min  
Lab File: VX022607.D  
Acq: 09 Jun 2021 12:36





Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0609WBS01

Tgt Ion: 91 Resp: 73259

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 91  | 100   |       |       |
| 92  | 54.7  | 27.4  | 82.0  |
| 134 | 23.5  | 13.2  | 39.5  |

Manual Integrations  
APPROVED




MMDadoda  
6/10/2021 12:11:56 PM




#90  
Hexachloroethane  
Concen: 16.635 ug/l  
RT: 12.542 min Scan# 1879  
Delta R.T. 0.000 min  
Lab File: VX022607.D  
Acq: 09 Jun 2021 12:36

Tgt Ion: 117 Resp: 9277

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 117 | 100   |       |       |
| 201 | 56.5  | 40.4  | 121.2 |









#91  
1,2-Dichlorobenzene  
Concen: 19.222 ug/l  
RT: 12.335 min Scan# 1845  
Delta R.T. 0.000 min  
Lab File: VX022607.D  
Acq: 09 Jun 2021 12:36

Tgt Ion:146 Resp: 36042  
Ion Ratio Lower Upper  
146 100  
111 47.0 20.8 62.2  
148 62.4 31.9 95.5

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0609WBS01




Manual Integrations  
APPROVED


MMDadoda  
6/10/2021 12:11:56 PM



#92  
1,2-Dibromo-3-Chloropropane  
Concen: 17.894 ug/l  
RT: 12.945 min Scan# 1945  
Delta R.T. -0.000 min  
Lab File: VX022607.D  
Acq: 09 Jun 2021 12:36


Tgt Ion: 75 Resp: 5912  
Ion Ratio Lower Upper  
75 100  
155 66.9 45.3 135.9  
157 86.2 57.7 173.1





#93  
1,2,4-Trichlorobenzene  
Concen: 18.409 ug/l  
RT: 13.591 min Scan# 2051  
Delta R.T. 0.000 min  
Lab File: VX022607.D  
Acq: 09 Jun 2021 12:36




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0609WBS01




Tgt Ion:180 Resp: 20752  
Ion Ratio Lower Upper  
180 100  
182 94.7 47.7 143.1  
145 34.5 15.0 45.0

Manual Integrations  
APPROVED



MMDadoda  
6/10/2021 12:11:56 PM




#94  
Hexachlorobutadiene  
Concen: 19.443 ug/l  
RT: 13.725 min Scan# 2073  
Delta R.T. -0.006 min  
Lab File: VX022607.D  
Acq: 09 Jun 2021 12:36




Tgt Ion:225 Resp: 7855  
Ion Ratio Lower Upper  
225 100  
223 57.0 31.1 93.2  
227 57.8 31.6 95.0







#95  
Naphthalene  
Concen: 18.201 ug/l  
RT: 13.780 min Scan# 2082  
Delta R.T. -0.000 min  
Lab File: VX022607.D  
Acq: 09 Jun 2021 12:36




Tgt Ion:128 Resp: 82255  
Ion Ratio Lower Upper  
128 100  
127 12.4 10.3 15.5  
129 10.9 8.8 13.2




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0609WBS01

Manual Integrations  
APPROVED

MMDadoda  
6/10/2021 12:11:56 PM



#96  
1,2,3-Trichlorobenzene  
Concen: 17.956 ug/l  
RT: 13.963 min Scan# 2112  
Delta R.T. -0.000 min  
Lab File: VX022607.D  
Acq: 09 Jun 2021 12:36



Tgt Ion:180 Resp: 20230  
Ion Ratio Lower Upper  
180 100  
182 95.4 47.6 142.9  
145 38.2 16.1 48.3

