

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060922\
 Data File : VX029341.D
 Acq On : 09 Jun 2022 11:27
 Operator : JC/MD
 Sample : VX0609MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0609MBL01

Quant Time: Jun 09 17:52:49 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 08 03:00:38 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	255324	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	429304	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	401840	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	207986	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	275197	71.563	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	143.120%#
35) Dibromofluoromethane	5.385	113	243133	66.617	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	133.240%#
50) Toluene-d8	8.646	98	786564	60.936	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	121.880%
62) 4-Bromofluorobenzene	11.079	95	338255	66.963	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	133.920%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060922\
Data File : VX029341.D
Acq On : 09 Jun 2022 11:27
Operator : JC/MD
Sample : VX0609MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0609MBL01

Quant Time: Jun 09 17:52:49 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060722W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 08 03:00:38 2022
Response via : Initial Calibration

