

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX060922\  
 Data File : VX029342.D  
 Acq On : 09 Jun 2022 13:42  
 Operator : JC/MD  
 Sample : VX0609WBL01  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 4 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 VX0609WBL01

Quant Time: Jun 09 17:53:13 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X060722W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Jun 08 03:00:38 2022  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon  | Response | Conc     | Units      | Dev(Min) |
|-----------------------------|--------|-------|----------|----------|------------|----------|
| -----                       |        |       |          |          |            |          |
| Internal Standards          |        |       |          |          |            |          |
| 1) Pentafluorobenzene       | 5.556  | 168   | 213980   | 50.000   | ug/l       | 0.00     |
| 34) 1,4-Difluorobenzene     | 6.763  | 114   | 360054   | 50.000   | ug/l       | 0.00     |
| 63) Chlorobenzene-d5        | 10.055 | 117   | 335834   | 50.000   | ug/l       | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 12.024 | 152   | 174826   | 50.000   | ug/l       | 0.00     |
| System Monitoring Compounds |        |       |          |          |            |          |
| 33) 1,2-Dichloroethane-d4   | 5.958  | 65    | 187004   | 57.681   | ug/l       | 0.00     |
| Spiked Amount               | 50.000 | Range | 61 - 141 | Recovery | = 115.360% |          |
| 35) Dibromofluoromethane    | 5.391  | 113   | 166349   | 54.345   | ug/l       | 0.00     |
| Spiked Amount               | 50.000 | Range | 69 - 133 | Recovery | = 108.700% |          |
| 50) Toluene-d8              | 8.653  | 98    | 511665   | 47.264   | ug/l       | 0.00     |
| Spiked Amount               | 50.000 | Range | 65 - 126 | Recovery | = 94.520%  |          |
| 62) 4-Bromofluorobenzene    | 11.085 | 95    | 234081   | 55.253   | ug/l       | 0.00     |
| Spiked Amount               | 50.000 | Range | 58 - 135 | Recovery | = 110.500% |          |

| Target Compounds | Qvalue |
|------------------|--------|
| -----            |        |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX060922\  
Data File : VX029342.D  
Acq On : 09 Jun 2022 13:42  
Operator : JC/MD  
Sample : VX0609WBL01  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 4 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0609WBL01

Quant Time: Jun 09 17:53:13 2022  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X060722W.M  
Quant Title : SW846 8260  
QLast Update : Wed Jun 08 03:00:38 2022  
Response via : Initial Calibration

