

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060922\
 Data File : VX029355.D
 Acq On : 09 Jun 2022 19:08
 Operator : JC/MD
 Sample : N3237-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-TB-20220607

Quant Time: Jun 10 08:10:27 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 08 03:00:38 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	209179	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	350075	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	327870	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	177944	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	184926	58.370	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 116.740%	
35) Dibromofluoromethane	5.391	113	163041	54.783	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 109.560%	
50) Toluene-d8	8.653	98	503081	47.795	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 95.600%	
62) 4-Bromofluorobenzene	11.085	95	239068	58.039	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 116.080%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060922\
Data File : VX029355.D
Acq On : 09 Jun 2022 19:08
Operator : JC/MD
Sample : N3237-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-TB-20220607

Quant Time: Jun 10 08:10:27 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060722W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 08 03:00:38 2022
Response via : Initial Calibration

