

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060925\
 Data File : VX046572.D
 Acq On : 09 Jun 2025 12:37
 Operator : JC/MD
 Sample : Q2240-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP-2

Quant Time: Jun 10 01:41:29 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060625W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 06 16:56:12 2025
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By : John Carlone 06/10/2025
 Supervised By : Mahesh Dadoda 06/10/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.568	168	111802	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	199705	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	183737	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	89206	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	96212	48.370	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	96.740%		
35) Dibromofluoromethane	5.397	113	69441	47.283	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	94.560%		
50) Toluene-d8	8.653	98	249299	51.488	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	102.980%		
62) 4-Bromofluorobenzene	11.079	95	97010	48.425	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	96.860%		
Target Compounds						
8) Diethyl Ether	2.148	74	24852	30.185	ug/l	91
16) Acetone	2.386	43	16622	25.484	ug/l	93
18) Methyl Acetate	2.721	43	9626	4.063	ug/l	98
20) Methylene Chloride	2.806	84	2784	1.745	ug/l #	87
37) Ethyl Acetate	4.751	43	5552m	2.712	ug/l	
43) Isopropyl Acetate	6.361	43	3903m	1.093	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060925\
Data File : VX046572.D
Acq On : 09 Jun 2025 12:37
Operator : JC/MD
Sample : Q2240-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP-2

Quant Time: Jun 10 01:41:29 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060625W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 06 16:56:12 2025
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :John Carlone 06/10/2025
Supervised By :Mahesh Dadoda 06/10/2025

