

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060925\
 Data File : VX046577.D
 Acq On : 09 Jun 2025 14:23
 Operator : JC/MD
 Sample : Q2244-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP03-MHC

Quant Time: Jun 10 01:43:40 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060625W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 06 16:56:12 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.568	168	144269	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	261183	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	235979	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	114518	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	113159	44.087	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	88.180%
35) Dibromofluoromethane	5.403	113	81570	42.468	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	84.940%
50) Toluene-d8	8.653	98	290791	45.921	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	91.840%
62) 4-Bromofluorobenzene	11.079	95	113910	43.477	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	86.960%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	35814	39.474	ug/l	98
20) Methylene Chloride	2.806	84	4147	2.014	ug/l	90
43) Isopropyl Acetate	6.354	43	10112	2.166	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060925\
Data File : VX046577.D
Acq On : 09 Jun 2025 14:23
Operator : JC/MD
Sample : Q2244-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP03-MHC

Quant Time: Jun 10 01:43:40 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060625W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 06 16:56:12 2025
Response via : Initial Calibration

