

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060925\
 Data File : VX046581.D
 Acq On : 09 Jun 2025 15:48
 Operator : JC/MD
 Sample : Q2265-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP11-MHL-WC

Quant Time: Jun 10 01:45:20 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060625W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 06 16:56:12 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.568	168	104693	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	194131	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	175205	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	86589	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	92218	49.510	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	99.020%
35) Dibromofluoromethane	5.397	113	66430	46.532	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.060%
50) Toluene-d8	8.653	98	233922	49.699	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.400%
62) 4-Bromofluorobenzene	11.079	95	91335	46.901	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	93.800%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	30769	45.889	ug/l	97
20) Methylene Chloride	2.806	84	4109	2.750	ug/l	96
43) Isopropyl Acetate	6.354	43	10539	3.037	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060925\
Data File : VX046581.D
Acq On : 09 Jun 2025 15:48
Operator : JC/MD
Sample : Q2265-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP11-MHL-WC

Quant Time: Jun 10 01:45:20 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060625W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 06 16:56:12 2025
Response via : Initial Calibration

