


BFB

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060925\
Data File : VX046566.D
Acq On : 09 Jun 2025 08:08
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060625W.M
Title : SW846 8260
Last Update : Fri Jun 06 16:56:12 2025



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1631

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	20.2	8666	PASS
75	95	30	60	54.3	23341	PASS
95	95	100	100	100.0	42968	PASS
96	95	5	9	6.4	2759	PASS
173	174	0.00	2	1.3	379	PASS
174	95	50	100	69.6	29896	PASS
175	174	5	9	7.4	2201	PASS
176	174	95	101	96.5	28840	PASS
177	176	5	9	6.8	1957	PASS