

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX061019\
 Data File : VX010212.D
 Acq On : 11 Jun 2019 05:47
 Operator : JC/SP
 Sample : K3163-08
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 52 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SU-04-060719-B

Quant Time: Jun 11 06:21:32 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X060719W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 08 02:12:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	254494	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	387901	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	355586	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	166405	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.06	65	113893	45.85	ug/l	0.00
Spiked Amount			Recovery	=	91.70%	
35) Dibromofluoromethane	5.50	113	118778	49.64	ug/l	0.00
Spiked Amount			Recovery	=	99.28%	
50) Toluene-d8	8.71	98	458530	51.13	ug/l	0.00
Spiked Amount			Recovery	=	102.26%	
62) 4-Bromofluorobenzene	11.13	95	157646	47.07	ug/l	0.00
Spiked Amount			Recovery	=	94.14%	


Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
3) Chloromethane	1.33	50	511	0.220	ug/l	90
5) Bromomethane	1.65	94	300	0.364	ug/l #	74
6) Chloroethane	1.76	64	529	0.480	ug/l #	43
8) Diethyl Ether	2.18	74	240	0.214	ug/l	65
11) Tert butyl alcohol	3.04	59	428	0.703	ug/l #	94
13) Acrolein	2.31	56	361	3.253	ug/l #	14
16) Acetone	2.44	43	14565	12.466	ug/l	90
18) Methyl Acetate	2.78	43	1602	0.623	ug/l #	72
20) Methylene Chloride	2.85	84	4931	1.922	ug/l #	75
25) 2-Butanone	4.70	43	1613	0.927	ug/l #	75
29) Tetrahydrofuran	5.16	42	1143	1.045	ug/l #	70
30) Chloroform	5.23	83	1134	0.273	ug/l	94
36) 1,1-Dichloropropene	5.66	75	15298	4.834	ug/l #	52
38) Carbon Tetrachloride	5.66	117	21614	6.034	ug/l #	16
43) Isopropyl Acetate	6.45	43	9378	1.703	ug/l #	89
44) Trichloroethene	7.22	130	623	0.211	ug/l	95
48) Methyl methacrylate	7.68	41	5616	2.181	ug/l #	20
49) 1,4-Dioxane	7.87	88	118	1.607	ug/l #	1
51) 4-Methyl-2-Pentanone	8.61	43	69120	21.016	ug/l #	54
54) cis-1,3-Dichloropropene	8.61	75	39815	9.364	ug/l #	52
57) 1,3-Dichloropropane	9.54	76	7896	1.904	ug/l #	43
59) 2-Hexanone	9.43	43	9163	3.494	ug/l #	27
76) 1,2,3-Trichloropropane	11.13	75	71348	21.436	ug/l #	34
80) 1,3,5-Trimethylbenzene	11.51	105	2575	0.263	ug/l	84
81) trans-1,4-Dichloro-2-buten	11.13	75	71348	47.361	ug/l #	10
84) 1,2,4-Trimethylbenzene	11.80	105	2216	0.224	ug/l	95
95) Naphthalene	13.83	128	17102	1.357	ug/l	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX061019\
Data File : VX010212.D
Acq On : 11 Jun 2019 05:47
Operator : JC/SP
Sample : K3163-08
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 52 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampled :
SU-04-060719-B

Quant Time: Jun 11 06:21:32 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X060719W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 08 02:12:13 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.66 min Scan# 739

Delta R.T. -0.01 min

Lab File: VX010212.D

Acq: 11 Jun 2019 05:47

Instrument :

MSVOA_X

ClientSampleId :


SU-04-060719-B

Tot Ion:168 Resp: 254494

Ion Ratio Lower Upper

168 100

99 48.3 43.0 64.6



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0

5.66

100000

80000



60000

40000

20000

0

Time-->



#3

Chloromethane

Concen: 0.220 ug/l

RT: 1.33 min Scan# 29

Delta R.T. 0.01 min

Lab File: VX010212.D


Acq: 11 Jun 2019 05:47

Tgt Ion: 50 Resp: 511

Ion Ratio Lower Upper

50 100

52 38.1 26.0 39.0



Abundance Ion 49.90 (49.60 to 50.60): VX0

Ion 51.90 (51.60 to 52.60): VX0

600

500

400


300


200

100

0

Time-->





#5

Bromomethane

Concen: 0.364 ug/l

RT: 1.65 min Scan# 81

Delta R.T. 0.01 min

Lab File: VX010212.D

Acq: 11 Jun 2019 05:47

Instrument :

MSVOA_X

ClientSampleId :


SU-04-060719-B

Tot Ion: 94 Resp: 300

Ion Ratio Lower Upper




94 100

96 120.2 76.2 114.4#



Abundance Ion 93.90 (93.60 to 94.60): VX0

Ion 95.90 (95.60 to 96.60): VX0



#6

Chloroethane

Concen: 0.480 ug/l

RT: 1.76 min Scan# 99

Delta R.T. 0.04 min

Lab File: VX010212.D


Acq: 11 Jun 2019 05:47

Tgt Ion: 64 Resp: 529

Ion Ratio Lower Upper



64 100


66 0.0 25.1 37.7#



Abundance Ion 63.90 (63.60 to 64.60): VX0

Ion 65.90 (65.60 to 66.60): VX0





#8

Diethyl Ether

Concen: 0.214 ug/l

RT: 2.18 min Scan# 169

Delta R.T. -0.01 min

Lab File: VX010212.D

Acq: 11 Jun 2019 05:47

Instrument :

MSVOA_X

ClientSampleId :


SU-04-060719-B

Tot Ion: 74 Resp: 240

Ion Ratio Lower Upper




74 100

45 75.0 56.1 168.3



Abundance Ion 74.00 (73.70 to 74.70): VX0

Ion 45.00 (44.70 to 45.70): VX0



#11

Tert butyl alcohol

Concen: 0.703 ug/l

RT: 3.04 min Scan# 309

Delta R.T. -0.01 min

Lab File: VX010212.D


Acq: 11 Jun 2019 05:47

Tgt Ion: 59 Resp: 428

Ion Ratio Lower Upper



59 100


57 12.6 8.2 12.2#



Abundance Ion 59.00 (58.70 to 59.70): VX0

Ion 57.00 (56.70 to 57.70): VX0





#13

Acrolein

Concen: 3.253 ug/l

RT: 2.31 min Scan# 190

Delta R.T. 0.02 min

Lab File: VX010212.D

Acq: 11 Jun 2019 05:47

Instrument :

MSVOA_X

ClientSampleId :





SU-04-060719-B

Tot Ion: 56 Resp: 361

Ion Ratio Lower Upper

56 100

55 0.0 57.0 85.4#



#16

Acetone

Concen: 12.466 ug/l

RT: 2.44 min Scan# 211

Delta R.T. -0.01 min

Lab File: VX010212.D




Acq: 11 Jun 2019 05:47


Tgt Ion: 43 Resp: 14565

Ion Ratio Lower Upper

43 100

58 34.2 23.2 34.8









#18
Methvl Acetate
Concen: 0.623 ug/l
RT: 2.78 min Scan# 266
Delta R.T. -0.00 min
Lab File: VX010212.D
Acq: 11 Jun 2019 05:47




Instrument :
MSVOA_X
ClientSampleId :
SU-04-060719-B


Tot Ion: 43 Resp: 1602
Ion Ratio Lower Upper
43 100
74 32.7 15.8 23.6#



#20
Methylene Chloride
Concen: 1.922 ug/l
RT: 2.85 min Scan# 279
Delta R.T. -0.00 min
Lab File: VX010212.D
Acq: 11 Jun 2019 05:47

Tgt Ion: 84 Resp: 4931
Ion Ratio Lower Upper
84 100
49 96.8 115.8 173.6#
51 32.9 34.4 51.6#
86 62.9 49.7 74.5





#25

2-Butanone

Concen: 0.927 ug/l

RT: 4.70 min Scan# 581

Delta R.T. 0.02 min

Lab File: VX010212.D

Acq: 11 Jun 2019 05:47

Instrument :

MSVOA_X

ClientSampled :





SU-04-060719-B

Tot Ion: 43 Resp: 1613

Ion Ratio Lower Upper

43 100

72 33.7 17.5 26.3#



#29

Tetrahydrofuran

Concen: 1.045 ug/l

RT: 5.16 min Scan# 657

Delta R.T. 0.03 min

Lab File: VX010212.D

Acq: 11 Jun 2019 05:47




Tgt Ion: 42 Resp: 1143


Ion Ratio Lower Upper

42 100

72 37.8 29.2 43.8

71 0.0 27.6 41.4#









#30
Chloroform
Concen: 0.273 ug/l
RT: 5.23 min Scan# 668
Delta R.T. 0.01 min
Lab File: VX010212.D
Acq: 11 Jun 2019 05:47




Instrument :
MSVOA_X
ClientSampleId :
SU-04-060719-B


Tot Ion: 83 Resp: 1134
Ion Ratio Lower Upper
83 100
85 60.8 52.6 78.8



#33
1,2-Dichloroethane-d4
Concen: 45.848 ug/l
RT: 6.06 min Scan# 805
Delta R.T. 0.00 min
Lab File: VX010212.D
Acq: 11 Jun 2019 05:47

Tgt Ion: 65 Resp: 113893
Ion Ratio Lower Upper
65 100
67 55.0 0.0 108.4





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.86 min Scan# 936

Delta R.T. -0.00 min

Lab File: VX010212.D

Acq: 11 Jun 2019 05:47

Instrument :

MSVOA_X

ClientSampleId :

SU-04-060719-B


Tot Ion:114 Resp: 387901

Ion Ratio Lower Upper

114 100

63 16.1 0.0 47.4




88 14.6 0.0 33.0



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VX0

Ion 87.90 (87.60 to 88.60): VX0



#35

Dibromofluoromethane

Concen: 49.643 ug/l

RT: 5.50 min Scan# 713

Delta R.T. -0.00 min

Lab File: VX010212.D

Acq: 11 Jun 2019 05:47


Tgt Ion:113 Resp: 118778

Ion Ratio Lower Upper

113 100

111 100.9 82.2 123.2



192 23.4 17.1 25.7




Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V









#36
 1,1-Dichloropropene
 Concen: 4.834 ug/l
 RT: 5.66 min Scan# 739
 Delta R.T. -0.14 min
 Lab File: VX010212.D
 Acq: 11 Jun 2019 05:47




Instrument :
 MSVOA_X
 ClientSampleId :
 SU-04-060719-B


Tot Ion: 75 Resp: 15298
 Ion Ratio Lower Upper
 75 100
 110 10.5 16.7 50.1#
 77 0.0 25.0 37.4#



#38
 Carbon Tetrachloride
 Concen: 6.034 ug/l
 RT: 5.66 min Scan# 739
 Delta R.T. -0.12 min
 Lab File: VX010212.D
 Acq: 11 Jun 2019 05:47

Tgt Ion: 117 Resp: 21614
 Ion Ratio Lower Upper
 117 100
 119 5.4 77.3 115.9#
 121 0.0 25.4 38.0#





#43

Isopropyl Acetate

Concen: 1.703 ug/l

RT: 6.45 min Scan# 869

Delta R.T. 0.01 min

Lab File: VX010212.D

Acq: 11 Jun 2019 05:47

Instrument :

MSVOA_X

ClientSampleId :

SU-04-060719-B


Tot Ion: 43 Resp: 9378

Ion Ratio Lower Upper

43 100

61 22.9 14.6 21.8#




87 15.0 8.7 13.1#



Abundance Ion 43.00 (42.70 to 43.70): VX0

Ion 61.00 (60.70 to 61.70): VX0

Ion 87.00 (86.70 to 87.70): VX0



#44

Trichloroethene

Concen: 0.211 ug/l

RT: 7.22 min Scan# 995

Delta R.T. 0.01 min

Lab File: VX010212.D


Acq: 11 Jun 2019 05:47

Tgt Ion:130 Resp: 623

Ion Ratio Lower Upper



130 100


95 98.7 0.0 207.4



Abundance Ion 129.80 (129.50 to 130.50): V

Ion 94.90 (94.60 to 95.60): VX0









#48
Methvl methacrylate
Concen: 2.181 ug/l
RT: 7.68 min Scan# 1070
Delta R.T. -0.09 min
Lab File: VX010212.D
Acq: 11 Jun 2019 05:47




Instrument :
MSVOA_X
ClientSampleId :
SU-04-060719-B


Tot Ion: 41 Resp: 5616
Ion Ratio Lower Upper
41 100
69 0.8 56.4 84.6#
39 0.0 40.6 60.8#



#49
1,4-Dioxane
Concen: 1.607 ug/l
RT: 7.87 min Scan# 1102
Delta R.T. 0.13 min
Lab File: VX010212.D
Acq: 11 Jun 2019 05:47

Tgt Ion: 88 Resp: 118
Ion Ratio Lower Upper
88 100
43 627423.7 30.2 45.4#
58 2821.2 64.0 96.0#





#50

Toluene-d8

Concen: 51.133 ug/l

RT: 8.71 min Scan# 1240

Delta R.T. 0.00 min

Lab File: VX010212.D

Acq: 11 Jun 2019 05:47

Instrument :

MSVOA_X

ClientSampled :


SU-04-060719-B

Tot Ion: 98 Resp: 458530

Ion Ratio Lower Upper




98 100

100 64.1 51.5 77.3



Abundance Ion 98.00 (97.70 to 98.70): VX0

Ion 100.00 (99.70 to 100.70): VX0



#51

4-Methyl-2-Pentanone

Concen: 21.016 ug/l

RT: 8.61 min Scan# 1223

Delta R.T. -0.03 min

Lab File: VX010212.D


Acq: 11 Jun 2019 05:47

Tgt Ion: 43 Resp: 69120

Ion Ratio Lower Upper



43 100


58 8.8 28.8 43.2#



Abundance Ion 43.00 (42.70 to 43.70): VX0

Ion 58.00 (57.70 to 58.70): VX0









#54

cis-1,3-Dichloropropene
 Concen: 9.364 ug/l
 RT: 8.61 min Scan# 1223
 Delta R.T. 0.18 min
 Lab File: VX010212.D
 Acq: 11 Jun 2019 05:47

Instrument :
 MSVOA_X
 ClientSampled :
 SU-04-060719-B




Tot Ion	Ratio	Lower	Upper
75	100		
77	0.7	24.8	37.2#
39	22.3	43.4	65.0#




#57

1,3-Dichloropropane
 Concen: 1.904 ug/l
 RT: 9.54 min Scan# 1376
 Delta R.T. 0.17 min
 Lab File: VX010212.D
 Acq: 11 Jun 2019 05:47

Tgt Ion	Ratio	Lower	Upper
76	100		
78	0.3	25.4	38.0#









#59
2-Hexanone
Concen: 3.494 ug/l
RT: 9.43 min Scan# 1358
Delta R.T. -0.05 min
Lab File: VX010212.D
Acq: 11 Jun 2019 05:47




Instrument :
MSVOA_X
ClientSampleId :
SU-04-060719-B


Tot Ion: 43 Resp: 9163
Ion Ratio Lower Upper
43 100
58 0.0 25.1 75.3#



#62
4-Bromofluorobenzene
Concen: 47.074 ug/l
RT: 11.13 min Scan# 1637
Delta R.T. 0.00 min
Lab File: VX010212.D
Acq: 11 Jun 2019 05:47

Tgt Ion: 95 Resp: 157646
Ion Ratio Lower Upper
95 100
174 94.5 0.0 160.4
176 92.4 0.0 154.6









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.11 min Scan# 1469
Delta R.T. -0.00 min
Lab File: VX010212.D
Acq: 11 Jun 2019 05:47

Instrument :
MSVOA_X
ClientSampled :
SU-04-060719-B

Tot Ion:117 Resp: 355586
Ion Ratio Lower Upper
117 100
82 50.2 49.2 73.8
119 31.8 25.2 37.8




Abundance Ion 116.90 (116.60 to 117.60): V
Ion 82.00 (81.70 to 82.70): VX0
Ion 118.90 (118.60 to 119.60): V






#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.08 min Scan# 1792
Delta R.T. -0.00 min
Lab File: VX010212.D
Acq: 11 Jun 2019 05:47

Tgt Ion:152 Resp: 166405
Ion Ratio Lower Upper
152 100
115 53.7 41.4 124.2
150 155.7 0.0 345.2



Abundance Ion 151.90 (151.60 to 152.60): V
Ion 114.90 (114.60 to 115.60): V
Ion 149.90 (149.60 to 150.60): V









#76
 1,2,3-Trichloropropane
 Concen: 21.436 ug/l
 RT: 11.13 min Scan# 1637
 Delta R.T. -0.16 min
 Lab File: VX010212.D
 Acq: 11 Jun 2019 05:47




Instrument :
 MSVOA_X
 ClientSampleId :
 SU-04-060719-B


Tot Ion: 75 Resp: 71348
 Ion Ratio Lower Upper
 75 100
 77 1.5 22.1 66.1#



#80
 1,3,5-Trimethylbenzene
 Concen: 0.263 ug/l
 RT: 11.51 min Scan# 1698
 Delta R.T. -0.00 min
 Lab File: VX010212.D
 Acq: 11 Jun 2019 05:47

Tgt Ion: 105 Resp: 2575
 Ion Ratio Lower Upper
 105 100
 120 58.8 24.1 72.3









#81
trans-1,4-Dichloro-2-butene
Concen: 47.361 ug/l
RT: 11.13 min Scan# 1637
Delta R.T. 0.06 min
Lab File: VX010212.D
Acq: 11 Jun 2019 05:47

Instrument :
MSVOA_X
ClientSampleId :
SU-04-060719-B

Tot Ion: 75 Resp: 71348
Ion Ratio Lower Upper
75 100
53 0.0 79.8 119.6#
89 0.0 34.2 51.4#




Abundance Ion 74.90 (74.60 to 75.60): VX0
Ion 53.00 (52.70 to 53.70): VX0
Ion 88.90 (88.60 to 89.60): VX0






#84
1,2,4-Trimethylbenzene
Concen: 0.224 ug/l
RT: 11.80 min Scan# 1747
Delta R.T. -0.00 min
Lab File: VX010212.D
Acq: 11 Jun 2019 05:47

Tgt Ion: 105 Resp: 2216
Ion Ratio Lower Upper
105 100
120 40.7 22.1 66.1



Abundance Ion 105.00 (104.70 to 105.70): V
Ion 120.00 (119.70 to 120.70): V





#95
Naphthalene
Concen: 1.357 ug/l
RT: 13.83 min Scan# 2080
Delta R.T. 0.01 min
Lab File: VX010212.D
Acq: 11 Jun 2019 05:47

Instrument :
MSVOA_X
ClientSampleId :
SU-04-060719-B

Tot Ion	Ratio	Lower	Upper
128	100		
127	13.9	10.6	16.0
129	11.7	8.8	13.2

