

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX061019\
 Data File : VX010178.D
 Acq On : 10 Jun 2019 15:22
 Operator : JC/SP
 Sample : K3249-09
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-DUP03-20190606

Quant Time: Jun 11 07:21:07 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X060719W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 08 02:12:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	261661	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	397211	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	350701	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	166038	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	117982	46.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.38%	
35) Dibromofluoromethane	5.49	113	122780	50.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.22%	
50) Toluene-d8	8.71	98	459441	50.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.06%	
62) 4-Bromofluorobenzene	11.13	95	156721	45.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.40%	
Target Compounds						
16) Acetone	2.44	43	2942	2.449	ug/l #	79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX061019\
Data File : VX010178.D
Acq On : 10 Jun 2019 15:22
Operator : JC/SP
Sample : K3249-09
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-DUP03-20190606

Quant Time: Jun 11 07:21:07 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X060719W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 08 02:12:13 2019
Response via : Initial Calibration

