

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX061019\
 Data File : VX010180.D
 Acq On : 10 Jun 2019 16:09
 Operator : JC/SP
 Sample : K3257-02DL 10X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP02-20190605DL

Quant Time: Jun 11 07:26:03 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X060719W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 08 02:12:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	253474	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	382445	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	351460	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	168722	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	114830	46.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.82%	
35) Dibromofluoromethane	5.49	113	118030	50.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.06%	
50) Toluene-d8	8.71	98	448194	50.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.38%	
62) 4-Bromofluorobenzene	11.13	95	159580	48.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.66%	
Target Compounds						
44) Trichloroethene	7.21	130	128393	44.148	ug/l	Qvalue 87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX061019\
Data File : VX010180.D
Acq On : 10 Jun 2019 16:09
Operator : JC/SP
Sample : K3257-02DL 10X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP02-20190605DL

Quant Time: Jun 11 07:26:03 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X060719W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 08 02:12:13 2019
Response via : Initial Calibration

