

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX061019\
 Data File : VX010206.D
 Acq On : 11 Jun 2019 03:27
 Operator : JC/SP
 Sample : K3279-09
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-EB03-20190607

Quant Time: Jun 11 08:36:11 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X060719W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 08 02:12:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	266265	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	402374	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	357815	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	165472	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	118113	45.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.88%	
35) Dibromofluoromethane	5.50	113	122246	49.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.50%	
50) Toluene-d8	8.71	98	470213	50.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.10%	
62) 4-Bromofluorobenzene	11.13	95	160069	46.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.16%	
Target Compounds						
16) Acetone	2.44	43	10515	8.602	ug/l	96
20) Methylene Chloride	2.85	84	1582	0.589	ug/l #	73
44) Trichloroethene	7.21	130	3039	0.993	ug/l	82
52) Toluene	8.79	92	23497	3.521	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX061019\
Data File : VX010206.D
Acq On : 11 Jun 2019 03:27
Operator : JC/SP
Sample : K3279-09
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 46 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-EB03-20190607

Quant Time: Jun 11 08:36:11 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X060719W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 08 02:12:13 2019
Response via : Initial Calibration

