

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX061019\
 Data File : VX010211.D
 Acq On : 11 Jun 2019 05:23
 Operator : JC/SP
 Sample : K3262-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 51 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STOCK-PILE-1

Quant Time: Jun 11 08:48:04 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X060719W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 08 02:12:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	257705	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	387587	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	346769	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	164423	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	115693	45.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.98%	
35) Dibromofluoromethane	5.50	113	119088	49.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.62%	
50) Toluene-d8	8.71	98	456838	50.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.98%	
62) 4-Bromofluorobenzene	11.13	95	153668	45.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.84%	
Target Compounds						
16) Acetone	2.45	43	11140	9.416	ug/l	90
20) Methylene Chloride	2.86	84	3477	1.338	ug/l #	77

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX061019\
Data File : VX010211.D
Acq On : 11 Jun 2019 05:23
Operator : JC/SP
Sample : K3262-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 51 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STOCK-PILE-1

Quant Time: Jun 11 08:48:04 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X060719W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 08 02:12:13 2019
Response via : Initial Calibration

