


Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX061019\
Data File : VX010193.D
Acq On : 10 Jun 2019 21:36
Operator : JC/SP
Sample : BFB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X060719W.M
Title : SW846 8260
Last Update : Sat Jun 08 02:12:13 2019



AutoFind: Scans 1636, 1637, 1638; Background Corrected with Scan 1630

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	15.4	7620	PASS
75	95	30	60	46.1	22738	PASS
95	95	100	100	100.0	49360	PASS
96	95	5	9	7.0	3457	PASS
173	174	0.00	2	1.1	515	PASS
174	95	50	100	91.7	45253	PASS
175	174	5	9	7.2	3253	PASS
176	174	95	101	96.7	43773	PASS
177	176	5	9	6.9	3027	PASS