

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX061020\
 Data File : VX016670.D
 Acq On : 10 Jun 2020 13:30
 Operator : JC/SP
 Sample : L2962-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SS003RW114R-200609

Quant Time: Jun 11 03:31:12 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X052720W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 27 16:31:05 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	220298	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	346302	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	348001	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	185773	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	129375	47.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.58%	
35) Dibromofluoromethane	5.46	113	105476	49.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.92%	
50) Toluene-d8	8.70	98	426588	51.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.80%	
62) 4-Bromofluorobenzene	11.12	95	173498	54.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.20%	
Target Compounds						
40) Benzene	6.12	78	3281	0.336	ug/l	93
65) Chlorobenzene	10.12	112	491598	69.290	ug/l	99
87) 1,3-Dichlorobenzene	12.01	146	137471	23.503	ug/l	99
88) 1,4-Dichlorobenzene	12.08	146	386333	64.583	ug/l	99
91) 1,2-Dichlorobenzene	12.38	146	2012302	355.110	ug/l	99
93) 1,2,4-Trichlorobenzene	13.63	180	10436	2.773	ug/l	97
96) 1,2,3-Trichlorobenzene	14.01	180	6267	1.700	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX061020\
Data File : VX016670.D
Acq On : 10 Jun 2020 13:30
Operator : JC/SP
Sample : L2962-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SS003RW114R-200609

Quant Time: Jun 11 03:31:12 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X052720W.M
Quant Title : SW846 8260
QLast Update : Wed May 27 16:31:05 2020
Response via : Initial Calibration

