

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX061020\
 Data File : VX016679.D
 Acq On : 10 Jun 2020 16:54
 Operator : JC/SP
 Sample : L2963-06
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 LF014MW009-200609

Quant Time: Jun 11 03:36:02 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X052720W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 27 16:31:05 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	223198	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	353451	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	361533	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	187634	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	131840	47.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.12%	
35) Dibromofluoromethane	5.47	113	109958	50.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.04%	
50) Toluene-d8	8.70	98	439087	51.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.68%	
62) 4-Bromofluorobenzene	11.12	95	177452	54.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.42%	
Target Compounds						
4) Vinyl Chloride	1.39	62	112026	39.908	ug/l	98
16) Acetone	2.42	43	3362	2.768	ug/l	98
27) cis-1,2-Dichloroethene	4.57	96	72417	26.351	ug/l	89
40) Benzene	6.12	78	2657	0.267	ug/l	90
91) 1,2-Dichlorobenzene	12.38	146	2370	0.414	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX061020\
Data File : VX016679.D
Acq On : 10 Jun 2020 16:54
Operator : JC/SP
Sample : L2963-06
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
LF014MW009-200609

Quant Time: Jun 11 03:36:02 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X052720W.M
Quant Title : SW846 8260
QLast Update : Wed May 27 16:31:05 2020
Response via : Initial Calibration

