

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX061020\
 Data File : VX016682.D
 Acq On : 10 Jun 2020 18:02
 Operator : JC/SP
 Sample : L2963-09
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 LF014MW080-200609

Quant Time: Jun 11 03:41:43 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X052720W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 27 16:31:05 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	216645	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	344916	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	357550	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	183625	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	128665	47.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.64%	
35) Dibromofluoromethane	5.47	113	105891	49.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.72%	
50) Toluene-d8	8.70	98	427027	51.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.32%	
62) 4-Bromofluorobenzene	11.12	95	174206	54.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.08%	
Target Compounds						
16) Acetone	2.42	43	2667	2.263	ug/l	95
27) cis-1,2-Dichloroethene	4.57	96	1107	0.415	ug/l	84
65) Chlorobenzene	10.12	112	24643	3.381	ug/l	99
91) 1,2-Dichlorobenzene	12.38	146	2286	0.408	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX061020\
Data File : VX016682.D
Acq On : 10 Jun 2020 18:02
Operator : JC/SP
Sample : L2963-09
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
LF014MW080-200609

Quant Time: Jun 11 03:41:43 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X052720W.M
Quant Title : SW846 8260
QLast Update : Wed May 27 16:31:05 2020
Response via : Initial Calibration

