

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX061020\
 Data File : VX016688.D
 Acq On : 10 Jun 2020 20:19
 Operator : JC/SP
 Sample : L2963-17
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 LF014RW133-200608

Quant Time: Jun 11 03:51:44 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X052720W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 27 16:31:05 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	213957	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	335637	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	348728	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	174501	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	129188	48.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.24%	
35) Dibromofluoromethane	5.47	113	104322	50.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.96%	
50) Toluene-d8	8.70	98	415630	51.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.34%	
62) 4-Bromofluorobenzene	11.12	95	168408	54.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.36%	
Target Compounds						
16) Acetone	2.42	43	2311	1.985	ug/l	99
27) cis-1,2-Dichloroethene	4.57	96	3118	1.184	ug/l	86
65) Chlorobenzene	10.12	112	9996	1.406	ug/l	95
91) 1,2-Dichlorobenzene	12.38	146	1585	0.298	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX061020\
Data File : VX016688.D
Acq On : 10 Jun 2020 20:19
Operator : JC/SP
Sample : L2963-17
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
LF014RW133-200608

Quant Time: Jun 11 03:51:44 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X052720W.M
Quant Title : SW846 8260
QLast Update : Wed May 27 16:31:05 2020
Response via : Initial Calibration

