

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061021\
 Data File : VX022639.D
 Acq On : 10 Jun 2021 13:13
 Operator : JC/MD
 Sample : VX0610MBL01
 Misc : 5.00g/10mL/100uL/5.0mL/MSVOA_X/MEOH
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0610MBL01

Quant Time: Jun 11 06:42:18 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061021W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 10 12:20:00 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	58966	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	119428	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	106025	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	39731	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	49367	46.551	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	93.100%
35) Dibromofluoromethane	5.391	113	36014	47.264	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.520%
50) Toluene-d8	8.653	98	147711	49.175	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.360%
62) 4-Bromofluorobenzene	11.085	95	50605	44.546	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.100%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061021\
Data File : VX022639.D
Acq On : 10 Jun 2021 13:13
Operator : JC/MD
Sample : VX0610MBL01
Misc : 5.00g/10mL/100uL/5.0mL/MSVOA_X/MEOH
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0610MBL01

Quant Time: Jun 11 06:42:18 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061021W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 10 12:20:00 2021
Response via : Initial Calibration

