

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061021\
 Data File : VX022640.D
 Acq On : 10 Jun 2021 13:37
 Operator : JC/MD
 Sample : VX0610WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0610WBL01

Quant Time: Jun 11 06:42:40 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061021W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 10 12:20:00 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	57428	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	118554	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	105469	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	41666	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	49843	48.259	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.520%
35) Dibromofluoromethane	5.397	113	35023	46.302	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.600%
50) Toluene-d8	8.653	98	145272	48.720	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.440%
62) 4-Bromofluorobenzene	11.085	95	50370	44.666	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.340%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061021\
Data File : VX022640.D
Acq On : 10 Jun 2021 13:37
Operator : JC/MD
Sample : VX0610WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0610WBL01

Quant Time: Jun 11 06:42:40 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061021W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 10 12:20:00 2021
Response via : Initial Calibration

