

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061021\
 Data File : VX022645.D
 Acq On : 10 Jun 2021 15:44
 Operator : JC/MD
 Sample : M2620-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-4

Quant Time: Jun 11 06:45:00 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061021W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 10 12:20:00 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	58505	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	117681	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	105720	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	40595	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.970	65	49357	46.909	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	93.820%
35) Dibromofluoromethane	5.397	113	35209	46.893	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.780%
50) Toluene-d8	8.653	98	145150	49.040	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.080%
62) 4-Bromofluorobenzene	11.085	95	49303	44.044	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.080%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061021\
Data File : VX022645.D
Acq On : 10 Jun 2021 15:44
Operator : JC/MD
Sample : M2620-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-4

Quant Time: Jun 11 06:45:00 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061021W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 10 12:20:00 2021
Response via : Initial Calibration

