

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061021\
 Data File : VX022647.D
 Acq On : 10 Jun 2021 16:31
 Operator : JC/MD
 Sample : M2655-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 0226-FIELD-BLANK-2

Quant Time: Jun 11 06:45:35 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061021W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 10 12:20:00 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.568	168	57332	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	118113	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	103243	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	39257	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	48926	47.450	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	94.900%
35) Dibromofluoromethane	5.397	113	34042	45.173	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.340%
50) Toluene-d8	8.653	98	143469	48.295	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.580%
62) 4-Bromofluorobenzene	11.085	95	49357	43.931	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	87.860%
Target Compounds						
16) Acetone	2.386	43	1140	2.469	ug/l	Qvalue # 82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061021\
Data File : VX022647.D
Acq On : 10 Jun 2021 16:31
Operator : JC/MD
Sample : M2655-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
0226-FIELD-BLANK-2

Quant Time: Jun 11 06:45:35 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061021W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 10 12:20:00 2021
Response via : Initial Calibration

