

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061021\
 Data File : VX022648.D
 Acq On : 10 Jun 2021 16:54
 Operator : JC/MD
 Sample : M2652-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PDB-BLANK-20210608

Quant Time: Jun 11 06:45:52 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061021W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 10 12:20:00 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	56102	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	112863	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	100303	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	37387	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	47648	47.224	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	94.440%
35) Dibromofluoromethane	5.397	113	33870	47.036	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.080%
50) Toluene-d8	8.653	98	139833	49.260	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.520%
62) 4-Bromofluorobenzene	11.085	95	47778	44.503	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.000%
Target Compounds						
						Qvalue
4) Vinyl Chloride	1.374	62	473	0.605	ug/l	95
16) Acetone	2.386	43	4451	9.852	ug/l	98
27) cis-1,2-Dichloroethene	4.495	96	7818	8.502	ug/l	93
44) Trichloroethene	7.135	130	4911	6.253	ug/l	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061021\
Data File : VX022648.D
Acq On : 10 Jun 2021 16:54
Operator : JC/MD
Sample : M2652-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PDB-BLANK-20210608

Quant Time: Jun 11 06:45:52 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061021W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 10 12:20:00 2021
Response via : Initial Calibration

