

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061021\
 Data File : VX022650.D
 Acq On : 10 Jun 2021 17:41
 Operator : JC/MD
 Sample : M2652-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP09-20210608

Quant Time: Jun 11 06:46:33 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061021W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 10 12:20:00 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.568	168	58347	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	121183	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	105435	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	41044	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	50107	47.750	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.500%
35) Dibromofluoromethane	5.391	113	35013	45.285	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.560%
50) Toluene-d8	8.653	98	145590	47.767	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.540%
62) 4-Bromofluorobenzene	11.085	95	49927	43.312	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	86.620%
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.172	85	1739	2.540	ug/l	97
4) Vinyl Chloride	1.374	62	19456	23.933	ug/l	97
12) 1,1-Dichloroethene	2.325	96	517	0.730	ug/l	83
16) Acetone	2.386	43	2045	4.352	ug/l #	78
21) trans-1,2-Dichloroethene	3.099	96	1442	1.837	ug/l	97
27) cis-1,2-Dichloroethene	4.495	96	218145	228.097	ug/l	93
44) Trichloroethene	7.135	130	2755	3.267	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061021\
Data File : VX022650.D
Acq On : 10 Jun 2021 17:41
Operator : JC/MD
Sample : M2652-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP09-20210608

Quant Time: Jun 11 06:46:33 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061021W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 10 12:20:00 2021
Response via : Initial Calibration

