



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061021\
 Data File : VX022630.D
 Acq On : 10 Jun 2021 09:19
 Operator : JC/MD
 Sample : BFB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061021W.M
 Title : SW846 8260
 Last Update : Thu Jun 10 12:20:00 2021



AutoFind: Scans 1639, 1640, 1641; Background Corrected with Scan 1632

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.3	5281	PASS
75	95	30	60	52.1	14259	PASS
95	95	100	100	100.0	27349	PASS
96	95	5	9	6.5	1775	PASS
173	174	0.00	2	0.8	140	PASS
174	95	50	100	62.9	17192	PASS
175	174	5	9	7.4	1272	PASS
176	174	95	101	95.4	16400	PASS
177	176	5	9	7.2	1188	PASS