

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061022\
 Data File : VX029367.D
 Acq On : 10 Jun 2022 14:30
 Operator : JC/MD
 Sample : N3237-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-TB-20220607

Quant Time: Jun 11 00:53:06 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 08 03:00:38 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	179433	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	299298	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	302884	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	176358	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	164491	60.595	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 121.180%	
35) Dibromofluoromethane	5.391	113	141341	55.549	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 111.100%	
50) Toluene-d8	8.653	98	463083	51.459	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 102.920%	
62) 4-Bromofluorobenzene	11.085	95	189891	53.921	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 107.840%	
Target Compounds						
						Qvalue
16) Acetone	2.385	43	4053	2.792	ug/l	97
20) Methylene Chloride	2.788	84	1100	0.339	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061022\
Data File : VX029367.D
Acq On : 10 Jun 2022 14:30
Operator : JC/MD
Sample : N3237-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-TB-20220607

Quant Time: Jun 11 00:53:06 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060722W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 08 03:00:38 2022
Response via : Initial Calibration

